

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072617\
 Data File : BE093610.D
 Acq On : 26 Jul 2017 18:28
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 27 08:35:37 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 21 16:13:20 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.664	0.686	-3.3	86	0.00
3	n-Nitrosodimethylamine	0.657	0.613	6.7	83	0.00
4 S	2-Fluorophenol	1.388	1.407	-1.4	85	0.00
5 S	Phenol-d6	1.945	2.133	-9.7	95	0.00
6	bis(2-Chloroethyl)ether	1.548	1.659	-7.2	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	83	0.00
8 S	Nitrobenzene-d5	0.318	0.347	-9.1	95	0.00
9	Naphthalene	4.659	5.093	-9.3	94	0.00
10	Hexachlorobutadiene	0.181	0.189	-4.4	89	-0.02
11 SURR	2-Methylnaphthalene-d10	0.647	0.737	-13.9	97	0.00
12	2-Methylnaphthalene	0.781	0.877	-12.3	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.145	0.103	29.0#	62	0.00
15 S	2-Fluorobiphenyl	1.574	1.687	-7.2	95	-0.02
16	Acenaphthylene	2.014	2.099	-4.2	98	0.00
17	Acenaphthene	1.368	1.455	-6.4	96	0.00
18	Fluorene	1.926	1.944	-0.9	91	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
20	4-Bromophenyl-phenylether	0.156	0.134	14.1	83	0.00
21	Hexachlorobenzene	0.150	0.138	8.0	86	0.00
22	Pentachlorophenol	0.088	0.086	2.3	104	0.00
23	Phenanthrene	0.907	0.936	-3.2	91	0.00
24	Anthracene	1.387	1.399	-0.9	89	0.00
25 SURR	Fluoranthene-d10	6.369	4.944	22.4	68	0.00
26	Fluoranthene	1.935	1.512	21.9	69	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	61	-0.01
28	Pyrene	1.296	1.450	-11.9	69	0.00
29 S	Terphenyl-d14	0.728	0.797	-9.5	67	0.00
30	Benzo(a)anthracene	1.286	1.327	-3.2	64	0.00
31	Chrysene	1.275	1.403	-10.0	67	0.00
32	Bis(2-ethylhexyl)phthalate	2.788	1.939	30.5#	46#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.206	1.419	-17.7	73	0.00
34 I	Perylene-d12	1.000	1.000	0.0	65	0.00
35	Benzo(b)fluoranthene	1.405	1.516	-7.9	70	0.00
36	Benzo(k)fluoranthene	1.388	1.542	-11.1	73	0.00
37 C	Benzo(a)pyrene	1.306	1.430	-9.5	73	0.00
38	Dibenzo(a,h)anthracene	1.235	1.370	-10.9	73	0.00
39	Benzo(g,h,i)perylene	1.240	1.324	-6.8	70	0.00

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072617\
Data File : BE093610.D
Acq On : 26 Jul 2017 18:28
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 27 08:35:37 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 21 16:13:20 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			