

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072716\
 Data File : BE092062.D
 Acq On : 27 Jul 2016 18:39
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 28 05:58:53 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 27 17:13:26 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	105	-0.02
2	1,4-Dioxane	0.400	0.417	-4.2	109	0.00
3	n-Nitrosodimethylamine	0.400	0.390	2.5	113	0.00
4 S	2-Fluorophenol	0.400	0.366	8.5	110	-0.02
5 S	Phenol-d6	0.400	0.343	14.2	102	0.00
6	bis(2-Chloroethyl)ether	0.400	0.374	6.5	108	0.00
7	n-Nitroso-di-n-propylamine	0.400	0.369	7.8	110	0.00
8 I	Naphthalene-d8	5.000	5.000	0.0	104	0.00
9 S	Nitrobenzene-d5	0.400	0.349	12.8	97	0.00
10	Nitrobenzene	0.400	0.372	7.0	109	0.00
11	bis(2-Chloroethoxy)methane	0.400	0.363	9.3	108	0.00
12	Naphthalene	0.400	0.404	-1.0	105	0.00
13	Hexachlorobutadiene	0.400	0.416	-4.0	110	0.00
14	2-Methylnaphthalene	0.400	0.396	1.0	110	0.00
15 I	Acenaphthene-d10	5.000	5.000	0.0	114	0.00
16 S	2,4,6-Tribromophenol	0.400	0.434	-8.5	117	0.00
17 S	2-Fluorobiphenyl	0.400	0.377	5.8	109	0.00
18	Acenaphthylene	0.400	0.369	7.8	114	0.00
19	2,6-Dinitrotoluene	0.400	0.415	-3.7	111	0.00
20	Acenaphthene	0.400	0.385	3.8	112	0.00
21	2,4-Dinitrotoluene	0.400	0.360	10.0	118	0.00
22	Fluorene	0.400	0.376	6.0	111	0.00
23	4-Chlorophenyl-phenylether	0.400	0.381	4.8	109	0.00
24 I	Phenanthrene-d10	5.000	5.000	0.0	108	0.00
25	4-Bromophenyl-phenylether	0.400	0.402	-0.5	112	0.00
26	Hexachlorobenzene	0.400	0.401	-0.3	111	0.00
27	Pentachlorophenol	0.400	0.504	-26.0#	121	0.00
28	Phenanthrene	0.400	0.405	-1.3	113	0.00
29	Anthracene	0.400	0.391	2.3	114	0.00
30	Fluoranthene	0.400	0.372	7.0	115	0.00
31 I	Chrysene-d12	5.000	5.000	0.0	110	0.00
32	Benzidine	0.400	0.351	12.3	102	0.00
33	Pyrene	0.400	0.365	8.8	110	0.00
34 S	Terphenyl-d14	0.400	0.324	19.0	100	0.00
35	Benzo(a)anthracene	0.400	0.410	-2.5	112	0.00
36	3,3'-Dichlorobenzidine	0.400	0.432	-8.0	106	0.00
37	Chrysene	0.400	0.368	8.0	117	0.00
38	Bis(2-ethylhexyl)phthalate	0.400	0.451	-12.7	126	0.00
39	Indeno(1,2,3-cd)pyrene	0.400	0.358	10.5	105	0.00
40 I	Perylene-d12	5.000	5.000	0.0	104	0.00
41	Benzo(b)fluoranthene	0.400	0.432	-8.0	114	0.00

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE072716\
Data File : BE092062.D
Acq On : 27 Jul 2016 18:39
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 28 05:58:53 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE072716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 27 17:13:26 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
42	Benzo(k)fluoranthene	0.400	0.467	-16.8	114	0.00
43 C	Benzo(a)pyrene	0.400	0.403	-0.8	108	0.01
44	Dibenzo(a,h)anthracene	0.400	0.372	7.0	105	0.02
45	Benzo(g,h,i)perylene	0.400	0.377	5.8	106	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0