

Data Path : Z:\HPCHEM1\BNA_E\Data\BE072915\
 Data File : BE090282.D
 Acq On : 29 Jul 2015 14:09
 Operator : TP/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.449

Quant Time: Jul 30 01:51:13 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE072815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 29 13:56:48 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
3	Naphthalene	1.057	1.048	0.9	98	0.00
4 SURR	2-Methylnaphthalene-d10	0.615	0.619	-0.7	102	0.00
5	2-Methylnaphthalene	0.684	0.688	-0.6	106	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
7	Acenaphthylene	8.346	8.283	0.8	101	0.00
8 C	Acenaphthene	6.150	6.008	2.3	99	0.00
9	Fluorene	1.655	1.582	4.4	97	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
11	Pentachlorophenol	0.026	0.017	34.6#	66	0.01
12	Phenanthrene	4.283	4.256	0.6	93	0.00
13	Anthracene	4.369	4.293	1.7	93	0.00
14 SURR	Fluoranthene-d10	1.107	1.066	3.7	88	0.00
15 C	Fluoranthene	5.028	4.802	4.5	88	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
17	Pyrene	7.425	7.249	2.4	87	0.00
18	Benzo(a)anthracene	0.594	0.492	17.2	73	0.00
19	Chrysene	1.743	1.774	-1.8	92	0.00
20 I	Perylene-d12	1.000	1.000	0.0	68	0.00
21	Benzo(b)fluoranthene	0.599	0.470	21.5#	61	0.00
22	Benzo(k)fluoranthene	1.890	2.000	-5.8	78	0.00
23 C	Benzo(a)pyrene	0.815	0.642	21.2#	68	0.00
24	Indeno(1,2,3-cd)pyrene	2.532	1.783	29.6#	60	0.00
25	Dibenzo(a,h)anthracene	0.300	0.236	21.3#	69	0.00
26	Benzo(g,h,i)perylene	3.557	2.903	18.4	58	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1