

Data Path : Z:\HPCHEM1\BNA_E\Data\BE072915\
 Data File : BE090290.D
 Acq On : 29 Jul 2015 20:57
 Operator : TP/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.450

Quant Time: Jul 30 02:42:04 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE072815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 30 01:55:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
3	Naphthalene	1.057	1.046	1.0	102	0.00
4 SURR	2-Methylnaphthalene-d10	0.615	0.620	-0.8	106	0.00
5	2-Methylnaphthalene	0.684	0.694	-1.5	111	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
7	Acenaphthylene	8.346	8.366	-0.2	106	0.00
8 C	Acenaphthene	6.150	5.955	3.2	102	0.00
9	Fluorene	1.655	1.592	3.8	102	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
11	Pentachlorophenol	0.026	0.020	23.1#	81	0.00
12	Phenanthrene	4.283	4.256	0.6	99	0.00
13	Anthracene	4.369	4.329	0.9	100	0.00
14 SURR	Fluoranthene-d10	1.107	1.105	0.2	97	0.00
15 C	Fluoranthene	5.028	4.942	1.7	96	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
17	Pyrene	7.425	7.447	-0.3	98	0.00
18	Benzo(a)anthracene	0.594	0.534	10.1	87	0.00
19	Chrysene	1.743	1.839	-5.5	104	0.00
20 I	Perylene-d12	1.000	1.000	0.0	81	0.00
21	Benzo(b)fluoranthene	0.599	0.422	29.5#	65	0.00
22	Benzo(k)fluoranthene	1.890	1.989	-5.2	91	0.00
23 C	Benzo(a)pyrene	0.815	0.626	23.2#	78	0.00
24	Indeno(1,2,3-cd)pyrene	2.532	1.593	37.1#	63	0.00
25	Dibenzo(a,h)anthracene	0.300	0.207	31.0#	71	0.00
26	Benzo(g,h,i)perylene	3.557	3.019	15.1	72	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1