

Data Path : Z:\HPCHEM1\BNA_E\Data\BE073015\
 Data File : BE090292.D
 Acq On : 30 Jul 2015 11:32
 Operator : TP/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.451

Quant Time: Jul 31 06:10:51 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE072815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 30 02:43:45 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
3	Naphthalene	1.057	1.062	-0.5	98	0.00
4 SURR	2-Methylnaphthalene-d10	0.615	0.639	-3.9	104	0.00
5	2-Methylnaphthalene	0.684	0.703	-2.8	106	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
7	Acenaphthylene	8.346	8.327	0.2	100	0.00
8 C	Acenaphthene	6.150	6.141	0.1	99	0.00
9	Fluorene	1.655	1.669	-0.8	101	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
11	Pentachlorophenol	0.026	0.016	38.5#	67	0.00
12	Phenanthrene	4.283	4.371	-2.1	101	0.00
13	Anthracene	4.369	4.311	1.3	99	0.00
14 SURR	Fluoranthene-d10	1.107	1.031	6.9	90	0.00
15 C	Fluoranthene	5.028	4.646	7.6	90	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	78	0.00
17	Pyrene	7.425	7.793	-5.0	88	0.00
18	Benzo(a)anthracene	0.594	0.458	22.9#	65	0.00
19	Chrysene	1.743	1.896	-8.8	93	0.00
20 I	Perylene-d12	1.000	1.000	0.0	53	0.00
21	Benzo(b)fluoranthene	0.599	0.539	10.0	55	0.00
22	Benzo(k)fluoranthene	1.890	2.408	-27.4#	72	0.00
23 C	Benzo(a)pyrene	0.815	0.637	21.8#	52	0.00
24	Indeno(1,2,3-cd)pyrene	2.532	1.929	23.8#	50	0.00
25	Dibenzo(a,h)anthracene	0.300	0.258	14.0	58	0.00
26	Benzo(g,h,i)perylene	3.557	2.981	16.2	46	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1