

Data Path : Z:\HPCHEM1\BNA_E\Data\BE073015\
 Data File : BE090300.D
 Acq On : 30 Jul 2015 17:03
 Operator : TP/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.452

Quant Time: Jul 31 06:49:28 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE072815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 31 06:10:41 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
3	Naphthalene	1.057	1.024	3.1	103	0.00
4 SURR	2-Methylnaphthalene-d10	0.615	0.608	1.1	107	0.00
5	2-Methylnaphthalene	0.684	0.681	0.4	112	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
7	Acenaphthylene	8.346	8.556	-2.5	106	0.00
8 C	Acenaphthene	6.150	6.089	1.0	102	0.00
9	Fluorene	1.655	1.602	3.2	100	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
11	Pentachlorophenol	0.026	0.020	23.1#	81	0.00
12	Phenanthrene	4.283	4.335	-1.2	99	0.00
13	Anthracene	4.369	4.360	0.2	98	0.00
14 SURR	Fluoranthene-d10	1.107	1.085	2.0	93	0.00
15 C	Fluoranthene	5.028	4.869	3.2	92	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
17	Pyrene	7.425	7.487	-0.8	93	0.00
18	Benzo(a)anthracene	0.594	0.497	16.3	77	0.00
19	Chrysene	1.743	1.862	-6.8	100	0.00
20 I	Perylene-d12	1.000	1.000	0.0	67	0.00
21	Benzo(b)fluoranthene	0.599	0.426	28.9#	55	0.00
22	Benzo(k)fluoranthene	1.890	2.035	-7.7	78	0.00
23 C	Benzo(a)pyrene	0.815	0.667	18.2	69	0.00
24	Indeno(1,2,3-cd)pyrene	2.532	1.790	29.3#	59	0.00
25	Dibenzo(a,h)anthracene	0.300	0.244	18.7	70	0.00
26	Benzo(g,h,i)perylene	3.557	2.596	27.0#	51	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0