

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080216\
 Data File : BE092111.D
 Acq On : 2 Aug 2016 14:41
 Operator : UM/SJ
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Aug 02 15:46:37 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 13:43:25 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.18	152	12006	5.00	ng	0.01
6) Naphthalene-d8	10.99	136	56943	5.00	ng	0.00
10) Acenaphthene-d10	14.85	164	32281	5.00	ng	0.00
16) Phenanthrene-d10	17.58	188	66637	5.00	ng	0.00
22) Chrysene-d12	21.75	240	84891	5.00	ng	0.00
28) Perylene-d12	24.14	264	84733	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.70	112	9010	2.30	ng	0.00
5) Phenol-d6	7.34	99	12567	2.67	ng	-0.01
7) Nitrobenzene-d5	9.34	82	12651	2.59	ng	-0.01
11) 2,4,6-Tribromophenol	16.33	330	3573	1.53	ng	0.00
12) 2-Fluorobiphenyl	13.46	172	29914	2.39	ng	0.00
24) Terphenyl-d14	20.19	244	37622	2.92	ng	-0.02


Target Compounds				Qvalue		
2) 1,4-Dioxane	3.44	88	4418	2.55	ng	92
3) n-Nitrosodimethylamine	3.78	42	6617	3.00	ng	# 97
8) Naphthalene	11.03	128	36625	2.79	ng	# 95
9) 2-Methylnaphthalene	12.65	142	24262	2.76	ng	98
13) Acenaphthylene	14.54	152	170345	2.38	ng	98
14) Acenaphthene	14.90	154	25524	2.58	ng	# 92
15) Fluorene	15.88	166	32794	2.21	ng	99
17) Hexachlorobenzene	16.89	284	9576	4.30	ng	# 61
18) Pentachlorophenol	17.24	266	3877	9.49	ng	# 82
19) Phenanthrene	17.60	178	56195	3.14	ng	98
20) Anthracene	17.70	178	57158	4.30	ng	99
21) Fluoranthene	19.63	202	245537	4.70	ng	96
23) Pvrene	19.99	202	265273	2.75	ng	99
25) Benzo(a)anthracene	21.72	228	273015	11.90	ng	# 93
26) Chrysene	21.77	228	269848	13.36	ng	# 84
27) Indeno(1,2,3-cd)pyrene	26.63	276	298225	3.31	ng	100
29) Benzo(b)fluoranthene	23.40	252	64618	2.75	ng	98
30) Benzo(k)fluoranthene	23.45	252	68288	2.84	ng	95
31) Benzo(a)pvrene	24.02	252	64488	2.95	ng	94
32) Dibenzo(a,h)anthracene	26.67	278	61075	3.01	ng	# 93
33) Benzo(g,h,i)perylene	27.42	276	254192	2.98	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080216\
Data File : BE092111.D
Acq On : 2 Aug 2016 14:41
Operator : UM/SJ
Sample : SSTDIC3.2
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Aug 02 15:46:37 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-PAH-BE080216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 02 13:43:25 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.18 min Scan# 480

Delta R.T. 0.01 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Instrument :

BNA_E

ClientSampleId :


Tot Ion:152 Resp: 12006

Ion Ratio Lower Upper

152 100

150 139.2 106.0 159.0




115 42.0 36.9 55.3



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 2.55 ng

RT: 3.44 min Scan# 74

Delta R.T. -0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41


Tgt Ion: 88 Resp: 4418

Ion Ratio Lower Upper

88 100

58 85.2 62.1 93.1



43 36.8 27.2 40.8




Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0






#3




n-Nitrosodimethylamine
Concen: 3.00 ng
RT: 3.78 min Scan# 103
Delta R.T. -0.03 min
Lab File: BE092111.D
Acq: 2 Aug 2016 14:41

Instrument :
BNA_E
ClientSampleId :

Tot Ion: 42 Resp: 6617
Ion Ratio Lower Upper
42 100
74 101.7 82.8 124.2
44 6.4 9.0 13.4#




Abundance Ion 42.00 (41.70 to 42.70): BE0
Ion 74.00 (73.70 to 74.70): BE0
Ion 44.00 (43.70 to 44.70): BE0





#4

2-Fluorophenol
Concen: 2.30 ng
RT: 5.70 min Scan# 267
Delta R.T. 0.00 min
Lab File: BE092111.D
Acq: 2 Aug 2016 14:41

Tgt Ion: 112 Resp: 9010
Ion Ratio Lower Upper
112 100
64 67.0 55.9 83.9
63 36.3 31.1 46.7



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BE0
Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 2.67 ng

RT: 7.34 min Scan# 409

Delta R.T. -0.01 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Instrument :

BNA_E

ClientSampleId :


Tot Ion: 99 Resp: 12567

Ion Ratio Lower Upper




99 100

42 25.9 22.4 33.6

71 35.3 28.9 43.3



Abundance Ion 99.00 (98.70 to 99.70): BE0
Ion 42.00 (41.70 to 42.70): BE0
Ion 71.00 (70.70 to 71.70): BE0



#6

Naphthalene-d8

Concen: 5.00 ng

RT: 10.99 min Scan# 704

Delta R.T. -0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Tgt Ion: 136 Resp: 56943


Ion Ratio Lower Upper

136 100



137 11.7 9.1 13.7

54 4.5 2.7 4.1#

68 3.7 2.2 3.4#



Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BE0
Ion 68.00 (67.70 to 68.70): BE0





#7

Nitrobenzene-d5

Concen: 2.59 ng

RT: 9.34 min Scan# 580

Delta R.T. -0.01 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Instrument :

BNA_E

ClientSampleId :





Tot Ion: 82 Resp: 12651

Ion Ratio Lower Upper

82 100

128 33.1 33.0 49.4

54 61.4 42.6 63.8



#8

Naphthalene

Concen: 2.79 ng

RT: 11.03 min Scan# 707

Delta R.T. -0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41




Tgt Ion: 128 Resp: 36625


Ion Ratio Lower Upper

128 100

129 11.1 11.1 16.7#

127 12.9 11.3 16.9





#9

2-Methylnaphthalene

Concen: 2.76 ng

RT: 12.65 min Scan# 834

Delta R.T. -0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Instrument :

BNA_E

ClientSampleId :


Tot Ion:142 Resp: 24262

Ion Ratio Lower Upper




142 100

141 81.4 65.1 97.7

115 30.1 27.0 40.4



Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 115.00 (114.70 to 115.70): E



#10

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.85 min Scan# 999

Delta R.T. 0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41


Tgt Ion:164 Resp: 32281

Ion Ratio Lower Upper



164 100


162 89.9 69.5 104.3

160 35.6 27.6 41.4



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E









#11
 2,4,6-Tribromophenol
 Concen: 1.53 ng
 RT: 16.33 min Scan# 1107
 Delta R.T. 0.00 min
 Lab File: BE092111.D
 Acq: 2 Aug 2016 14:41




Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:330 Resp: 3573
 Ion Ratio Lower Upper
 330 100
 332 98.0 74.2 111.4
 141 44.1 139.8 209.8#



#12
 2-Fluorobiphenyl
 Concen: 2.39 ng
 RT: 13.46 min Scan# 904
 Delta R.T. -0.00 min
 Lab File: BE092111.D
 Acq: 2 Aug 2016 14:41

Tgt Ion:172 Resp: 29914
 Ion Ratio Lower Upper
 172 100
 171 32.7 30.3 45.5
 170 21.1 21.6 32.4#









#13
Acenaphthylene
Concen: 2.38 ng
RT: 14.54 min Scan# 981
Delta R.T. -0.02 min
Lab File: BE092111.D
Acq: 2 Aug 2016 14:41




Instrument :
BNA_E
ClientSampleId :


Tot Ion:152 Resp: 170345
Ion Ratio Lower Upper
152 100
151 5.0 4.0 6.0
153 13.0 9.6 14.4



#14
Acenaphthene
Concen: 2.58 ng
RT: 14.90 min Scan# 1002
Delta R.T. 0.00 min
Lab File: BE092111.D
Acq: 2 Aug 2016 14:41

Tgt Ion:154 Resp: 25524
Ion Ratio Lower Upper
154 100
153 452.0 345.1 517.7
152 225.3 0.0 0.0#





#15

Fluorene

Concen: 2.21 ng

RT: 15.88 min Scan# 1068

Delta R.T. -0.01 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Instrument :

BNA_E

ClientSampleId :





Tot Ion:166 Resp: 32794

Ion Ratio Lower Upper

166 100

165 99.4 78.6 118.0

167 13.5 11.1 16.7



#16

Phenanthrene-d10

Concen: 5.00 ng

RT: 17.58 min Scan# 1172

Delta R.T. -0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41




Tgt Ion:188 Resp: 66637

Ion Ratio Lower Upper

188 100

94 2.3 3.9 5.9#

80 1.8 3.4 5.0#









#17
Hexachlorobenzene
Concen: 4.30 ng
RT: 16.89 min Scan# 1142
Delta R.T. -0.00 min
Lab File: BE092111.D
Acq: 2 Aug 2016 14:41




Instrument :
BNA_E
ClientSampleId :


Tot Ion:284 Resp: 9576
Ion Ratio Lower Upper
284 100
142 0.0 33.3 49.9#
249 29.3 25.4 38.0



#18
Pentachlorophenol
Concen: 9.49 ng
RT: 17.24 min Scan# 1157
Delta R.T. -0.00 min
Lab File: BE092111.D
Acq: 2 Aug 2016 14:41

Tgt Ion:266 Resp: 3877
Ion Ratio Lower Upper
266 100
268 67.3 62.7 94.1
264 59.8 63.5 95.3#





#19

Phenanthrene

Concen: 3.14 ng

RT: 17.60 min Scan# 1173

Delta R.T. -0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41


Instrument :

BNA_E

ClientSampleId :

Tot Ion:178 Resp: 56195




Ion	Ratio	Lower	Upper
178	100		
176	20.1	15.4	23.0
179	15.0	12.7	19.1



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#20

Anthracene

Concen: 4.30 ng

RT: 17.70 min Scan# 1177


Delta R.T. -0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Tgt Ion:178 Resp: 57158



Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.2	22.8
179	14.8	12.2	18.4



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#21

Fluoranthene

Concen: 4.70 ng

RT: 19.63 min Scan# 1284

Delta R.T. 0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Instrument :

BNA_E

ClientSampleId :


Tot Ion:202 Resp: 245537

Ion Ratio Lower Upper

202 100

101 2.8 2.2 3.2




203 18.0 13.0 19.4



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#22

Chrysene-d12

Concen: 5.00 ng

RT: 21.75 min Scan# 1407

Delta R.T. 0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41


Tgt Ion:240 Resp: 84891

Ion Ratio Lower Upper

240 100

120 1.3 0.6 1.0#



236 22.6 18.2 27.4




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#23

Pvrene

Concen: 2.75 ng

RT: 19.99 min Scan# 1305

Delta R.T. 0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Instrument :

BNA_E

ClientSampleId :


Tot Ion:202 Resp: 265273

Ion Ratio Lower Upper

202 100

200 5.3 4.4 6.6




203 17.5 13.4 20.0



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#24

Terphenyl-d14

Concen: 2.92 ng

RT: 20.19 min Scan# 1317

Delta R.T. -0.02 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41


Tgt Ion:244 Resp: 37622

Ion Ratio Lower Upper

244 100

212 9.6 9.1 13.7



122 12.3 11.0 16.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#25

Benzo(a)anthracene

Concen: 11.90 ng

RT: 21.72 min Scan# 1405

Delta R.T. 0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Instrument :

BNA_E

ClientSampleId :


Tot Ion:228 Resp: 273015

Ion Ratio Lower Upper

228 100

226 30.1 22.1 33.1




229 14.7 15.1 22.7#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#26

Chrysene

Concen: 13.36 ng

RT: 21.77 min Scan# 1408

Delta R.T. 0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41


Tgt Ion:228 Resp: 269848

Ion Ratio Lower Upper

228 100

226 36.6 22.3 33.5#



229 13.9 16.9 25.3#



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#27
 Indeno(1,2,3-cd)pyrene
 Concen: 3.31 ng
 RT: 26.63 min Scan# 1760
 Delta R.T. -0.02 min
 Lab File: BE092111.D
 Acq: 2 Aug 2016 14:41




Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:276 Resp: 298225
 Ion Ratio Lower Upper
 276 100
 138 26.4 21.2 31.8
 277 24.8 19.8 29.8



#28
 Perylene-d12
 Concen: 5.00 ng
 RT: 24.14 min Scan# 1591
 Delta R.T. 0.00 min
 Lab File: BE092111.D
 Acq: 2 Aug 2016 14:41

Tgt Ion:264 Resp: 84733
 Ion Ratio Lower Upper
 264 100
 260 25.2 20.8 31.2
 265 21.7 16.6 25.0





#29

Benzo(b)fluoranthene

Concen: 2.75 ng

RT: 23.40 min Scan# 1527

Delta R.T. 0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Instrument :

BNA_E

ClientSampleId :


Tot Ion:252 Resp: 64618

Ion Ratio Lower Upper




252 100

253 21.3 17.4 26.0

125 10.6 10.2 15.2



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#30

Benzo(k)fluoranthene

Concen: 2.84 ng

RT: 23.45 min Scan# 1531

Delta R.T. -0.01 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41


Tgt Ion:252 Resp: 68288

Ion Ratio Lower Upper



252 100


253 21.0 19.4 29.0

125 10.6 8.9 13.3



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#31

Benzo(a)pyrene

Concen: 2.95 ng

RT: 24.02 min Scan# 1581

Delta R.T. 0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Instrument :

BNA_E

ClientSampleId :


Tot Ion:252 Resp: 64488

Ion Ratio Lower Upper

252 100

253 21.3 19.8 29.8




125 11.9 10.4 15.6



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#32

Dibenzo(a,h)anthracene

Concen: 3.01 ng

RT: 26.67 min Scan# 1762

Delta R.T. -0.00 min

Lab File: BE092111.D

Acq: 2 Aug 2016 14:41


Tgt Ion:278 Resp: 61075

Ion Ratio Lower Upper

278 100

139 14.8 16.7 25.1#



279 24.3 20.3 30.5




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#33

Benzo(a,h,i)perylene

Concen: 2.98 ng

RT: 27.42 min Scan# 1806

Delta R.T. -0.00 min


Lab File: BE092111.D

Acq: 2 Aug 2016 14:41

Instrument :

BNA_E

ClientSampleId :



Tot Ion: 276 Resp: 254192

Ion Ratio Lower Upper

276 100

277 24.8 19.5 29.3

138 20.9 20.8 31.2

Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

