

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080216\
 Data File : BE092134.D
 Acq On : 3 Aug 2016 5:36
 Operator : UM/SJ
 Sample : H4205-05
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 GW-38-6.5-15.0

Quant Time: Aug 03 07:56:49 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	11116	5.00	ng	0.00
6) Naphthalene-d8	10.99	136	47923	5.00	ng	0.00
10) Acenaphthene-d10	14.85	164	25002	5.00	ng	0.00
16) Phenanthrene-d10	17.56	188	52906	5.00	ng	-0.02
22) Chrysene-d12	21.75	240	62659	5.00	ng	0.00
28) Perylene-d12	24.14	264	66359	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.70	112	21579	7.56	ng	0.00
5) Phenol-d6	7.34	99	18478	4.86	ng	-0.01
7) Nitrobenzene-d5	9.34	82	24629	6.51	ng	-0.01
11) 2,4,6-Tribromophenol	16.33	330	11241	13.85	ng	0.00
12) 2-Fluorobiphenyl	13.46	172	52541	6.63	ng	0.00
24) Terphenyl-d14	20.19	244	70122	7.79	ng	-0.02
Target Compounds						
20) Anthracene	17.70	178	668	0.05	ng	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080216\
Data File : BE092134.D
Acq On : 3 Aug 2016 5:36
Operator : UM/SJ
Sample : H4205-05
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
GW-38-6.5-15.0

Quant Time: Aug 03 07:56:49 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-PAH-BE080216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 02 16:59:37 2016
Response via : Initial Calibration

