

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080216\
 Data File : BE092116.D
 Acq On : 2 Aug 2016 18:13
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 03 03:53:18 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.01
2	1,4-Dioxane	0.680	0.725	-6.6	91	0.00
3	n-Nitrosodimethylamine	0.942	0.716	24.0	83	0.00
4 S	2-Fluorophenol	1.284	1.067	16.9	83	0.01
5 S	Phenol-d6	1.711	1.459	14.7	82	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
7 S	Nitrobenzene-d5	0.395	0.364	7.8	86	0.00
8	Naphthalene	1.159	1.102	4.9	87	0.00
9	2-Methylnaphthalene	0.743	0.673	9.4	86	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
11 S	2,4,6-Tribromophenol	0.162	0.128	21.0	83	0.00
12 S	2-Fluorobiphenyl	1.584	1.507	4.9	87	0.00
13	Acenaphthylene	8.821	8.126	7.9	86	0.00
14	Acenaphthene	1.406	1.362	3.1	87	0.00
15	Fluorene	1.783	1.634	8.4	85	0.00
16 I	Phenanthrene-d10	1.000	1.000	0.0	87	-0.02
17	Hexachlorobenzene	0.245	0.248	-1.2	87	0.00
18	Pentachlorophenol	0.068	0.048	29.4#	75	0.00
19	Phenanthrene	1.460	1.388	4.9	87	0.00
20	Anthracene	1.331	1.193	10.4	83	0.00
21	Fluoranthene	6.116	5.654	7.6	84	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
23	Pyrene	5.228	4.832	7.6	85	0.00
24 S	Terphenyl-d14	0.718	0.668	7.0	86	0.00
25	Benzo(a)anthracene	5.496	5.035	8.4	83	0.00
26	Chrysene	5.505	5.286	4.0	86	0.00
27	Indeno(1,2,3-cd)pyrene	5.978	5.673	5.1	87	0.00
28 I	Perylene-d12	1.000	1.000	0.0	89	0.00
29	Benzo(b)fluoranthene	1.368	1.187	13.2	75	0.00
30	Benzo(k)fluoranthene	1.369	1.299	5.1	97	0.00
31 C	Benzo(a)pyrene	1.257	1.177	6.4	86	0.00
32	Dibenzo(a,h)anthracene	1.200	1.094	8.8	86	0.00
33	Benzo(g,h,i)perylene	5.122	4.824	5.8	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0