

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080216\
 Data File : BE092116.D
 Acq On : 2 Aug 2016 18:13
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 03 03:53:18 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	88	0.01
2	1,4-Dioxane	0.400	0.426	-6.5	91	0.00
3	n-Nitrosodimethylamine	0.400	0.388	3.0	83	0.00
4 S	2-Fluorophenol	0.400	0.332	17.0	83	0.01
5 S	Phenol-d6	0.400	0.341	14.8	82	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	88	0.00
7 S	Nitrobenzene-d5	0.400	0.370	7.5	86	0.00
8	Naphthalene	0.400	0.399	0.3	87	0.00
9	2-Methylnaphthalene	0.400	0.391	2.3	86	0.00
10 I	Acenaphthene-d10	5.000	5.000	0.0	88	0.00
11 S	2,4,6-Tribromophenol	0.400	0.316	21.0	83	0.00
12 S	2-Fluorobiphenyl	0.400	0.380	5.0	87	0.00
13	Acenaphthylene	0.400	0.368	8.0	86	0.00
14	Acenaphthene	0.400	0.388	3.0	87	0.00
15	Fluorene	0.400	0.367	8.3	85	0.00
16 I	Phenanthrene-d10	5.000	5.000	0.0	87	-0.02
17	Hexachlorobenzene	0.400	0.405	-1.3	87	0.00
18	Pentachlorophenol	0.400	0.396	1.0	75	0.00
19	Phenanthrene	0.400	0.372	7.0	87	0.00
20	Anthracene	0.400	0.359	10.3	83	0.00
21	Fluoranthene	0.400	0.370	7.5	84	0.00
22 I	Chrysene-d12	5.000	5.000	0.0	85	0.00
23	Pyrene	0.400	0.370	7.5	85	0.00
24 S	Terphenyl-d14	0.400	0.372	7.0	86	0.00
25	Benzo(a)anthracene	0.400	0.366	8.5	83	0.00
26	Chrysene	0.400	0.384	4.0	86	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.380	5.0	87	0.00
28 I	Perylene-d12	5.000	5.000	0.0	89	0.00
29	Benzo(b)fluoranthene	0.400	0.347	13.3	75	0.00
30	Benzo(k)fluoranthene	0.400	0.379	5.3	97	0.00
31 C	Benzo(a)pyrene	0.400	0.374	6.5	86	0.00
32	Dibenzo(a,h)anthracene	0.400	0.365	8.8	86	0.00
33	Benzo(g,h,i)perylene	0.400	0.377	5.8	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0