

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080216\
 Data File : BE092132.D
 Acq On : 3 Aug 2016 4:19
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 03 07:14:45 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	97	0.00
2	1,4-Dioxane	0.400	0.396	1.0	93	0.00
3	n-Nitrosodimethylamine	0.400	0.411	-2.7	99	0.00
4 S	2-Fluorophenol	0.400	0.360	10.0	99	0.00
5 S	Phenol-d6	0.400	0.377	5.8	99	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	100	-0.01
7 S	Nitrobenzene-d5	0.400	0.379	5.3	99	0.00
8	Naphthalene	0.400	0.400	0.0	98	0.00
9	2-Methylnaphthalene	0.400	0.396	1.0	99	0.00
10 I	Acenaphthene-d10	5.000	5.000	0.0	101	-0.02
11 S	2,4,6-Tribromophenol	0.400	0.344	14.0	104	0.00
12 S	2-Fluorobiphenyl	0.400	0.384	4.0	101	0.00
13	Acenaphthylene	0.400	0.378	5.5	101	-0.02
14	Acenaphthene	0.400	0.392	2.0	102	0.00
15	Fluorene	0.400	0.374	6.5	100	0.00
16 I	Phenanthrene-d10	5.000	5.000	0.0	104	-0.02
17	Hexachlorobenzene	0.400	0.383	4.3	99	0.00
18	Pentachlorophenol	0.400	0.367	8.3	77	0.00
19	Phenanthrene	0.400	0.365	8.8	102	0.00
20	Anthracene	0.400	0.369	7.8	103	0.00
21	Fluoranthene	0.400	0.376	6.0	102	0.00
22 I	Chrysene-d12	5.000	5.000	0.0	100	0.00
23	Pyrene	0.400	0.389	2.8	106	0.00
24 S	Terphenyl-d14	0.400	0.393	1.8	108	-0.02
25	Benzo(a)anthracene	0.400	0.384	4.0	103	0.00
26	Chrysene	0.400	0.377	5.8	100	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.374	6.5	102	-0.02
28 I	Perylene-d12	5.000	5.000	0.0	102	0.00
29	Benzo(b)fluoranthene	0.400	0.372	7.0	93	0.00
30	Benzo(k)fluoranthene	0.400	0.375	6.3	110	-0.01
31 C	Benzo(a)pyrene	0.400	0.387	3.3	102	0.00
32	Dibenzo(a,h)anthracene	0.400	0.375	6.3	101	0.00
33	Benzo(g,h,i)perylene	0.400	0.372	7.0	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0