

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080217\
 Data File : BE093629.D
 Acq On : 2 Aug 2017 16:50
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 02 17:37:24 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 02 10:58:31 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.648	0.635	2.0	106	0.00
3	n-Nitrosodimethylamine	0.607	0.551	9.2	104	0.02
4 S	2-Fluorophenol	1.192	1.246	-4.5	117	0.00
5 S	Phenol-d6	1.794	1.817	-1.3	113	0.00
6	bis(2-Chloroethyl)ether	1.508	1.485	1.5	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.305	0.314	-3.0	116	0.00
9	Naphthalene	4.630	4.626	0.1	105	0.00
10	Hexachlorobutadiene	0.172	0.170	1.2	106	-0.02
11 SURR	2-Methylnaphthalene-d10	0.632	0.641	-1.4	107	0.00
12	2-Methylnaphthalene	0.766	0.763	0.4	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.093	0.088	5.4	123	0.00
15 S	2-Fluorobiphenyl	1.601	1.579	1.4	104	-0.01
16	Acenaphthylene	1.912	1.874	2.0	108	0.00
17	Acenaphthene	1.349	1.342	0.5	107	0.00
18	Fluorene	1.803	1.750	2.9	102	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
20	4-Bromophenyl-phenylether	0.126	0.117	7.1	94	0.00
21	Hexachlorobenzene	0.129	0.131	-1.6	106	0.00
22	Pentachlorophenol	0.055	0.070	-27.3#	150	0.00
23	Phenanthrene	0.880	0.841	4.4	105	0.00
24	Anthracene	1.203	1.257	-4.5	111	0.00
25 SURR	Fluoranthene-d10	4.541	4.508	0.7	103	0.00
26	Fluoranthene	1.401	1.375	1.9	102	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	-0.01
28	Pyrene	1.292	1.241	3.9	102	0.00
29 S	Terphenyl-d14	0.719	0.684	4.9	100	0.00
30	Benzo(a)anthracene	1.176	1.183	-0.6	110	0.00
31	Chrysene	1.311	1.275	2.7	103	0.00
32	Bis(2-ethylhexyl)phthalate	1.603	1.859	-16.0	143	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.249	-5.2	110	0.00
34 I	Perylene-d12	1.000	1.000	0.0	110	0.00
35	Benzo(b)fluoranthene	1.427	1.394	2.3	110	0.00
36	Benzo(k)fluoranthene	1.459	1.384	5.1	108	0.00
37 C	Benzo(a)pyrene	1.293	1.297	-0.3	114	0.00
38	Dibenzo(a,h)anthracene	1.248	1.217	2.5	109	0.00
39	Benzo(g,h,i)perylene	1.266	1.230	2.8	109	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080217\
Data File : BE093629.D
Acq On : 2 Aug 2017 16:50
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 02 17:37:24 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 02 10:58:31 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		