

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080217\
 Data File : BE093631.D
 Acq On : 2 Aug 2017 18:05
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 03 02:55:28 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 02 10:58:31 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	-0.01
2	1,4-Dioxane	0.648	0.708	-9.3	114	0.00
3	n-Nitrosodimethylamine	0.607	0.562	7.4	102	0.00
4 S	2-Fluorophenol	1.192	1.243	-4.3	113	0.00
5 S	Phenol-d6	1.794	1.831	-2.1	110	-0.01
6	bis(2-Chloroethyl)ether	1.508	1.469	2.6	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.305	0.300	1.6	108	0.00
9	Naphthalene	4.630	4.531	2.1	101	0.00
10	Hexachlorobutadiene	0.172	0.168	2.3	103	-0.02
11 SURR	2-Methylnaphthalene-d10	0.632	0.633	-0.2	103	0.00
12	2-Methylnaphthalene	0.766	0.760	0.8	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.093	0.098	-5.4	133	0.00
15 S	2-Fluorobiphenyl	1.601	1.590	0.7	102	-0.02
16	Acenaphthylene	1.912	1.919	-0.4	107	0.00
17	Acenaphthene	1.349	1.342	0.5	105	0.00
18	Fluorene	1.803	1.832	-1.6	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	126	0.00
20	4-Bromophenyl-phenylether	0.126	0.104	17.5	94	0.00
21	Hexachlorobenzene	0.129	0.118	8.5	106	-0.03
22	Pentachlorophenol	0.055	0.060	-9.1	145	0.00
23	Phenanthrene	0.880	0.785	10.8	110	0.00
24	Anthracene	1.203	1.230	-2.2	122	0.00
25 SURR	Fluoranthene-d10	4.541	4.143	8.8	105	0.00
26	Fluoranthene	1.401	1.289	8.0	107	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	107	-0.01
28	Pyrene	1.292	1.265	2.1	105	0.00
29 S	Terphenyl-d14	0.719	0.704	2.1	104	0.00
30	Benzo(a)anthracene	1.176	1.215	-3.3	114	0.00
31	Chrysene	1.311	1.285	2.0	105	0.00
32	Bis(2-ethylhexyl)phthalate	1.603	1.908	-19.0	148	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.269	-6.9	113	0.00
34 I	Perylene-d12	1.000	1.000	0.0	115	0.00
35	Benzo(b)fluoranthene	1.427	1.327	7.0	110	0.00
36	Benzo(k)fluoranthene	1.459	1.353	7.3	110	0.00
37 C	Benzo(a)pyrene	1.293	1.263	2.3	116	0.00
38	Dibenzo(a,h)anthracene	1.248	1.200	3.8	112	0.00
39	Benzo(g,h,i)perylene	1.266	1.174	7.3	109	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080217\
Data File : BE093631.D
Acq On : 2 Aug 2017 18:05
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 03 02:55:28 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 02 10:58:31 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		