

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080217\
 Data File : BE093629.D
 Acq On : 2 Aug 2017 16:50
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

Sohil
 8/3/2017 6:06:54 PM


Quant Time: Aug 02 17:37:24 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 02 10:58:31 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	2419	0.40	ng	0.00
7) Naphthalene-d8	10.71	136	11551	0.40	ng	0.00
13) Acenaphthene-d10	14.53	164	6771	0.40	ng	0.00
19) Phenanthrene-d10	17.24	188	16517	0.40	ng	0.00
27) Chrysene-d12	21.40	240	18787	0.40	ng	-0.01
34) Perylene-d12	23.91	264	16449	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.52	112	3014	0.42	ng	0.00
5) Phenol-d6	7.09	99	4396	0.41	ng	0.00
8) Nitrobenzene-d5	9.06	82	3631	0.41	ng	0.00
11) 2-Methylnaphthalene-d10	12.29	152	7399	0.41	ng	0.00
14) 2,4,6-Tribromophenol	16.00	330	595	0.40	ng	0.00
15) 2-Fluorobiphenyl	13.15	172	10694	0.39	ng	-0.01
25) Fluoranthene-d10	19.27	212	74464	0.40	ng	0.00
29) Terphenyl-d14	19.86	244	12858	0.38	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.44	88	1536	0.392	ng	95
3) n-Nitrosodimethylamine	3.75	42	1333	0.363	ng	82
6) bis(2-Chloroethyl)ether	7.34	93	3593	0.394	ng	# 97
9) Naphthalene	10.76	128	53437	0.400	ng	99
10) Hexachlorobutadiene	11.04	225	1964	0.396	ng	99
12) 2-Methylnaphthalene	12.36	142	8816	0.399	ng	97
16) Acenaphthylene	14.24	152	12691	0.392	ng	# 87
17) Acenaphthene	14.59	154	9084	0.398	ng	99
18) Fluorene	15.56	166	11851	0.388	ng	# 87
20) 4-Bromophenyl-phenylether	16.45	248	1934	0.380	ng	# 94
21) Hexachlorobenzene	16.58	284	2163	0.426	ng	# 100
22) Pentachlorophenol	16.91	266	1152	0.508	ng	96
23) Phenanthrene	17.27	178	13889m	0.382	ng	
24) Anthracene	17.37	178	20763m	0.418	ng	
26) Fluoranthene	19.30	202	22711	0.393	ng	99
28) Pyrene	19.67	202	23324	0.384	ng	# 99
30) Benzo(a)anthracene	21.39	228	22224	0.402	ng	96
31) Chrysene	21.45	228	23945	0.389	ng	96
32) Bis(2-ethylhexyl)phthalate	21.30	149	34926	0.464	ng	98
33) Indeno(1,2,3-cd)pyrene	26.55	276	23459	0.421	ng	94
35) Benzo(b)fluoranthene	23.14	252	22933	0.391	ng	# 95
36) Benzo(k)fluoranthene	23.19	252	22773	0.380	ng	# 93
37) Benzo(a)pyrene	23.80	252	21334	0.401	ng	95
38) Dibenzo(a,h)anthracene	26.57	278	20020	0.390	ng	# 89
39) Benzo(g,h,i)perylene	27.36	276	20231	0.389	ng	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1


1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.92 min Scan# 402
Delta R.T. -0.01 min
Lab File: BE093629.D
Acq: 2 Aug 2017 16:50

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion	Ratio	Lower	Upper
152	100		
150	160.6	125.1	187.7
115	73.3	55.7	83.5

Manual Integrations
APPROVED

Sohil
8/3/2017 6:06:54 PM




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 0.392 ng
RT: 3.44 min Scan# 46
Delta R.T. 0.00 min
Lab File: BE093629.D
Acq: 2 Aug 2017 16:50

Tgt Ion	Ratio	Lower	Upper
88	100		
58	82.1	62.2	93.4
43	32.0	23.2	34.8



Abundance Ion 88.00 (87.70 to 88.70): BE0
Ion 58.00 (57.70 to 58.70): BE0
Ion 43.00 (42.70 to 43.70): BE0





#3





n-Nitrosodimethylamine
 Concen: 0.363 ng
 RT: 3.75 min Scan# 71
 Delta R.T. 0.02 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:	42	Resp:	1333
Ion	Ratio	Lower	Upper
42	100		
74	145.0	99.1	148.7
44	7.6	7.4	11.2

Manual Integrations
 APPROVED




Sohil
 8/3/2017 6:06:54 PM






#4

2-Fluorophenol
 Concen: 0.418 ng
 RT: 5.52 min Scan# 211
 Delta R.T. 0.00 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Tgt Ion:	112	Resp:	3014
Ion	Ratio	Lower	Upper
112	100		
64	72.3	56.4	84.6
63	140.5	29.3	43.9#





#5

Phenol-d6

Concen: 0.405 ng

RT: 7.09 min Scan# 336

Delta R.T. -0.01 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Tgt Ion:	99	Resp:	4396
Ion Ratio	Lower	Upper	
99	100		
42	19.7	0.0	0.0#
71	33.3	0.0	0.0#

Instrument :





BNA_E

ClientSampleId :

SSTDCCC0.4EC

Manual Integrations
APPROVED

Sohil
8/3/2017 6:06:54 PM



#6

bis(2-Chloroethyl)ether

Concen: 0.394 ng


RT: 7.34 min Scan# 356


Delta R.T. 0.00 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Tgt Ion:	93	Resp:	3593
Ion Ratio	Lower	Upper	
93	100		
63	349.9	0.0	0.0#
95	33.2	25.4	38.0





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.71 min Scan# 577

Delta R.T. -0.00 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion	Ratio	Resp	Lower	Upper
136	100	11551		
137	11.8	9.8	14.8	
54	32.8	10.3	15.5#	
68	6.7	9.0	13.4#	

Manual Integrations
APPROVED

Sohil
8/3/2017 6:06:54 PM



#8

Nitrobenzene-d5

Concen: 0.412 ng




RT: 9.06 min Scan# 479


Delta R.T. -0.00 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	3631		
128	104.4	17.0	25.4#	
54	245.2	53.8	80.6#	





#9

Naphthalene

Concen: 0.400 ng

RT: 10.76 min Scan# 580

Delta R.T. -0.00 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	128	Resp:	53437
Ion	Ratio	Lower	Upper
128	100		
129	2.9	2.4	3.6
127	3.3	2.4	3.6

Manual Integrations
APPROVED

Sohil
8/3/2017 6:06:54 PM



#10

Hexachlorobutadiene

Concen: 0.396 ng




RT: 11.04 min Scan# 597


Delta R.T. -0.02 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Tgt Ion:	225	Resp:	1964
Ion	Ratio	Lower	Upper
225	100		
223	63.0	49.7	74.5
227	64.2	50.3	75.5









#11
 2-Methylnaphthalene-d10
 Concen: 0.405 ng
 RT: 12.29 min Scan# 825
 Delta R.T. -0.00 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:152 Resp: 7399
 Ion Ratio Lower Upper
 152 100
 151 19.0 15.1 22.7




Manual Integrations
 APPROVED


Sohil
 8/3/2017 6:06:54 PM



#12
 2-Methylnaphthalene
 Concen: 0.399 ng
 RT: 12.36 min Scan# 843
 Delta R.T. -0.01 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Tgt Ion:142 Resp: 8816
 Ion Ratio Lower Upper
 142 100
 141 89.3 72.6 108.8
 115 37.0 32.2 48.2









#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.53 min Scan# 1112
 Delta R.T. 0.00 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTCCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
164	100		
162	101.2	84.6	127.0
160	45.2	41.8	62.6




Manual Integrations
 APPROVED


Sohil
 8/3/2017 6:06:54 PM



#14
 2,4,6-Tribromophenol
 Concen: 0.400 ng
 RT: 16.00 min Scan# 1205
 Delta R.T. 0.00 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Tgt Ion	Ratio	Lower	Upper
330	100		
332	89.7	77.7	116.5
141	51.9	56.2	84.4#






#15
2-Fluorobiphenyl
Concen: 0.395 ng
RT: 13.15 min Scan# 1019
Delta R.T. -0.01 min
Lab File: BE093629.D
Acq: 2 Aug 2017 16:50

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.1	29.8	44.6
170	24.5	20.7	31.1

Manual Integrations
APPROVED

Sohil
8/3/2017 6:06:54 PM




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#16
Acenaphthylene
Concen: 0.392 ng
RT: 14.24 min Scan# 1093
Delta R.T. 0.00 min
Lab File: BE093629.D
Acq: 2 Aug 2017 16:50

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.8	4.0	6.0#
153	13.2	10.3	15.5



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E





#17

Acenaphthene

Concen: 0.398 ng

RT: 14.59 min Scan# 1116

Delta R.T. 0.00 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4EC




Tgt Ion:	154	Resp:	9084
Ion	Ratio	Lower	Upper
154	100		
153	112.0	91.2	136.8
152	55.5	44.8	67.2

Manual Integrations
APPROVED

Sohil
8/3/2017 6:06:54 PM



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E



#18

Fluorene

Concen: 0.388 ng


RT: 15.56 min Scan# 1182

Delta R.T. 0.00 min



Lab File: BE093629.D




Acq: 2 Aug 2017 16:50

Tgt Ion:	166	Resp:	11851
Ion	Ratio	Lower	Upper
166	100		
165	99.1	78.6	117.8
167	53.4	11.1	16.7#



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.24 min Scan# 1243
Delta R.T. -0.00 min
Lab File: BE093629.D
Acq: 2 Aug 2017 16:50

Tgt Ion:	188	Resp:	16517
Ion Ratio	Lower	Upper	
188	100		
94	15.7	11.3	16.9
80	16.5	11.7	17.5

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Manual Integrations
APPROVED


Sohil
8/3/2017 6:06:54 PM



#20
4-Bromophenyl-phenylether
Concen: 0.380 ng
RT: 16.45 min Scan# 1219
Delta R.T. 0.00 min
Lab File: BE093629.D
Acq: 2 Aug 2017 16:50

Tgt Ion:	248	Resp:	1934
Ion Ratio	Lower	Upper	
248	100		
250	166.9	128.0	192.0
141	18.8	12.0	18.0#





#21

Hexachlorobenzene

Concen: 0.426 ng

RT: 16.58 min Scan# 1223

Delta R.T. -0.00 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4EC




Tgt Ion:	284	Resp:	2163
Ion Ratio	Lower	Upper	
284	100		
142	0.0	0.0	0.0
249	0.0	0.0	0.0

Manual Integrations
APPROVED

Sohil
8/3/2017 6:06:54 PM



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E



#22

Pentachlorophenol

Concen: 0.508 ng


RT: 16.91 min Scan# 1233

Delta R.T. -0.00 min



Lab File: BE093629.D


Acq: 2 Aug 2017 16:50

Tgt Ion:	266	Resp:	1152
Ion Ratio	Lower	Upper	
266	100		
264	65.5	56.0	84.0
268	62.9	52.0	78.0



Abundance Ion 266.00 (265.70 to 266.70): E
Ion 264.00 (263.70 to 264.70): E
Ion 268.00 (267.70 to 268.70): E





#23

Phenanthrene

Concen: 0.382 ng m

RT: 17.27 min Scan# 1244

Delta R.T. -0.00 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	178	Resp:	13889
Ion Ratio	Lower	Upper	
178	100		
176	0.0	15.0	22.4#
179	0.0	12.7	19.1#

Manual Integrations
APPROVED

Sohil
8/3/2017 6:06:54 PM



#24

Anthracene

Concen: 0.418 ng m




RT: 17.37 min Scan# 1247


Delta R.T. -0.00 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Tgt Ion:	178	Resp:	20763
Ion Ratio	Lower	Upper	
178	100		
176	18.9	14.6	22.0
179	23.0	12.0	18.0#





#25

Fluoranthene-d10

Concen: 0.397 ng

RT: 19.27 min Scan# 1432

Delta R.T. -0.00 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	212	Resp:	74464
Ion Ratio	Lower	Upper	
212	100		
106	3.5	2.4	3.6
104	2.0	1.6	2.4

Manual Integrations
APPROVED

Sohil
8/3/2017 6:06:54 PM



#26

Fluoranthene

Concen: 0.393 ng




RT: 19.30 min Scan# 1436


Delta R.T. -0.00 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Tgt Ion:	202	Resp:	22711
Ion Ratio	Lower	Upper	
202	100		
101	12.1	9.6	14.4
203	17.1	14.4	21.6









#27
 Chrysene-d12
 Concen: 0.400 ng
 RT: 21.40 min Scan# 1654
 Delta R.T. -0.01 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:240 Resp: 18787
 Ion Ratio Lower Upper
 240 100
 120 11.8 4.6 7.0#
 236 27.0 20.8 31.2




Manual Integrations
 APPROVED


Sohil
 8/3/2017 6:06:54 PM



#28
 Pyrene
 Concen: 0.384 ng
 RT: 19.67 min Scan# 1485
 Delta R.T. -0.00 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Tgt Ion:202 Resp: 23324
 Ion Ratio Lower Upper
 202 100
 200 20.7 0.0 0.0#
 203 17.5 13.7 20.5









#29
 Terphenyl-d14
 Concen: 0.381 ng
 RT: 19.86 min Scan# 1511
 Delta R.T. -0.01 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
244	100		
212	35.0	10.6	16.0
122	14.0	15.4	23.0




Manual Integrations
 APPROVED


Sohil
 8/3/2017 6:06:54 PM



#30
 Benzo(a)anthracene
 Concen: 0.402 ng
 RT: 21.39 min Scan# 1653
 Delta R.T. 0.00 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Tgt Ion	Ratio	Lower	Upper
228	100		
226	25.4	19.7	29.5
229	20.4	19.2	28.8





#31

Chrysene

Concen: 0.389 ng

RT: 21.45 min Scan# 1658

Delta R.T. 0.00 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	228	Resp:	23945
Ion	Ratio	Lower	Upper
228	100		
226	28.6	21.5	32.3
229	20.7	18.6	28.0

Manual Integrations
APPROVED

Sohil
8/3/2017 6:06:54 PM



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.464 ng




RT: 21.30 min Scan# 1645


Delta R.T. -0.01 min

Lab File: BE093629.D

Acq: 2 Aug 2017 16:50

Tgt Ion:	149	Resp:	34926
Ion	Ratio	Lower	Upper
149	100		
167	6.6	4.8	7.2
279	0.9	0.8	1.2









#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.421 ng
 RT: 26.55 min Scan# 2697
 Delta R.T. -0.00 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
276	100		
138	28.5	27.4	41.2
277	25.0	20.1	30.1




Manual Integrations
 APPROVED


Sohil
 8/3/2017 6:06:54 PM



#34
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.91 min Scan# 2121
 Delta R.T. -0.01 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.0	21.0	31.4
265	26.0	24.6	36.8









#35
Benzo(b)fluoranthene
Concen: 0.391 ng
RT: 23.14 min Scan# 1953
Delta R.T. -0.00 min
Lab File: BE093629.D
Acq: 2 Aug 2017 16:50

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.2	18.5	27.7
125	12.1	13.4	20.0




Manual Integrations
APPROVED


Sohil
8/3/2017 6:06:54 PM



#36
Benzo(k)fluoranthene
Concen: 0.380 ng
RT: 23.19 min Scan# 1964
Delta R.T. -0.00 min
Lab File: BE093629.D
Acq: 2 Aug 2017 16:50

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.4	20.3	30.5
125	11.3	12.2	18.4









#37
Benzo(a)pyrene
Concen: 0.401 ng
RT: 23.80 min Scan# 2097
Delta R.T. -0.00 min
Lab File: BE093629.D
Acq: 2 Aug 2017 16:50

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.3	19.9	29.9
125	15.0	14.6	21.8




Manual Integrations
APPROVED


Sohil
8/3/2017 6:06:54 PM



#38
Dibenzo(a,h)anthracene
Concen: 0.390 ng
RT: 26.57 min Scan# 2702
Delta R.T. -0.00 min
Lab File: BE093629.D
Acq: 2 Aug 2017 16:50

Tgt Ion	Ratio	Lower	Upper
278	100		
139	18.8	21.4	32.0#
279	24.2	22.2	33.2





#39
 Benzo(g,h,i)perylene
 Concen: 0.389 ng
 RT: 27.36 min Scan# 2875
 Delta R.T. -0.01 min
 Lab File: BE093629.D
 Acq: 2 Aug 2017 16:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:	276	Resp:	20231
Ion Ratio	Lower	Upper	
276	100		
277	24.5	21.2	31.8
138	24.8	24.5	36.7

Manual Integrations
 APPROVED

Sohil
 8/3/2017 6:06:54 PM

