


Data Path : Z:\HPCHEM1\BNA_E\Data\BE080217\
 Data File : BE093630.D
 Acq On : 2 Aug 2017 17:29
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080117.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Aug 02 10:58:31 2017



AutoFind: Scans 2570, 2571, 2572; Background Corrected with Scan 2562

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.8	26971	PASS
68	69	0.00	2	0.7	206	PASS
69	198	0.00	100	41.4	28069	PASS
70	69	0.00	2	0.5	144	PASS
127	198	10	80	48.2	32691	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	67769	PASS
199	198	5	9	6.6	4485	PASS
275	198	10	60	24.7	16728	PASS
365	198	1	100	2.7	1838	PASS
441	442	0.01	24	16.8	7823	PASS
442	198	50	100	68.6	46501	PASS
443	442	15	24	19.0	8840	PASS