

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080315\
 Data File : BE090325.D
 Acq On : 4 Aug 2015 00:51
 Operator : TP/UM
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE080315

Quant Time: Aug 04 02:31:40 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 04 02:31:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	136	0.00
2	1,4-Dioxane	0.796	0.890	-11.8	150#	0.00
3	n-Nitrosodimethylamine	0.845	0.799	5.4	127	0.00
4 S	2-Fluorophenol	0.956	0.967	-1.2	141	0.00
5 S	Phenol-d6	1.175	1.118	4.9	139	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	138	0.00
7 S	Nitrobenzene-d5	0.295	0.286	3.1	139	0.00
8	Nitrobenzene	0.305	0.284	6.9	136	0.00
9	Naphthalene	1.074	1.070	0.4	138	0.00
10	Hexachlorobutadiene	0.144	0.147	-2.1	138	0.00
11	2-Methylnaphthalene	0.665	0.695	-4.5	145	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	142	0.00
13 S	2,4,6-Tribromophenol	0.067	0.059	11.9	174#	0.00
14 S	2-Fluorobiphenyl	1.449	1.485	-2.5	144	0.00
15	Acenaphthylene	8.730	8.853	-1.4	145	0.00
16	Acenaphthene	1.262	1.247	1.2	141	0.00
17	Fluorene	1.533	1.503	2.0	149	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	142	0.00
19	4-Bromophenyl-phenylether	0.188	0.195	-3.7	143	0.00
20	Hexachlorobenzene	0.710	0.731	-3.0	143	0.00
21	Pentachlorophenol	0.038	0.028	26.3#	163#	0.00
22	Phenanthrene	1.139	1.139	0.0	148	0.00
23	Anthracene	1.018	1.003	1.5	148	0.00
24	Fluoranthene	1.146	1.184	-3.3	153#	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	152#	0.00
26	Pyrene	1.352	1.405	-3.9	151#	0.00
27 S	Terphenyl-d14	0.936	0.983	-5.0	153#	0.00
28	Benzo(a)anthracene	0.951	0.981	-3.2	176#	0.00
29	Chrysene	1.320	1.432	-8.5	158#	0.00
30	Bis(2-ethylhexyl)phthalate	0.335	0.262	21.8	168#	0.00
31	Indeno(1,2,3-cd)pyrene	0.548	0.716	-30.7#	277#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	154#	0.00
33	Benzo(b)fluoranthene	0.676	0.851	-25.9#	226#	0.00
34	Benzo(k)fluoranthene	1.357	1.737	-28.0#	184#	0.00
35 C	Benzo(a)pyrene	0.673	0.840	-24.8#	218#	0.00
36	Dibenzo(a,h)anthracene	0.431	0.534	-23.9	272#	0.00
37	Benzo(g,h,i)perylene	0.744	0.950	-27.7#	203#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1