

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080315\
 Data File : BE090325.D
 Acq On : 4 Aug 2015 00:51
 Operator : TP/UM
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE080315

Quant Time: Aug 04 02:31:40 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 04 02:31:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	136	0.00
2	1,4-Dioxane	1.000	1.371	-37.1#	150	0.00
3	n-Nitrosodimethylamine	1.000	0.946	5.4	127	0.00
4 S	2-Fluorophenol	1.000	1.012	-1.2	141	0.00
5 S	Phenol-d6	1.000	0.952	4.8	139	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	138	0.00
7 S	Nitrobenzene-d5	1.000	0.970	3.0	139	0.00
8	Nitrobenzene	1.000	0.933	6.7	136	0.00
9	Naphthalene	1.000	0.995	0.5	138	0.00
10	Hexachlorobutadiene	1.000	1.024	-2.4	138	0.00
11	2-Methylnaphthalene	1.000	1.045	-4.5	145	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	142	0.00
13 S	2,4,6-Tribromophenol	1.000	1.019	-1.9	174	0.00
14 S	2-Fluorobiphenyl	1.000	1.025	-2.5	144	0.00
15	Acenaphthylene	1.000	1.014	-1.4	145	0.00
16	Acenaphthene	1.000	0.988	1.2	141	0.00
17	Fluorene	1.000	0.980	2.0	149	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	142	0.00
19	4-Bromophenyl-phenylether	1.000	1.037	-3.7	143	0.00
20	Hexachlorobenzene	1.000	1.029	-2.9	143	0.00
21	Pentachlorophenol	1.000	0.952	4.8	163	0.00
22	Phenanthrene	1.000	1.000	0.0	148	0.00
23	Anthracene	1.000	0.986	1.4	148	0.00
24	Fluoranthene	1.000	1.033	-3.3	153	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	152	0.00
26	Pyrene	1.000	1.040	-4.0	151	0.00
27 S	Terphenyl-d14	1.000	1.051	-5.1	153	0.00
28	Benzo(a)anthracene	1.000	1.032	-3.2	176	0.00
29	Chrysene	1.000	1.218	-21.8	158	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.937	6.3	168	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.362	-36.2#	277	0.00
32 I	Perylene-d12	5.000	5.000	0.0	154	0.00
33	Benzo(b)fluoranthene	1.000	1.172	-17.2	226	0.00
34	Benzo(k)fluoranthene	1.000	1.280	-28.0#	184	0.00
35 C	Benzo(a)pyrene	1.000	1.190	-19.0	218	0.00
36	Dibenzo(a,h)anthracene	1.000	1.285	-28.5#	272	0.00
37	Benzo(g,h,i)perylene	1.000	1.216	-21.6	203	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0