

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080315\
 Data File : BE090333.D
 Acq On : 4 Aug 2015 5:40
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 04 07:51:04 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 04 02:31:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	60	0.00
2	1,4-Dioxane	0.796	0.868	-9.0	65	-0.01
3	n-Nitrosodimethylamine	0.845	0.921	-9.0	65	0.00
4 S	2-Fluorophenol	0.956	1.151	-20.4	75	0.00
5 S	Phenol-d6	1.175	1.507	-28.3#	83	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	61	0.00
7 S	Nitrobenzene-d5	0.295	0.300	-1.7	65	0.00
8	Nitrobenzene	0.305	0.298	2.3	64	0.00
9	Naphthalene	1.074	1.194	-11.2	69	0.00
10	Hexachlorobutadiene	0.144	0.156	-8.3	65	0.00
11	2-Methylnaphthalene	0.665	0.741	-11.4	69	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	55	0.00
13 S	2,4,6-Tribromophenol	0.067	0.085	-26.9#	98	0.00
14 S	2-Fluorobiphenyl	1.449	1.812	-25.1#	68	0.00
15	Acenaphthylene	8.730	10.656	-22.1	67	0.00
16	Acenaphthene	1.262	1.479	-17.2	65	0.00
17	Fluorene	1.533	1.766	-15.2	67	0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	49#	0.02
19	4-Bromophenyl-phenylether	0.188	0.257	-36.7#	65	0.00
20	Hexachlorobenzene	0.710	0.945	-33.1#	64	0.00
21	Pentachlorophenol	0.038	0.044	-15.8	87	0.00
22	Phenanthrene	1.139	1.365	-19.8	61	0.00
23	Anthracene	1.018	1.235	-21.3	63	0.00
24	Fluoranthene	1.146	1.374	-19.9	61	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	45#	0.00
26	Pyrene	1.352	1.868	-38.2#	60	0.00
27 S	Terphenyl-d14	0.936	1.252	-33.8#	58	0.00
28	Benzo(a)anthracene	0.951	1.079	-13.5	58	0.00
29	Chrysene	1.320	1.942	-47.1#	64	0.00
30	Bis(2-ethylhexyl)phthalate	0.335	0.363	-8.4	69	0.00
31	Indeno(1,2,3-cd)pyrene	0.548	0.985	-79.7#	113	0.00
32 I	Perylene-d12	1.000	1.000	0.0	48#	0.00
33	Benzo(b)fluoranthene	0.676	1.116	-65.1#	92	0.00
34	Benzo(k)fluoranthene	1.357	2.151	-58.5#	71	0.00
35 C	Benzo(a)pyrene	0.673	1.111	-65.1#	90	0.00
36	Dibenzo(a,h)anthracene	0.431	0.706	-63.8#	112	0.00
37	Benzo(g,h,i)perylene	0.744	1.403	-88.6#	93	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1