

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080315\
 Data File : BE090333.D
 Acq On : 4 Aug 2015 5:40
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 04 07:51:04 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 04 02:31:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	60	0.00
2	1,4-Dioxane	1.000	1.332	-33.2#	65	-0.01
3	n-Nitrosodimethylamine	1.000	1.090	-9.0	65	0.00
4 S	2-Fluorophenol	1.000	1.204	-20.4	75	0.00
5 S	Phenol-d6	1.000	1.283	-28.3#	83	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	61	0.00
7 S	Nitrobenzene-d5	1.000	1.015	-1.5	65	0.00
8	Nitrobenzene	1.000	0.978	2.2	64	0.00
9	Naphthalene	1.000	1.111	-11.1	69	0.00
10	Hexachlorobutadiene	1.000	1.085	-8.5	65	0.00
11	2-Methylnaphthalene	1.000	1.114	-11.4	69	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	55	0.00
13 S	2,4,6-Tribromophenol	1.000	1.391	-39.1#	98	0.00
14 S	2-Fluorobiphenyl	1.000	1.250	-25.0	68	0.00
15	Acenaphthylene	1.000	1.221	-22.1	67	0.00
16	Acenaphthene	1.000	1.172	-17.2	65	0.00
17	Fluorene	1.000	1.151	-15.1	67	0.02
18 I	Phenanthrene-d10	5.000	5.000	0.0	49	0.02
19	4-Bromophenyl-phenylether	1.000	1.366	-36.6#	65	0.00
20	Hexachlorobenzene	1.000	1.330	-33.0#	64	0.00
21	Pentachlorophenol	1.000	1.358	-35.8#	87	0.00
22	Phenanthrene	1.000	1.199	-19.9	61	0.00
23	Anthracene	1.000	1.214	-21.4	63	0.00
24	Fluoranthene	1.000	1.199	-19.9	61	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	45	0.00
26	Pyrene	1.000	1.382	-38.2#	60	0.00
27 S	Terphenyl-d14	1.000	1.338	-33.8#	58	0.00
28	Benzo(a)anthracene	1.000	1.134	-13.4	58	0.00
29	Chrysene	1.000	1.674	-67.4#	64	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.215	-21.5	69	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.785	-78.5#	113	0.00
32 I	Perylene-d12	5.000	5.000	0.0	48	0.00
33	Benzo(b)fluoranthene	1.000	1.427	-42.7#	92	0.00
34	Benzo(k)fluoranthene	1.000	1.585	-58.5#	71	0.00
35 C	Benzo(a)pyrene	1.000	1.476	-47.6#	90	0.00
36	Dibenzo(a,h)anthracene	1.000	1.605	-60.5#	112	0.00
37	Benzo(g,h,i)perylene	1.000	1.680	-68.0#	93	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1