

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080315\
 Data File : BE090335.D
 Acq On : 4 Aug 2015 7:37
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 04 08:13:38 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 04 02:31:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	46#	0.00
2	1,4-Dioxane	0.796	0.799	-0.4	46#	0.00
3	n-Nitrosodimethylamine	0.845	0.894	-5.8	48#	0.00
4 S	2-Fluorophenol	0.956	1.011	-5.8	50	0.00
5 S	Phenol-d6	1.175	1.049	10.7	44#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	45#	0.00
7 S	Nitrobenzene-d5	0.295	0.288	2.4	46#	0.00
8	Nitrobenzene	0.305	0.273	10.5	43#	0.00
9	Naphthalene	1.074	1.077	-0.3	46#	0.00
10	Hexachlorobutadiene	0.144	0.143	0.7	44#	0.00
11	2-Methylnaphthalene	0.665	0.585	12.0	40#	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	38#	0.00
13 S	2,4,6-Tribromophenol	0.067	0.054	19.4	42#	0.00
14 S	2-Fluorobiphenyl	1.449	1.545	-6.6	40#	0.00
15	Acenaphthylene	8.730	8.916	-2.1	39#	0.00
16	Acenaphthene	1.262	1.238	1.9	37#	0.00
17	Fluorene	1.533	1.444	5.8	38#	0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	34#	0.02
19	4-Bromophenyl-phenylether	0.188	0.214	-13.8	37#	0.00
20	Hexachlorobenzene	0.710	0.827	-16.5	38#	0.00
21	Pentachlorophenol	0.038	0.447	-1076.3#	613#	0.00
22	Phenanthrene	1.139	1.112	2.4	34#	0.00
23	Anthracene	1.018	0.984	3.3	35#	0.00
24	Fluoranthene	1.146	1.091	4.8	33#	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	30#	0.00
26	Pyrene	1.352	1.590	-17.6	34#	0.00
27 S	Terphenyl-d14	0.936	1.062	-13.5	33#	0.00
28	Benzo(a)anthracene	0.951	0.885	6.9	31#	0.00
29	Chrysene	1.320	1.556	-17.9	34#	0.00
30	Bis(2-ethylhexyl)phthalate	0.335	0.284	15.2	36#	0.00
31	Indeno(1,2,3-cd)pyrene	0.548	0.702	-28.1#	54	0.00
32 I	Perylene-d12	1.000	1.000	0.0	30#	0.00
33	Benzo(b)fluoranthene	0.676	0.767	-13.5	39#	0.00
34	Benzo(k)fluoranthene	1.357	1.617	-19.2	33#	0.00
35 C	Benzo(a)pyrene	0.673	0.858	-27.5#	43#	0.00
36	Dibenzo(a,h)anthracene	0.431	0.615	-42.7#	61	0.00
37	Benzo(g,h,i)perylene	0.744	1.062	-42.7#	44#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1