

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080315\
 Data File : BE090335.D
 Acq On : 4 Aug 2015 7:37
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 04 08:13:38 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 04 02:31:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	46	0.00
2	1,4-Dioxane	1.000	1.207	-20.7	46	0.00
3	n-Nitrosodimethylamine	1.000	1.058	-5.8	48	0.00
4 S	2-Fluorophenol	1.000	1.058	-5.8	50	0.00
5 S	Phenol-d6	1.000	0.893	10.7	44	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	45	0.00
7 S	Nitrobenzene-d5	1.000	0.974	2.6	46	0.00
8	Nitrobenzene	1.000	0.895	10.5	43	0.00
9	Naphthalene	1.000	1.002	-0.2	46	0.00
10	Hexachlorobutadiene	1.000	0.992	0.8	44	0.00
11	2-Methylnaphthalene	1.000	0.879	12.1	40	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	38	0.00
13 S	2,4,6-Tribromophenol	1.000	0.942	5.8	42	0.00
14 S	2-Fluorobiphenyl	1.000	1.066	-6.6	40	0.00
15	Acenaphthylene	1.000	1.021	-2.1	39	0.00
16	Acenaphthene	1.000	0.981	1.9	37	0.00
17	Fluorene	1.000	0.942	5.8	38	0.02
18 I	Phenanthrene-d10	5.000	5.000	0.0	34	0.02
19	4-Bromophenyl-phenylether	1.000	1.140	-14.0	37	0.00
20	Hexachlorobenzene	1.000	1.163	-16.3	38	0.00
21	Pentachlorophenol	1.000	7.459	-645.9#	613	0.00
22	Phenanthrene	1.000	0.977	2.3	34	0.00
23	Anthracene	1.000	0.967	3.3	35	0.00
24	Fluoranthene	1.000	0.952	4.8	33	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	30	0.00
26	Pyrene	1.000	1.176	-17.6	34	0.00
27 S	Terphenyl-d14	1.000	1.134	-13.4	33	0.00
28	Benzo(a)anthracene	1.000	0.930	7.0	31	0.00
29	Chrysene	1.000	1.329	-32.9#	34	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.997	0.3	36	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.340	-34.0#	54	0.00
32 I	Perylene-d12	5.000	5.000	0.0	30	0.00
33	Benzo(b)fluoranthene	1.000	1.091	-9.1	39	0.00
34	Benzo(k)fluoranthene	1.000	1.191	-19.1	33	0.00
35 C	Benzo(a)pyrene	1.000	1.210	-21.0#	43	0.00
36	Dibenzo(a,h)anthracene	1.000	1.438	-43.8#	61	0.00
37	Benzo(g,h,i)perylene	1.000	1.331	-33.1#	44	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1