

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080315\
 Data File : BE090342.D
 Acq On : 4 Aug 2015 11:50
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 04 13:22:23 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 04 02:31:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	47#	0.00
2	1,4-Dioxane	0.796	1.242	-56.0#	73	0.00
3	n-Nitrosodimethylamine	0.845	0.881	-4.3	49#	0.00
4 S	2-Fluorophenol	0.956	1.349	-41.1#	69	0.00
5 S	Phenol-d6	1.175	1.535	-30.6#	67	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	53	0.00
7 S	Nitrobenzene-d5	0.295	0.431	-46.1#	80	0.00
8	Nitrobenzene	0.305	0.245	19.7	45#	0.00
9	Naphthalene	1.074	0.885	17.6	44#	0.00
10	Hexachlorobutadiene	0.144	0.121	16.0	43#	0.00
11	2-Methylnaphthalene	0.665	0.502	24.5	40#	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	44#	0.00
13 S	2,4,6-Tribromophenol	0.067	0.132	-97.0#	121	0.00
14 S	2-Fluorobiphenyl	1.449	2.564	-76.9#	77	0.00
15	Acenaphthylene	8.730	7.161	18.0	36#	0.00
16	Acenaphthene	1.262	0.971	23.1	34#	0.00
17	Fluorene	1.533	1.122	26.8#	34#	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	39#	0.00
19	4-Bromophenyl-phenylether	0.188	0.164	12.8	33#	0.00
20	Hexachlorobenzene	0.710	0.619	12.8	33#	0.00
21	Pentachlorophenol	0.038	0.034	10.5	53	0.00
22	Phenanthrene	1.139	0.885	22.3	31#	0.00
23	Anthracene	1.018	0.756	25.7#	30#	0.00
24	Fluoranthene	1.146	0.869	24.2	31#	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	36#	0.00
26	Pyrene	1.352	1.170	13.5	30#	0.00
27 S	Terphenyl-d14	0.936	1.794	-91.7#	66	0.00
28	Benzo(a)anthracene	0.951	0.662	30.4#	28#	0.00
29	Chrysene	1.320	1.045	20.8	27#	0.00
30	Bis(2-ethylhexyl)phthalate	0.335	0.283	15.5	43#	0.00
31	Indeno(1,2,3-cd)pyrene	0.548	0.602	-9.9	56	0.00
32 I	Perylene-d12	1.000	1.000	0.0	40#	0.00
33	Benzo(b)fluoranthene	0.676	0.617	8.7	43#	0.00
34	Benzo(k)fluoranthene	1.357	0.973	28.3#	27#	0.00
35 C	Benzo(a)pyrene	0.673	0.608	9.7	41#	0.00
36	Dibenzo(a,h)anthracene	0.431	0.351	18.6	47#	0.00
37	Benzo(g,h,i)perylene	0.744	0.736	1.1	41#	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0