

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080315\
 Data File : BE090321.D
 Acq On : 3 Aug 2015 22:25
 Operator : TP/UM
 Sample : SSTDICCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICCC001

Manual Integrations
 APPROVED

MMDadoda
 8/4/2015 3:26:49 PM

Quant Time: Aug 04 02:15:08 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 04 01:38:34 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	6236	5.00	ng	0.00
6) Naphthalene-d8	10.36	136	24930	5.00	ng	0.00
12) Acenaphthene-d10	14.21	164	12621	5.00	ng	0.00
18) Phenanthrene-d10	16.95	188	26451	5.00	ng	0.00
25) Chrysene-d12	21.11	240	20918	5.00	ng	0.00
32) Perylene-d12	23.43	264	16784	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.22	112	1158	0.96	ng	0.00
5) Phenol-d6	6.78	99	1359	0.96	ng	0.00
7) Nitrobenzene-d5	8.74	82	1418	0.93	ng	0.00
13) 2,4,6-Tribromophenol	15.73	330	121	0.81	ng	0.00
14) 2-Fluorobiphenyl	12.83	172	3696	1.02	ng	0.00
27) Terphenyl-d14	19.58	244	4095	1.03	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.21	88	1003	1.09	ng	100
3) n-Nitrosodimethylamine	3.51	42	1068	1.00	ng	100
8) Nitrobenzene	8.78	77	1433	0.94	ng	100
9) Naphthalene	10.40	128	5317	1.00	ng	100
10) Hexachlorobutadiene	10.69	225	735	1.01	ng	100
11) 2-Methylnaphthalene	12.02	142	3283	0.99	ng	100
15) Acenaphthylene	13.93	152	21917	1.00	ng	100
16) Acenaphthene	14.27	154	3161	1.00	ng	100
17) Fluorene	15.26	166	3623	0.95	ng	100
19) 4-Bromophenyl-phenylether	16.16	248	1022	1.02	ng	100
20) Hexachlorobenzene	16.27	284	3839	1.02	ng	100
21) Pentachlorophenol	16.63	266	130	0.75	ng	100
22) Phenanthrene	17.00	178	5776	0.96	ng	100
23) Anthracene	17.08	178	5076	0.95	ng	100
24) Fluoranthene	19.00	202	5810	0.97	ng	100
26) Pyrene	19.36	202	5937	0.97	ng	100
28) Benzo(a)anthracene	21.10	228	3554	1.07	ng	100
29) Chrysene	21.15	228	5777m	0.93	ng	
30) Bis(2-ethylhexyl)phthalate	21.04	149	994	0.78	ng	100
31) Indeno(1,2,3-cd)pyrene	25.83	276	1645	1.22	ng	100
33) Benzo(b)fluoranthene	22.72	252	1948	1.28	ng	100
34) Benzo(k)fluoranthene	22.76	252	4880m	1.21	ng	
35) Benzo(a)pyrene	23.32	252	1996	1.14	ng	100
36) Dibenzo(a,h)anthracene	25.86	278	1017	1.17	ng	100
37) Benzo(g,h,i)perylene	26.56	276	2421	1.30	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080315\
Data File : BE090321.D
Acq On : 3 Aug 2015 22:25
Operator : TP/UM
Sample : SSTDICCC001
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICCC001

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:49 PM

Quant Time: Aug 04 02:15:08 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 04 01:38:34 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.60 min Scan# 848

Delta R.T. 0.00 min

Lab File: BE090321.D


Acq: 3 Aug 2015 22:25

Instrument :

BNA_E

ClientSampleId :

SSTDICCC001



Tgt Ion:152 Resp: 6236

Ion Ratio Lower Upper




152 100

150 155.2 124.2 186.2

115 63.0 50.4 75.6

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:49 PM



#2

1,4-Dioxane

Concen: 1.09 ng

RT: 3.21 min Scan# 91

Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25




Tgt Ion: 88 Resp: 1003


Ion Ratio Lower Upper

88 100

58 62.7 50.2 75.2

43 31.3 25.0 37.6





#3





n-Nitrosodimethylamine
 Concen: 1.00 ng
 RT: 3.51 min Scan# 134
 Delta R.T. 0.00 min
 Lab File: BE090321.D
 Acq: 3 Aug 2015 22:25

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC001

Tgt Ion	Ratio	Lower	Upper
42	100		
74	91.5	73.2	109.8
44	4.7	3.8	5.6

Manual Integrations
 APPROVED




MMDadoda
 8/4/2015 3:26:49 PM






#4

2-Fluorophenol
 Concen: 0.96 ng
 RT: 5.22 min Scan# 387
 Delta R.T. 0.00 min
 Lab File: BE090321.D
 Acq: 3 Aug 2015 22:25

Tgt Ion	Ratio	Lower	Upper
112	100		
64	60.8	49.0	73.4
63	34.6	27.7	41.5





#5

Phenol-d6

Concen: 0.96 ng

RT: 6.78 min Scan# 669

Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Tgt Ion:	99	Resp:	1359
Ion	Ratio	Lower	Upper
99	100		
42	27.7	22.2	33.2
71	36.5	29.2	43.8

Instrument :





BNA_E

ClientSampleId :

SSTDICCC001

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:49 PM



#6

Naphthalene-d8

Concen: 5.00 ng


RT: 10.36 min Scan# 1373


Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Tgt Ion:	136	Resp:	24930
Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.9	13.3
54	7.5	6.0	9.0
68	4.9	3.9	5.9





#7

Nitrobenzene-d5

Concen: 0.93 ng

RT: 8.74 min Scan# 1095

Delta R.T. 0.00 min

Lab File: BE090321.D


Acq: 3 Aug 2015 22:25

Instrument :

BNA_E

ClientSampleId :




SSTDICCC001



Tgt Ion:	82	Resp:	1418
Ion	Ratio	Lower	Upper
82	100		
128	38.5	30.8	46.2
54	59.4	47.5	71.3

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:49 PM



#8

Nitrobenzene

Concen: 0.94 ng




RT: 8.78 min Scan# 1104


Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Tgt Ion:	77	Resp:	1433
Ion	Ratio	Lower	Upper
77	100		
123	37.0	29.6	44.4
65	13.0	10.4	15.6





#9

Naphthalene

Concen: 1.00 ng

RT: 10.40 min Scan# 1379

Delta R.T. 0.00 min

Lab File: BE090321.D


Acq: 3 Aug 2015 22:25

Instrument :

BNA_E

ClientSampleId :




SSTDICCC001



Tgt Ion:	128	Resp:	5317
Ion	Ratio	Lower	Upper
128	100		
129	12.6	10.1	15.1
127	14.1	11.3	16.9

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:49 PM



#10

Hexachlorobutadiene

Concen: 1.01 ng




RT: 10.69 min Scan# 1418


Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Tgt Ion:	225	Resp:	735
Ion	Ratio	Lower	Upper
225	100		
223	63.0	50.4	75.6
227	62.0	49.6	74.4





#11

2-Methylnaphthalene

Concen: 0.99 ng

RT: 12.02 min Scan# 1652

Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Instrument :

BNA_E





ClientSampleId :

SSTDICCC001

Tgt Ion:	142	Resp:	3283
Ion Ratio	Lower	Upper	
142	100		
141	93.3	74.6	112.0
115	40.8	32.6	49.0

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:49 PM



#12

Acenaphthene-d10

Concen: 5.00 ng




RT: 14.21 min Scan# 2077


Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25


Tgt Ion:	164	Resp:	12621
Ion Ratio	Lower	Upper	
164	100		
162	103.4	82.7	124.1
160	53.8	43.0	64.6





#13
 2,4,6-Tribromophenol
 Concen: 0.81 ng
 RT: 15.73 min Scan# 2208
 Delta R.T. 0.00 min
 Lab File: BE090321.D
 Acq: 3 Aug 2015 22:25




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC001



Tgt Ion: 330 Resp: 121
 Ion Ratio Lower Upper
 330 100
 332 99.2 79.4 119.0
 141 307.4 245.9 368.9




Manual Integrations
 APPROVED


MMDadoda
 8/4/2015 3:26:49 PM



#14
 2-Fluorobiphenyl
 Concen: 1.02 ng
 RT: 12.83 min Scan# 1831
 Delta R.T. 0.00 min
 Lab File: BE090321.D
 Acq: 3 Aug 2015 22:25

Tgt Ion: 172 Resp: 3696
 Ion Ratio Lower Upper
 172 100
 171 36.9 29.5 44.3
 170 26.6 21.3 31.9





#15

Acenaphthylene

Concen: 1.00 ng

RT: 13.93 min Scan# 2033

Delta R.T. 0.00 min

Lab File: BE090321.D


Acq: 3 Aug 2015 22:25

Instrument :

BNA_E

ClientSampleId :




SSTDICCC001



Tgt Ion	Ratio	Lower	Upper
152	100		
151	5.0	4.0	6.0
153	12.8	10.2	15.4

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:49 PM



#16

Acenaphthene

Concen: 1.00 ng




RT: 14.27 min Scan# 2086


Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Tgt Ion	Ratio	Lower	Upper
154	100		
153	472.5	378.0	567.0
152	265.1	212.7	319.1









#17
Fluorene
 Concen: 0.95 ng
 RT: 15.26 min Scan# 2181
 Delta R.T. 0.00 min
 Lab File: BE090321.D
 Acq: 3 Aug 2015 22:25

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC001

Tgt Ion	Ratio	Lower	Upper
166	100		
165	103.6	82.6	123.8
167	14.9	11.8	17.8




Manual Integrations
 APPROVED


MMDadoda
 8/4/2015 3:26:49 PM



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 16.95 min Scan# 2278
 Delta R.T. 0.00 min
 Lab File: BE090321.D
 Acq: 3 Aug 2015 22:25

Tgt Ion	Ratio	Lower	Upper
188	100		
94	11.9	9.5	14.3
80	12.5	10.0	15.0





#19

4-Bromophenyl-phenylether

Concen: 1.02 ng

RT: 16.16 min Scan# 2233

Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Instrument :

BNA_E


ClientSampleId :

SSTDICCC001




Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:49 PM

Tgt Ion:	248	Resp:	1022
Ion Ratio	Lower	Upper	
248	100		
250	92.5	74.0	111.0
141	323.6	258.9	388.3



Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E



#20

Hexachlorobenzene

Concen: 1.02 ng


RT: 16.27 min Scan# 2239

Delta R.T. 0.00 min



Lab File: BE090321.D


Acq: 3 Aug 2015 22:25

Tgt Ion:	284	Resp:	3839
Ion Ratio	Lower	Upper	
284	100		
142	55.3	44.2	66.4
249	33.1	26.4	39.6



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E









#21
 Pentachlorophenol
 Concen: 0.75 ng
 RT: 16.63 min Scan# 2260
 Delta R.T. 0.00 min
 Lab File: BE090321.D
 Acq: 3 Aug 2015 22:25

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC001

Tgt Ion	Ratio	Lower	Upper
266	100		
268	82.7	66.2	99.2
264	86.4	69.1	103.7




Manual Integrations
 APPROVED


MMDadoda
 8/4/2015 3:26:49 PM



#22
 Phenanthrene
 Concen: 0.96 ng
 RT: 17.00 min Scan# 2281
 Delta R.T. 0.00 min
 Lab File: BE090321.D
 Acq: 3 Aug 2015 22:25


Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.5	23.3
179	15.7	12.6	18.8





#23
 Anthracene
 Concen: 0.95 ng
 RT: 17.08 min Scan# 2286
 Delta R.T. 0.00 min
 Lab File: BE090321.D
 Acq: 3 Aug 2015 22:25




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC001



Tgt Ion:178 Resp: 5076
 Ion Ratio Lower Upper
 178 100
 176 18.9 15.1 22.7
 179 15.1 12.1 18.1




Manual Integrations
 APPROVED


MMDadoda
 8/4/2015 3:26:49 PM



#24
 Fluoranthene
 Concen: 0.97 ng
 RT: 19.00 min Scan# 2509
 Delta R.T. 0.00 min
 Lab File: BE090321.D
 Acq: 3 Aug 2015 22:25

Tgt Ion:202 Resp: 5810
 Ion Ratio Lower Upper
 202 100
 101 10.4 8.3 12.5
 203 17.3 13.8 20.8









#25
Chrysene-d12
Concen: 5.00 ng
RT: 21.11 min Scan# 2867
Delta R.T. 0.00 min
Lab File: BE090321.D
Acq: 3 Aug 2015 22:25

Instrument :
BNA_E
ClientSampleId :
SSTDICCC001

Tgt Ion	Ratio	Lower	Upper
240	100		
120	9.4	7.5	11.3
236	26.2	21.0	31.4




Manual Integrations
APPROVED


MMDadoda
8/4/2015 3:26:49 PM



#26
Pyrene
Concen: 0.97 ng
RT: 19.36 min Scan# 2577
Delta R.T. 0.00 min
Lab File: BE090321.D
Acq: 3 Aug 2015 22:25

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.7	16.6	24.8
203	17.2	13.8	20.6





#27

Terphenyl-d14

Concen: 1.03 ng

RT: 19.58 min Scan# 2618

Delta R.T. 0.00 min

Lab File: BE090321.D


Acq: 3 Aug 2015 22:25

Instrument :

BNA_E

ClientSampleId :




SSTDICCC001



Tgt Ion:	244	Resp:	4095
Ion Ratio	Lower	Upper	
244	100		
212	9.3	7.4	11.2
122	13.0	10.4	15.6

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:49 PM



#28

Benzo(a)anthracene

Concen: 1.07 ng




RT: 21.10 min Scan# 2865


Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Tgt Ion:	228	Resp:	3554
Ion Ratio	Lower	Upper	
228	100		
226	28.7	23.0	34.4
229	21.8	17.4	26.2





#29

Chrysene

Concen: 0.93 ng m

RT: 21.15 min Scan# 2872

Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Instrument :

BNA_E





ClientSampleId :

SSTDICCC001

Tgt Ion:	228	Resp:	5777
Ion Ratio	Lower	Upper	
228	100		
226	31.1	24.9	37.3
229	20.8	16.6	25.0

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:49 PM



#30

Bis(2-ethylhexyl)phthalate

Concen: 0.78 ng




RT: 21.04 min Scan# 2856


Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Tgt Ion:	149	Resp:	994
Ion Ratio	Lower	Upper	
149	100		
167	20.7	16.6	24.8
279	3.4	2.7	4.1





#31

Indeno(1,2,3-cd)pyrene

Concen: 1.22 ng

RT: 25.83 min Scan# 3839

Delta R.T. 0.00 min

Lab File: BE090321.D


Acq: 3 Aug 2015 22:25

Instrument :

BNA_E

ClientSampleId :

SSTDICCC001



Tgt Ion: 276 Resp: 1645

Ion Ratio Lower Upper


276 100

138 13.3 10.6 16.0

277 18.3 14.6 22.0

Manual Integrations
APPROVED



MMDadoda
8/4/2015 3:26:49 PM



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#32

Perylene-d12

Concen: 5.00 ng

RT: 23.43 min Scan# 3314

Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25


Tgt Ion: 264 Resp: 16784

Ion Ratio Lower Upper

264 100

260 25.1 20.1 30.1



265 23.3 18.6 28.0




Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E





#33

Benzo(b)fluoranthene

Concen: 1.28 ng

RT: 22.72 min Scan# 3160

Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Instrument :

BNA_E





ClientSampleId :

SSTDICCC001

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:49 PM

Tgt Ion:	252	Resp:	1948
Ion Ratio	Lower	Upper	
252	100		
253	26.9	21.5	32.3
125	14.8	11.8	17.8



#34

Benzo(k)fluoranthene

Concen: 1.21 ng m




RT: 22.76 min Scan# 3169


Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Tgt Ion:	252	Resp:	4880
Ion Ratio	Lower	Upper	
252	100		
253	24.9	19.9	29.9
125	14.0	11.2	16.8









#35
Benzo(a)pyrene
Concen: 1.14 ng
RT: 23.32 min Scan# 3291
Delta R.T. 0.00 min
Lab File: BE090321.D
Acq: 3 Aug 2015 22:25

Instrument :
BNA_E
ClientSampleId :
SSTDICCC001

Tgt Ion	Ratio	Lower	Upper
252	100		
253	28.7	23.0	34.4
125	17.7	14.2	21.2




Manual Integrations
APPROVED




MMDadoda
8/4/2015 3:26:49 PM



#36
Dibenzo(a,h)anthracene
Concen: 1.17 ng
RT: 25.86 min Scan# 3845
Delta R.T. 0.00 min
Lab File: BE090321.D
Acq: 3 Aug 2015 22:25

Tgt Ion	Ratio	Lower	Upper
278	100		
139	34.7	27.8	41.6
279	38.6	30.9	46.3





#37

Benzo(g,h,i)perylene

Concen: 1.30 ng

RT: 26.56 min Scan# 3999

Delta R.T. 0.00 min

Lab File: BE090321.D

Acq: 3 Aug 2015 22:25

Instrument :

BNA_E

ClientSampleId :

SSTDICCC001

Tgt Ion:	276	Resp:	2421
Ion Ratio	Lower	Upper	
276	100		
277	32.0	25.6	38.4
138	27.5	22.0	33.0

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:49 PM

