

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080315\
 Data File : BE090322.D
 Acq On : 3 Aug 2015 23:01
 Operator : TP/UM
 Sample : SSTDICC002
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Manual Integrations
 APPROVED

MMDadoda
 8/4/2015 3:26:56 PM

Quant Time: Aug 04 02:16:35 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 04 01:38:34 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	14347	5.00	ng	0.00
6) Naphthalene-d8	10.35	136	61466	5.00	ng	0.00
12) Acenaphthene-d10	14.21	164	32202	5.00	ng	0.00
18) Phenanthrene-d10	16.94	188	68646	5.00	ng	0.00
25) Chrysene-d12	21.11	240	63300	5.00	ng	0.00
32) Perylene-d12	23.42	264	53223	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.23	112	4895	1.76	ng	0.00
5) Phenol-d6	6.78	99	6347	1.94	ng	0.00
7) Nitrobenzene-d5	8.73	82	6956	1.85	ng	0.00
13) 2,4,6-Tribromophenol	15.71	330	785	2.06	ng	-0.02
14) 2-Fluorobiphenyl	12.83	172	16532	1.78	ng	0.00
27) Terphenyl-d14	19.57	244	20923	1.74	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.21	88	3324	1.57	ng	94
3) n-Nitrosodimethylamine	3.50	42	4722	1.93	ng	# 97
8) Nitrobenzene	8.78	77	7373	1.97	ng	98
9) Naphthalene	10.40	128	22876	1.74	ng	97
10) Hexachlorobutadiene	10.69	225	2973	1.66	ng	99
11) 2-Methylnaphthalene	12.01	142	14858	1.81	ng	99
15) Acenaphthylene	13.92	152	99379	1.78	ng	100
16) Acenaphthene	14.27	154	14290	1.76	ng	98
17) Fluorene	15.26	166	17937	1.85	ng	99
19) 4-Bromophenyl-phenylether	16.16	248	4559	1.75	ng	96
20) Hexachlorobenzene	16.27	284	16182	1.65	ng	99
21) Pentachlorophenol	16.61	266	901	2.01	ng	# 83
22) Phenanthrene	17.00	178	27267	1.75	ng	100
23) Anthracene	17.08	178	25406	1.83	ng	99
24) Fluoranthene	18.99	202	28099	1.81	ng	100
26) Pyrene	19.36	202	29265	1.58	ng	99
28) Benzo(a)anthracene	21.09	228	21525	2.13	ng	97
29) Chrysene	21.14	228	26577	1.42	ng	98
30) Bis(2-ethylhexyl)phthalate	21.04	149	7559	1.95	ng	99
31) Indeno(1,2,3-cd)pyrene	25.82	276	13115	3.22	ng	# 84
33) Benzo(b)fluoranthene	22.72	252	13733	2.85	ng	# 91
34) Benzo(k)fluoranthene	22.76	252	25817m	2.02	ng	
35) Benzo(a)pyrene	23.32	252	13272	2.39	ng	# 88
36) Dibenzo(a,h)anthracene	25.84	278	8981	3.27	ng	# 75
37) Benzo(g,h,i)perylene	26.55	276	14720	2.49	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080315\
Data File : BE090322.D
Acq On : 3 Aug 2015 23:01
Operator : TP/UM
Sample : SSTDICC002
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC002

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:56 PM

Quant Time: Aug 04 02:16:35 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 04 01:38:34 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.60 min Scan# 848

Delta R.T. -0.00 min

Lab File: BE090322.D


Acq: 3 Aug 2015 23:01

Instrument :

BNA_E

ClientSampleId :




SSTDICC002



Tgt Ion:	152	Resp:	14347
Ion	Ratio	Lower	Upper
152	100		
150	154.0	124.2	186.2
115	62.3	50.4	75.6

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:56 PM



#2

1,4-Dioxane

Concen: 1.57 ng




RT: 3.21 min Scan# 90


Delta R.T. -0.01 min

Lab File: BE090322.D

Acq: 3 Aug 2015 23:01

Tgt Ion:	88	Resp:	3324
Ion	Ratio	Lower	Upper
88	100		
58	68.1	50.2	75.2
43	33.7	25.0	37.6





#3





n-Nitrosodimethylamine
 Concen: 1.93 ng
 RT: 3.50 min Scan# 133
 Delta R.T. -0.01 min
 Lab File: BE090322.D
 Acq: 3 Aug 2015 23:01

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Manual Integrations
 APPROVED

MMDadoda
 8/4/2015 3:26:56 PM




Tgt Ion	Ratio	Lower	Upper
42	100		
74	89.2	73.2	109.8
44	6.2	3.8	5.6#




#4

2-Fluorophenol
 Concen: 1.76 ng
 RT: 5.23 min Scan# 388
 Delta R.T. 0.01 min
 Lab File: BE090322.D
 Acq: 3 Aug 2015 23:01

Tgt Ion	Ratio	Lower	Upper
112	100		
64	57.7	49.0	73.4
63	34.5	27.7	41.5





#5

Phenol-d6

Concen: 1.94 ng

RT: 6.78 min Scan# 668

Delta R.T. -0.01 min

Lab File: BE090322.D

Acq: 3 Aug 2015 23:01

Instrument :

BNA_E





ClientSampleId :

SSTDICC002

Tgt Ion:	99	Resp:	6347
Ion	Ratio	Lower	Upper
99	100		
42	26.6	22.2	33.2
71	36.2	29.2	43.8

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:56 PM



#6

Naphthalene-d8

Concen: 5.00 ng




RT: 10.35 min Scan# 1373


Delta R.T. -0.00 min

Lab File: BE090322.D

Acq: 3 Aug 2015 23:01

Tgt Ion:	136	Resp:	61466
Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.9	13.3
54	7.2	6.0	9.0
68	4.7	3.9	5.9





#7

Nitrobenzene-d5

Concen: 1.85 ng

RT: 8.73 min Scan# 1094

Delta R.T. -0.00 min

Lab File: BE090322.D

Acq: 3 Aug 2015 23:01

Instrument :

BNA_E





ClientSampleId :

SSTDICC002

Tgt Ion:	82	Resp:	6956
Ion	Ratio	Lower	Upper
82	100		
128	36.8	30.8	46.2
54	55.4	47.5	71.3

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:56 PM



#8

Nitrobenzene

Concen: 1.97 ng




RT: 8.78 min Scan# 1103

Delta R.T. -0.00 min

Lab File: BE090322.D

Acq: 3 Aug 2015 23:01

Tgt Ion:	77	Resp:	7373
Ion	Ratio	Lower	Upper
77	100		
123	38.7	29.6	44.4
65	12.8	10.4	15.6





#9

Naphthalene

Concen: 1.74 ng

RT: 10.40 min Scan# 1379

Delta R.T. -0.00 min

Lab File: BE090322.D


Acq: 3 Aug 2015 23:01

Instrument :

BNA_E

ClientSampleId :




SSTDICC002



Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.4	10.1	15.1
127	13.2	11.3	16.9

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:56 PM



#10

Hexachlorobutadiene

Concen: 1.66 ng




RT: 10.69 min Scan# 1418


Delta R.T. -0.00 min

Lab File: BE090322.D

Acq: 3 Aug 2015 23:01

Tgt Ion	Ratio	Lower	Upper
225	100		
223	63.3	50.4	75.6
227	64.0	49.6	74.4





#11

2-Methylnaphthalene

Concen: 1.81 ng

RT: 12.01 min Scan# 1651

Delta R.T. -0.00 min

Lab File: BE090322.D


Acq: 3 Aug 2015 23:01

Instrument :

BNA_E

ClientSampleId :

SSTDICC002



Tgt Ion:142 Resp: 14858

Ion Ratio Lower Upper




142 100

141 92.1 74.6 112.0

115 39.5 32.6 49.0

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:56 PM



#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.21 min Scan# 2076

Delta R.T. -0.00 min

Lab File: BE090322.D

Acq: 3 Aug 2015 23:01




Tgt Ion:164 Resp: 32202


Ion Ratio Lower Upper

164 100

162 107.5 82.7 124.1

160 54.2 43.0 64.6









#13
 2,4,6-Tribromophenol
 Concen: 2.06 ng
 RT: 15.71 min Scan# 2206
 Delta R.T. -0.02 min
 Lab File: BE090322.D
 Acq: 3 Aug 2015 23:01

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Tgt Ion	Ratio	Lower	Upper
330	100		
332	92.5	79.4	119.0
141	274.9	245.9	368.9




Manual Integrations
 APPROVED


MMDadoda
 8/4/2015 3:26:56 PM



#14
 2-Fluorobiphenyl
 Concen: 1.78 ng
 RT: 12.83 min Scan# 1830
 Delta R.T. -0.00 min
 Lab File: BE090322.D
 Acq: 3 Aug 2015 23:01

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.5	29.5	44.3
170	24.7	21.3	31.9









#15
Acenaphthylene
Concen: 1.78 ng
RT: 13.92 min Scan# 2032
Delta R.T. -0.00 min
Lab File: BE090322.D
Acq: 3 Aug 2015 23:01

Instrument :
BNA_E
ClientSampleId :
SSTDICC002

Tgt Ion	Ratio	Lower	Upper
152	100		
151	5.0	4.0	6.0
153	13.0	10.2	15.4




Manual Integrations
APPROVED


MMDadoda
8/4/2015 3:26:56 PM



#16
Acenaphthene
Concen: 1.76 ng
RT: 14.27 min Scan# 2086
Delta R.T. 0.00 min
Lab File: BE090322.D
Acq: 3 Aug 2015 23:01

Tgt Ion	Ratio	Lower	Upper
154	100		
153	473.1	378.0	567.0
152	254.3	212.7	319.1









#17
Fluorene
 Concen: 1.85 ng
 RT: 15.26 min Scan# 2180
 Delta R.T. -0.00 min
 Lab File: BE090322.D
 Acq: 3 Aug 2015 23:01

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Tgt Ion	Ratio	Lower	Upper
166	100		
165	102.8	82.6	123.8
167	13.6	11.8	17.8




Manual Integrations
 APPROVED


MMDadoda
 8/4/2015 3:26:56 PM



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 16.94 min Scan# 2277
 Delta R.T. -0.00 min
 Lab File: BE090322.D
 Acq: 3 Aug 2015 23:01


Tgt Ion	Ratio	Lower	Upper
188	100		
94	11.4	9.5	14.3
80	12.3	10.0	15.0





#19
4-Bromophenyl-phenylether
Concen: 1.75 ng
RT: 16.16 min Scan# 2232
Delta R.T. -0.00 min
Lab File: BE090322.D
Acq: 3 Aug 2015 23:01




Instrument :
BNA_E
ClientSampleId :
SSTDICC002



Tgt Ion:248 Resp: 4559
Ion Ratio Lower Upper
248 100
250 99.8 74.0 111.0
141 318.7 258.9 388.3




Manual Integrations
APPROVED


MMDadoda
8/4/2015 3:26:56 PM



#20
Hexachlorobenzene
Concen: 1.65 ng
RT: 16.27 min Scan# 2238
Delta R.T. -0.00 min
Lab File: BE090322.D
Acq: 3 Aug 2015 23:01

Tgt Ion:284 Resp: 16182
Ion Ratio Lower Upper
284 100
142 54.4 44.2 66.4
249 32.7 26.4 39.6









#21
 Pentachlorophenol
 Concen: 2.01 ng
 RT: 16.61 min Scan# 2258
 Delta R.T. -0.02 min
 Lab File: BE090322.D
 Acq: 3 Aug 2015 23:01

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Tgt Ion	Ratio	Lower	Upper
266	100		
268	65.2	66.2	99.2#
264	73.0	69.1	103.7




Manual Integrations
 APPROVED


MMDadoda
 8/4/2015 3:26:56 PM



#22
 Phenanthrene
 Concen: 1.75 ng
 RT: 17.00 min Scan# 2280
 Delta R.T. -0.00 min
 Lab File: BE090322.D
 Acq: 3 Aug 2015 23:01

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.5	23.3
179	15.4	12.6	18.8





#23

Anthracene

Concen: 1.83 ng

RT: 17.08 min Scan# 2285

Delta R.T. -0.00 min

Lab File: BE090322.D

Acq: 3 Aug 2015 23:01

Instrument :

BNA_E





ClientSampleId :

SSTDIC002

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:56 PM

Tgt Ion:	178	Resp:	25406
Ion Ratio	Lower	Upper	
178	100		
176	19.2	15.1	22.7
179	15.2	12.1	18.1



#24

Fluoranthene

Concen: 1.81 ng




RT: 18.99 min Scan# 2507


Delta R.T. -0.01 min

Lab File: BE090322.D

Acq: 3 Aug 2015 23:01

Tgt Ion:	202	Resp:	28099
Ion Ratio	Lower	Upper	
202	100		
101	10.4	8.3	12.5
203	17.2	13.8	20.8





#25

Chrysene-d12

Concen: 5.00 ng

RT: 21.11 min Scan# 2866

Delta R.T. -0.01 min

Lab File: BE090322.D


Acq: 3 Aug 2015 23:01

Instrument :

BNA_E

ClientSampleId :




SSTDICC002



Tgt Ion:	240	Resp:	63300
Ion Ratio	Lower	Upper	
240	100		
120	10.9	7.5	11.3
236	26.8	21.0	31.4

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:56 PM



#26

Pyrene

Concen: 1.58 ng




RT: 19.36 min Scan# 2576


Delta R.T. -0.00 min

Lab File: BE090322.D

Acq: 3 Aug 2015 23:01

Tgt Ion:	202	Resp:	29265
Ion Ratio	Lower	Upper	
202	100		
200	21.0	16.6	24.8
203	17.5	13.8	20.6





#27

Terphenyl-d14

Concen: 1.74 ng

RT: 19.57 min Scan# 2616

Delta R.T. -0.01 min

Lab File: BE090322.D


Acq: 3 Aug 2015 23:01

Instrument :

BNA_E

ClientSampleId :




SSTDICC002



Tgt Ion:	244	Resp:	20923
Ion Ratio	Lower	Upper	
244	100		
212	7.5	7.4	11.2
122	10.6	10.4	15.6

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:56 PM



#28

Benzo(a)anthracene

Concen: 2.13 ng




RT: 21.09 min Scan# 2864


Delta R.T. -0.01 min

Lab File: BE090322.D

Acq: 3 Aug 2015 23:01

Tgt Ion:	228	Resp:	21525
Ion Ratio	Lower	Upper	
228	100		
226	27.7	23.0	34.4
229	19.4	17.4	26.2









#29
Chrysene
 Concen: 1.42 ng
 RT: 21.14 min Scan# 2871
 Delta R.T. -0.01 min
 Lab File: BE090322.D
 Acq: 3 Aug 2015 23:01

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Tgt Ion	Ratio	Resp	Lower	Upper
228	100	26577		
226	30.5		24.9	37.3
229	19.4		16.6	25.0




Manual Integrations
 APPROVED


MMDadoda
 8/4/2015 3:26:56 PM



#30
Bis(2-ethylhexyl)phthalate
 Concen: 1.95 ng
 RT: 21.04 min Scan# 2856
 Delta R.T. 0.00 min
 Lab File: BE090322.D
 Acq: 3 Aug 2015 23:01

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	7559		
167	21.2		16.6	24.8
279	2.8		2.7	4.1









#31
Indeno(1,2,3-cd)pyrene
Concen: 3.22 ng
RT: 25.82 min Scan# 3835
Delta R.T. -0.01 min
Lab File: BE090322.D
Acq: 3 Aug 2015 23:01

Instrument :
BNA_E
ClientSampleId :
SSTDICC002

Tgt Ion	Ratio	Lower	Upper
276	100		
138	21.1	10.6	16.0#
277	24.7	14.6	22.0#




Manual Integrations
APPROVED


MMDadoda
8/4/2015 3:26:56 PM



#32
Perylene-d12
Concen: 5.00 ng
RT: 23.42 min Scan# 3313
Delta R.T. -0.00 min
Lab File: BE090322.D
Acq: 3 Aug 2015 23:01

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.1	20.1	30.1
265	21.7	18.6	28.0





#33

Benzo(b)fluoranthene

Concen: 2.85 ng

RT: 22.72 min Scan# 3158

Delta R.T. -0.01 min

Lab File: BE090322.D


Acq: 3 Aug 2015 23:01

Instrument :

BNA_E

ClientSampleId :



SSTDICC002



Tgt Ion:	252	Resp:	13733
Ion Ratio	Lower	Upper	
252	100		
253	22.7	21.5	32.3
125	10.8	11.8	17.8#

Manual Integrations
APPROVED

MMDadoda
8/4/2015 3:26:56 PM



#34

Benzo(k)fluoranthene

Concen: 2.02 ng m




RT: 22.76 min Scan# 3168


Delta R.T. -0.00 min

Lab File: BE090322.D

Acq: 3 Aug 2015 23:01

Tgt Ion:	252	Resp:	25817
Ion Ratio	Lower	Upper	
252	100		
253	22.3	19.9	29.9
125	10.6	11.2	16.8#









#35
Benzo(a)pyrene
Concen: 2.39 ng
RT: 23.32 min Scan# 3290
Delta R.T. -0.00 min
Lab File: BE090322.D
Acq: 3 Aug 2015 23:01

Instrument :
BNA_E
ClientSampleId :
SSTDICC002

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.6	23.0	34.4#
125	11.8	14.2	21.2#




Manual Integrations
APPROVED


MMDadoda
8/4/2015 3:26:56 PM



#36
Dibenzo(a,h)anthracene
Concen: 3.27 ng
RT: 25.84 min Scan# 3840
Delta R.T. -0.02 min
Lab File: BE090322.D
Acq: 3 Aug 2015 23:01

Tgt Ion	Ratio	Lower	Upper
278	100		
139	17.3	27.8	41.6#
279	26.7	30.9	46.3#





#37
 Benzo(g,h,i)perylene
 Concen: 2.49 ng
 RT: 26.55 min Scan# 3996
 Delta R.T. -0.01 min
 Lab File: BE090322.D
 Acq: 3 Aug 2015 23:01

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.9	25.6	38.4#
138	21.5	22.0	33.0#

Manual Integrations
 APPROVED

MMDadoda
 8/4/2015 3:26:56 PM

