

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080316\
 Data File : BE092169.D
 Acq On : 4 Aug 2016 6:13
 Operator : UM/SJ
 Sample : H4221-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 GW-43-8.7-15.0

Quant Time: Aug 04 07:21:52 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	12544	5.00	ng	0.00
6) Naphthalene-d8	10.98	136	60095	5.00	ng	-0.01
10) Acenaphthene-d10	14.85	164	34597	5.00	ng	0.00
16) Phenanthrene-d10	17.56	188	76736	5.00	ng	-0.02
22) Chrysene-d12	21.75	240	89701	5.00	ng	0.00
28) Perylene-d12	24.14	264	89202	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.69	112	8623	2.68	ng	0.00
5) Phenol-d6	7.35	99	8733	2.03	ng	0.00
7) Nitrobenzene-d5	9.34	82	11357	2.40	ng	-0.01
11) 2,4,6-Tribromophenol	16.33	330	7934	7.06	ng	0.00
12) 2-Fluorobiphenyl	13.46	172	30345	2.77	ng	0.00
24) Terphenyl-d14	20.19	244	59012	4.58	ng	-0.02

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080316\
Data File : BE092169.D
Acq On : 4 Aug 2016 6:13
Operator : UM/SJ
Sample : H4221-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
GW-43-8.7-15.0

Quant Time: Aug 04 07:21:52 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-PAH-BE080216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 02 16:59:37 2016
Response via : Initial Calibration

