

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080316\
 Data File : BE092149.D
 Acq On : 3 Aug 2016 17:26
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 04 03:56:32 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	91	0.00
2	1,4-Dioxane	0.400	0.407	-1.7	89	0.00
3	n-Nitrosodimethylamine	0.400	0.379	5.3	83	0.00
4 S	2-Fluorophenol	0.400	0.314	21.5	81	0.00
5 S	Phenol-d6	0.400	0.342	14.5	85	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	95	-0.01
7 S	Nitrobenzene-d5	0.400	0.367	8.3	92	0.00
8	Naphthalene	0.400	0.401	-0.3	94	0.00
9	2-Methylnaphthalene	0.400	0.391	2.3	93	0.00
10 I	Acenaphthene-d10	5.000	5.000	0.0	96	-0.02
11 S	2,4,6-Tribromophenol	0.400	0.310	22.5	89	0.00
12 S	2-Fluorobiphenyl	0.400	0.380	5.0	95	0.00
13	Acenaphthylene	0.400	0.379	5.3	96	0.00
14	Acenaphthene	0.400	0.392	2.0	97	0.00
15	Fluorene	0.400	0.369	7.8	94	0.00
16 I	Phenanthrene-d10	5.000	5.000	0.0	99	-0.02
17	Hexachlorobenzene	0.400	0.380	5.0	93	0.00
18	Pentachlorophenol	0.400	0.373	6.8	75	0.00
19	Phenanthrene	0.400	0.381	4.8	101	0.00
20	Anthracene	0.400	0.356	11.0	94	0.00
21	Fluoranthene	0.400	0.365	8.8	94	0.00
22 I	Chrysene-d12	5.000	5.000	0.0	94	0.00
23	Pyrene	0.400	0.374	6.5	95	0.00
24 S	Terphenyl-d14	0.400	0.409	-2.2	105	0.00
25	Benzo(a)anthracene	0.400	0.362	9.5	91	0.00
26	Chrysene	0.400	0.387	3.3	96	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.371	7.3	94	0.00
28 I	Perylene-d12	5.000	5.000	0.0	96	0.00
29	Benzo(b)fluoranthene	0.400	0.353	11.8	83	0.00
30	Benzo(k)fluoranthene	0.400	0.374	6.5	103	0.00
31 C	Benzo(a)pyrene	0.400	0.382	4.5	95	0.00
32	Dibenzo(a,h)anthracene	0.400	0.361	9.8	91	0.00
33	Benzo(g,h,i)perylene	0.400	0.369	7.8	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0