

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080316\
 Data File : BE092166.D
 Acq On : 4 Aug 2016 4:17
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 04 06:05:12 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	100	0.00
2	1,4-Dioxane	0.400	0.395	1.3	95	0.00
3	n-Nitrosodimethylamine	0.400	0.397	0.8	97	0.00
4 S	2-Fluorophenol	0.400	0.347	13.3	98	0.00
5 S	Phenol-d6	0.400	0.382	4.5	104	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	104	-0.01
7 S	Nitrobenzene-d5	0.400	0.384	4.0	105	0.00
8	Naphthalene	0.400	0.404	-1.0	103	0.00
9	2-Methylnaphthalene	0.400	0.410	-2.5	106	0.00
10 I	Acenaphthene-d10	5.000	5.000	0.0	107	-0.02
11 S	2,4,6-Tribromophenol	0.400	0.361	9.8	116	0.00
12 S	2-Fluorobiphenyl	0.400	0.388	3.0	109	0.00
13	Acenaphthylene	0.400	0.380	5.0	108	-0.02
14	Acenaphthene	0.400	0.396	1.0	109	0.00
15	Fluorene	0.400	0.380	5.0	108	0.00
16 I	Phenanthrene-d10	5.000	5.000	0.0	114	-0.02
17	Hexachlorobenzene	0.400	0.379	5.3	107	0.00
18	Pentachlorophenol	0.400	0.409	-2.2	104	0.00
19	Phenanthrene	0.400	0.379	5.3	116	0.00
20	Anthracene	0.400	0.359	10.3	109	0.00
21	Fluoranthene	0.400	0.373	6.8	111	0.00
22 I	Chrysene-d12	5.000	5.000	0.0	102	0.00
23	Pyrene	0.400	0.401	-0.3	111	0.00
24 S	Terphenyl-d14	0.400	0.492	-23.0	137	-0.02
25	Benzo(a)anthracene	0.400	0.402	-0.5	110	0.00
26	Chrysene	0.400	0.392	2.0	105	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.374	6.5	103	-0.02
28 I	Perylene-d12	5.000	5.000	0.0	103	0.00
29	Benzo(b)fluoranthene	0.400	0.373	6.8	94	0.00
30	Benzo(k)fluoranthene	0.400	0.386	3.5	114	0.00
31 C	Benzo(a)pyrene	0.400	0.393	1.8	105	0.00
32	Dibenzo(a,h)anthracene	0.400	0.380	5.0	103	0.00
33	Benzo(g,h,i)perylene	0.400	0.375	6.3	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0