

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080316\
 Data File : BE092166.D
 Acq On : 4 Aug 2016 4:17
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 04 06:05:12 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.680	0.671	1.3	95	0.00
3	n-Nitrosodimethylamine	0.942	0.737	21.8	97	0.00
4 S	2-Fluorophenol	1.284	1.113	13.3	98	0.00
5 S	Phenol-d6	1.711	1.635	4.4	104	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	104	-0.01
7 S	Nitrobenzene-d5	0.395	0.379	4.1	105	0.00
8	Naphthalene	1.159	1.115	3.8	103	0.00
9	2-Methylnaphthalene	0.743	0.706	5.0	106	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	107	-0.02
11 S	2,4,6-Tribromophenol	0.162	0.146	9.9	116	0.00
12 S	2-Fluorobiphenyl	1.584	1.535	3.1	109	0.00
13	Acenaphthylene	8.821	8.374	5.1	108	-0.02
14	Acenaphthene	1.406	1.392	1.0	109	0.00
15	Fluorene	1.783	1.692	5.1	108	0.00
16 I	Phenanthrene-d10	1.000	1.000	0.0	114	-0.02
17	Hexachlorobenzene	0.245	0.232	5.3	107	0.00
18	Pentachlorophenol	0.068	0.051	25.0#	104	0.00
19	Phenanthrene	1.460	1.410	3.4	116	0.00
20	Anthracene	1.331	1.195	10.2	109	0.00
21	Fluoranthene	6.116	5.704	6.7	111	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
23	Pyrene	5.228	5.239	-0.2	111	0.00
24 S	Terphenyl-d14	0.718	0.883	-23.0	137	-0.02
25	Benzo(a)anthracene	5.496	5.527	-0.6	110	0.00
26	Chrysene	5.505	5.393	2.0	105	0.00
27	Indeno(1,2,3-cd)pyrene	5.978	5.590	6.5	103	-0.02
28 I	Perylene-d12	1.000	1.000	0.0	103	0.00
29	Benzo(b)fluoranthene	1.368	1.276	6.7	94	0.00
30	Benzo(k)fluoranthene	1.369	1.322	3.4	114	0.00
31 C	Benzo(a)pyrene	1.257	1.234	1.8	105	0.00
32	Dibenzo(a,h)anthracene	1.200	1.141	4.9	103	0.00
33	Benzo(g,h,i)perylene	5.122	4.797	6.3	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0