

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080316\
 Data File : BE092182.D
 Acq On : 4 Aug 2016 14:40
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 04 15:15:49 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.680	0.638	6.2	89	0.00
3	n-Nitrosodimethylamine	0.942	0.750	20.4	98	0.00
4 S	2-Fluorophenol	1.284	1.080	15.9	94	0.01
5 S	Phenol-d6	1.711	1.603	6.3	101	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	106	-0.02
7 S	Nitrobenzene-d5	0.395	0.374	5.3	105	0.00
8	Naphthalene	1.159	1.098	5.3	103	0.00
9	2-Methylnaphthalene	0.743	0.697	6.2	107	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	107	-0.02
11 S	2,4,6-Tribromophenol	0.162	0.137	15.4	109	0.00
12 S	2-Fluorobiphenyl	1.584	1.527	3.6	108	0.00
13	Acenaphthylene	8.821	8.386	4.9	108	-0.02
14	Acenaphthene	1.406	1.386	1.4	109	0.00
15	Fluorene	1.783	1.674	6.1	107	-0.01
16 I	Phenanthrene-d10	1.000	1.000	0.0	114	-0.02
17	Hexachlorobenzene	0.245	0.225	8.2	104	0.00
18	Pentachlorophenol	0.068	0.048	29.4#	99	0.00
19	Phenanthrene	1.460	1.357	7.1	111	0.00
20	Anthracene	1.331	1.191	10.5	109	0.00
21	Fluoranthene	6.116	5.413	11.5	105	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
23	Pyrene	5.228	5.250	-0.4	108	0.00
24 S	Terphenyl-d14	0.718	0.771	-7.4	116	-0.02
25	Benzo(a)anthracene	5.496	5.171	5.9	100	0.00
26	Chrysene	5.505	5.210	5.4	99	0.00
27	Indeno(1,2,3-cd)pyrene	5.978	5.349	10.5	96	0.00
28 I	Perylene-d12	1.000	1.000	0.0	95	0.00
29	Benzo(b)fluoranthene	1.368	1.255	8.3	86	0.00
30	Benzo(k)fluoranthene	1.369	1.329	2.9	107	-0.01
31 C	Benzo(a)pyrene	1.257	1.242	1.2	98	0.00
32	Dibenzo(a,h)anthracene	1.200	1.141	4.9	96	0.00
33	Benzo(g,h,i)perylene	5.122	4.800	6.3	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0