

Data Path : Z:\HPCHEM1\BNA E\Data\BE080317\
 Data File : BE093660.D
 Acq On : 3 Aug 2017 17:12
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 04 04:56:53 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 02 10:58:31 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	-0.01
2	1,4-Dioxane	0.648	0.644	0.6	108	0.00
3	n-Nitrosodimethylamine	0.607	0.624	-2.8	118	0.00
4 S	2-Fluorophenol	1.192	1.316	-10.4	124	0.00
5 S	Phenol-d6	1.794	2.015	-12.3	126	-0.01
6	bis(2-Chloroethyl)ether	1.508	1.531	-1.5	113	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
8 S	Nitrobenzene-d5	0.305	0.313	-2.6	128	0.00
9	Naphthalene	4.630	4.502	2.8	114	0.00
10	Hexachlorobutadiene	0.172	0.166	3.5	115	-0.02
11 SURR	2-Methylnaphthalene-d10	0.632	0.632	0.0	117	0.00
12	2-Methylnaphthalene	0.766	0.760	0.8	116	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
14 S	2,4,6-Tribromophenol	0.093	0.100	-7.5	155#	0.00
15 S	2-Fluorobiphenyl	1.601	1.556	2.8	114	-0.01
16	Acenaphthylene	1.912	1.942	-1.6	124	0.00
17	Acenaphthene	1.349	1.327	1.6	118	0.00
18	Fluorene	1.803	1.776	1.5	115	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	127	0.00
20	4-Bromophenyl-phenylether	0.126	0.115	8.7	104	0.00
21	Hexachlorobenzene	0.129	0.123	4.7	112	0.00
22	Pentachlorophenol	0.055	0.078	-41.8#	188#	0.00
23	Phenanthrene	0.880	0.838	4.8	117	0.00
24	Anthracene	1.203	1.253	-4.2	125	0.00
25 SURR	Fluoranthene-d10	4.541	4.601	-1.3	118	0.00
26	Fluoranthene	1.401	1.400	0.1	117	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	116	-0.01
28	Pvrene	1.292	1.303	-0.9	117	0.00
29 S	Terphenyl-d14	0.719	0.713	0.8	113	0.00
30	Benzo(a)anthracene	1.176	1.268	-7.8	129	0.00
31	Chrysene	1.311	1.277	2.6	112	0.00
32	Bis(2-ethylhexyl)phthalate	1.603	2.646	-65.1#	222#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.265	-6.6	122	0.00
34 I	Perylene-d12	1.000	1.000	0.0	123	0.00
35	Benzo(b)fluoranthene	1.427	1.355	5.0	120	0.00
36	Benzo(k)fluoranthene	1.459	1.402	3.9	122	0.00
37 C	Benzo(a)pyrene	1.293	1.297	-0.3	128	0.00
38	Dibenzo(a,h)anthracene	1.248	1.196	4.2	120	0.00
39	Benzo(a,h,i)perylene	1.266	1.209	4.5	120	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE080317\
Data File : BE093660.D
Acq On : 3 Aug 2017 17:12
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 04 04:56:53 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 02 10:58:31 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			