

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080415\
 Data File : BE090344.D
 Acq On : 4 Aug 2015 15:34
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 04 17:41:38 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 04 02:31:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	-0.01
2	1,4-Dioxane	0.796	0.799	-0.4	94	-0.02
3	n-Nitrosodimethylamine	0.845	0.985	-16.6	109	-0.02
4 S	2-Fluorophenol	0.956	1.086	-13.6	111	-0.01
5 S	Phenol-d6	1.175	1.234	-5.0	108	-0.01
6 I	Naphthalene-d8	1.000	1.000	0.0	108	-0.02
7 S	Nitrobenzene-d5	0.295	0.284	3.7	108	0.00
8	Nitrobenzene	0.305	0.294	3.6	110	0.00
9	Naphthalene	1.074	1.053	2.0	106	-0.02
10	Hexachlorobutadiene	0.144	0.132	8.3	96	0.00
11	2-Methylnaphthalene	0.665	0.649	2.4	106	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	107	-0.01
13 S	2,4,6-Tribromophenol	0.067	0.074	-10.4	164#	-0.02
14 S	2-Fluorobiphenyl	1.449	1.490	-2.8	109	-0.01
15	Acenaphthylene	8.730	8.557	2.0	105	-0.01
16	Acenaphthene	1.262	1.213	3.9	104	-0.01
17	Fluorene	1.533	1.509	1.6	112	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
19	4-Bromophenyl-phenylether	0.188	0.184	2.1	106	-0.02
20	Hexachlorobenzene	0.710	0.678	4.5	104	0.00
21	Pentachlorophenol	0.038	0.794	-1989.5#	3601#	-0.03
22	Phenanthrene	1.139	1.113	2.3	114	-0.02
23	Anthracene	1.018	0.988	2.9	115	-0.02
24	Fluoranthene	1.146	1.149	-0.3	117	-0.01
25 I	Chrysene-d12	1.000	1.000	0.0	142	-0.01
26	Pyrene	1.352	1.195	11.6	120	-0.01
27 S	Terphenyl-d14	0.936	0.915	2.2	133	-0.02
28	Benzo(a)anthracene	0.951	0.968	-1.8	162#	-0.01
29	Chrysene	1.320	1.142	13.5	118	-0.01
30	Bis(2-ethylhexyl)phthalate	0.335	0.385	-14.9	230#	0.00
31	Indeno(1,2,3-cd)pyrene	0.548	0.891	-62.6#	322#	-0.04
32 I	Perylene-d12	1.000	1.000	0.0	165#	-0.02
33	Benzo(b)fluoranthene	0.676	0.820	-21.3	234#	-0.02
34	Benzo(k)fluoranthene	1.357	1.172	13.6	133	-0.01
35 C	Benzo(a)pyrene	0.673	0.801	-19.0	223#	-0.02
36	Dibenzo(a,h)anthracene	0.431	0.705	-63.6#	385#	-0.05
37	Benzo(g,h,i)perylene	0.744	0.924	-24.2	212#	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0