

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080415\
 Data File : BE090344.D
 Acq On : 4 Aug 2015 15:34
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 04 17:41:38 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 04 02:31:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	95	-0.01
2	1,4-Dioxane	1.000	1.208	-20.8	94	-0.02
3	n-Nitrosodimethylamine	1.000	1.165	-16.5	109	-0.02
4 S	2-Fluorophenol	1.000	1.137	-13.7	111	-0.01
5 S	Phenol-d6	1.000	1.050	-5.0	108	-0.01
6 I	Naphthalene-d8	5.000	5.000	0.0	108	-0.02
7 S	Nitrobenzene-d5	1.000	0.961	3.9	108	0.00
8	Nitrobenzene	1.000	0.964	3.6	110	0.00
9	Naphthalene	1.000	0.980	2.0	106	-0.02
10	Hexachlorobutadiene	1.000	0.915	8.5	96	0.00
11	2-Methylnaphthalene	1.000	0.976	2.4	106	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	107	-0.01
13 S	2,4,6-Tribromophenol	1.000	1.231	-23.1	164	-0.02
14 S	2-Fluorobiphenyl	1.000	1.028	-2.8	109	-0.01
15	Acenaphthylene	1.000	0.980	2.0	105	-0.01
16	Acenaphthene	1.000	0.961	3.9	104	-0.01
17	Fluorene	1.000	0.984	1.6	112	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	111	0.00
19	4-Bromophenyl-phenylether	1.000	0.978	2.2	106	-0.02
20	Hexachlorobenzene	1.000	0.954	4.6	104	0.00
21	Pentachlorophenol	1.000	10.710	-971.0#	3601	-0.03
22	Phenanthrene	1.000	0.978	2.2	114	-0.02
23	Anthracene	1.000	0.970	3.0	115	-0.02
24	Fluoranthene	1.000	1.003	-0.3	117	-0.01
25 I	Chrysene-d12	5.000	5.000	0.0	142	-0.01
26	Pyrene	1.000	0.884	11.6	120	-0.01
27 S	Terphenyl-d14	1.000	0.978	2.2	133	-0.02
28	Benzo(a)anthracene	1.000	1.017	-1.7	162	-0.01
29	Chrysene	1.000	0.959	4.1	118	-0.01
30	Bis(2-ethylhexyl)phthalate	1.000	1.275	-27.5#	230	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.639	-63.9#	322	-0.04
32 I	Perylene-d12	5.000	5.000	0.0	165	-0.02
33	Benzo(b)fluoranthene	1.000	1.143	-14.3	234	-0.02
34	Benzo(k)fluoranthene	1.000	0.863	13.7	133	-0.01
35 C	Benzo(a)pyrene	1.000	1.150	-15.0	223	-0.02
36	Dibenzo(a,h)anthracene	1.000	1.603	-60.3#	385	-0.05
37	Benzo(g,h,i)perylene	1.000	1.189	-18.9	212	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0