

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080416\
 Data File : BE092184.D
 Acq On : 4 Aug 2016 16:01
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 05 03:20:01 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	0.00
2	1,4-Dioxane	0.680	0.719	-5.7	75	0.00
3	n-Nitrosodimethylamine	0.942	0.736	21.9	71	0.00
4 S	2-Fluorophenol	1.284	1.069	16.7	69	0.01
5 S	Phenol-d6	1.711	1.413	17.4	66	0.01
6 I	Naphthalene-d8	1.000	1.000	0.0	77	-0.01
7 S	Nitrobenzene-d5	0.395	0.315	20.3	64	0.00
8	Naphthalene	1.159	1.121	3.3	77	0.00
9	2-Methylnaphthalene	0.743	0.698	6.1	77	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	77	-0.02
11 S	2,4,6-Tribromophenol	0.162	0.134	17.3	76	0.00
12 S	2-Fluorobiphenyl	1.584	1.545	2.5	78	0.00
13	Acenaphthylene	8.821	8.482	3.8	79	0.00
14	Acenaphthene	1.406	1.397	0.6	79	0.00
15	Fluorene	1.783	1.685	5.5	77	0.00
16 I	Phenanthrene-d10	1.000	1.000	0.0	80	-0.02
17	Hexachlorobenzene	0.245	0.241	1.6	78	0.00
18	Pentachlorophenol	0.068	0.046	32.4#	66	0.00
19	Phenanthrene	1.460	1.424	2.5	82	0.00
20	Anthracene	1.331	1.164	12.5	75	0.00
21	Fluoranthene	6.116	5.610	8.3	76	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	74	0.00
23	Pyrene	5.228	5.023	3.9	77	0.00
24 S	Terphenyl-d14	0.718	0.798	-11.1	90	0.00
25	Benzo(a)anthracene	5.496	5.193	5.5	75	0.00
26	Chrysene	5.505	5.360	2.6	76	0.00
27	Indeno(1,2,3-cd)pyrene	5.978	5.333	10.8	71	0.00
28 I	Perylene-d12	1.000	1.000	0.0	72	0.00
29	Benzo(b)fluoranthene	1.368	1.248	8.8	64	0.00
30	Benzo(k)fluoranthene	1.369	1.363	0.4	82	-0.01
31 C	Benzo(a)pyrene	1.257	1.217	3.2	72	0.00
32	Dibenzo(a,h)anthracene	1.200	1.132	5.7	72	0.00
33	Benzo(g,h,i)perylene	5.122	4.865	5.0	72	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0