

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080416\
 Data File : BE092184.D
 Acq On : 4 Aug 2016 16:01
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 05 03:20:01 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	73	0.00
2	1,4-Dioxane	0.400	0.423	-5.7	75	0.00
3	n-Nitrosodimethylamine	0.400	0.396	1.0	71	0.00
4 S	2-Fluorophenol	0.400	0.333	16.8	69	0.01
5 S	Phenol-d6	0.400	0.330	17.5	66	0.01
6 I	Naphthalene-d8	5.000	5.000	0.0	77	-0.01
7 S	Nitrobenzene-d5	0.400	0.319	20.3	64	0.00
8	Naphthalene	0.400	0.407	-1.7	77	0.00
9	2-Methylnaphthalene	0.400	0.405	-1.3	77	0.00
10 I	Acenaphthene-d10	5.000	5.000	0.0	77	-0.02
11 S	2,4,6-Tribromophenol	0.400	0.330	17.5	76	0.00
12 S	2-Fluorobiphenyl	0.400	0.390	2.5	78	0.00
13	Acenaphthylene	0.400	0.385	3.8	79	0.00
14	Acenaphthene	0.400	0.397	0.8	79	0.00
15	Fluorene	0.400	0.378	5.5	77	0.00
16 I	Phenanthrene-d10	5.000	5.000	0.0	80	-0.02
17	Hexachlorobenzene	0.400	0.394	1.5	78	0.00
18	Pentachlorophenol	0.400	0.386	3.5	66	0.00
19	Phenanthrene	0.400	0.383	4.3	82	0.00
20	Anthracene	0.400	0.350	12.5	75	0.00
21	Fluoranthene	0.400	0.367	8.3	76	0.00
22 I	Chrysene-d12	5.000	5.000	0.0	74	0.00
23	Pyrene	0.400	0.384	4.0	77	0.00
24 S	Terphenyl-d14	0.400	0.445	-11.2	90	0.00
25	Benzo(a)anthracene	0.400	0.378	5.5	75	0.00
26	Chrysene	0.400	0.389	2.8	76	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.357	10.8	71	0.00
28 I	Perylene-d12	5.000	5.000	0.0	72	0.00
29	Benzo(b)fluoranthene	0.400	0.365	8.8	64	0.00
30	Benzo(k)fluoranthene	0.400	0.399	0.3	82	-0.01
31 C	Benzo(a)pyrene	0.400	0.387	3.3	72	0.00
32	Dibenzo(a,h)anthracene	0.400	0.378	5.5	72	0.00
33	Benzo(g,h,i)perylene	0.400	0.380	5.0	72	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0