

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080416\
 Data File : BE092201.D
 Acq On : 5 Aug 2016 3:06
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 05 04:07:30 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2	1,4-Dioxane	0.680	0.696	-2.4	76	-0.01
3	n-Nitrosodimethylamine	0.942	0.750	20.4	76	0.00
4 S	2-Fluorophenol	1.284	1.106	13.9	75	0.00
5 S	Phenol-d6	1.711	1.489	13.0	73	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	83	-0.01
7 S	Nitrobenzene-d5	0.395	0.320	19.0	70	0.00
8	Naphthalene	1.159	1.118	3.5	82	0.00
9	2-Methylnaphthalene	0.743	0.708	4.7	85	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	84	-0.02
11 S	2,4,6-Tribromophenol	0.162	0.136	16.0	84	0.00
12 S	2-Fluorobiphenyl	1.584	1.538	2.9	85	0.00
13	Acenaphthylene	8.821	8.546	3.1	87	-0.02
14	Acenaphthene	1.406	1.369	2.6	84	0.00
15	Fluorene	1.783	1.704	4.4	85	-0.01
16 I	Phenanthrene-d10	1.000	1.000	0.0	89	-0.02
17	Hexachlorobenzene	0.245	0.232	5.3	84	0.00
18	Pentachlorophenol	0.068	0.054	20.6	87	0.00
19	Phenanthrene	1.460	1.409	3.5	91	0.00
20	Anthracene	1.331	1.168	12.2	84	0.00
21	Fluoranthene	6.116	5.355	12.4	81	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	75	0.00
23	Pyrene	5.228	5.282	-1.0	82	0.00
24 S	Terphenyl-d14	0.718	0.840	-17.0	96	0.00
25	Benzo(a)anthracene	5.496	5.314	3.3	78	0.00
26	Chrysene	5.505	5.418	1.6	78	0.00
27	Indeno(1,2,3-cd)pyrene	5.978	5.456	8.7	74	0.00
28 I	Perylene-d12	1.000	1.000	0.0	73	0.00
29	Benzo(b)fluoranthene	1.368	1.268	7.3	67	0.00
30	Benzo(k)fluoranthene	1.369	1.350	1.4	83	0.00
31 C	Benzo(a)pyrene	1.257	1.232	2.0	75	0.00
32	Dibenzo(a,h)anthracene	1.200	1.132	5.7	73	0.00
33	Benzo(g,h,i)perylene	5.122	4.902	4.3	74	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0