

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080416\
 Data File : BE092216.D
 Acq On : 5 Aug 2016 12:42
 Operator : UM/SJ
 Sample : SSTDCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCC0.4

Quant Time: Aug 05 18:03:48 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 16:59:37 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2	1,4-Dioxane	0.680	0.706	-3.8	79	0.00
3	n-Nitrosodimethylamine	0.942	0.752	20.2	79	0.00
4 S	2-Fluorophenol	1.284	1.150	10.4	80	0.01
5 S	Phenol-d6	1.711	1.534	10.3	77	0.01
6 I	Naphthalene-d8	1.000	1.000	0.0	84	-0.01
7 S	Nitrobenzene-d5	0.395	0.322	18.5	72	0.00
8	Naphthalene	1.159	1.116	3.7	84	0.00
9	2-Methylnaphthalene	0.743	0.709	4.6	87	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	87	-0.02
11 S	2,4,6-Tribromophenol	0.162	0.147	9.3	94	0.00
12 S	2-Fluorobiphenyl	1.584	1.513	4.5	87	0.00
13	Acenaphthylene	8.821	8.355	5.3	88	-0.02
14	Acenaphthene	1.406	1.368	2.7	87	0.00
15	Fluorene	1.783	1.683	5.6	87	0.00
16 I	Phenanthrene-d10	1.000	1.000	0.0	90	-0.02
17	Hexachlorobenzene	0.245	0.223	9.0	81	0.00
18	Pentachlorophenol	0.068	0.058	14.7	93	0.00
19	Phenanthrene	1.460	1.411	3.4	91	0.00
20	Anthracene	1.331	1.249	6.2	90	0.00
21	Fluoranthene	6.116	5.677	7.2	87	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
23	Pyrene	5.228	5.093	2.6	87	0.00
24 S	Terphenyl-d14	0.718	0.744	-3.6	93	-0.02
25	Benzo(a)anthracene	5.496	5.312	3.3	85	0.00
26	Chrysene	5.505	5.317	3.4	84	0.00
27	Indeno(1,2,3-cd)pyrene	5.978	5.370	10.2	80	0.00
28 I	Perylene-d12	1.000	1.000	0.0	79	0.00
29	Benzo(b)fluoranthene	1.368	1.271	7.1	72	0.00
30	Benzo(k)fluoranthene	1.369	1.370	-0.1	91	-0.01
31 C	Benzo(a)pyrene	1.257	1.249	0.6	82	0.00
32	Dibenzo(a,h)anthracene	1.200	1.151	4.1	80	0.00
33	Benzo(g,h,i)perylene	5.122	4.898	4.4	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0