

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080515\
 Data File : BE090382.D
 Acq On : 6 Aug 2015 9:07
 Operator : TP/UM
 Sample : G3180-10
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 080415-D506-E-D-S

Manual Integrations
 APPROVED

MMDadoda
 8/10/2015 10:11:16 AM

Quant Time: Aug 06 13:45:04 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 02:22:05 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	21322	5.00	ng	0.00
6) Naphthalene-d8	10.34	136	95167	5.00	ng	0.00
12) Acenaphthene-d10	14.20	164	62158	5.00	ng	0.00
18) Phenanthrene-d10	16.94	188	135760	5.00	ng	0.00
25) Chrysene-d12	21.10	240	160578	5.00	ng	0.00
32) Perylene-d12	23.40	264	147004	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.22	112	10011	2.67	ng	0.00
5) Phenol-d6	6.77	99	8846	1.48	ng	0.00
7) Nitrobenzene-d5	8.73	82	25736	3.39	ng	0.00
13) 2,4,6-Tribromophenol	15.71	330	11281	6.52	ng	0.00
14) 2-Fluorobiphenyl	12.82	172	68652	4.02	ng	0.00
27) Terphenyl-d14	19.56	244	135499	5.73	ng	0.00
Target Compounds						
30) Bis(2-ethylhexyl)phthalate	21.03	149	1570	0.29	ng	Qvalue # 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080515\
Data File : BE090382.D
Acq On : 6 Aug 2015 9:07
Operator : TP/UM
Sample : G3180-10
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
080415-D506-E-D-S

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:11:16 AM

Quant Time: Aug 06 13:45:04 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 06 02:22:05 2015
Response via : Initial Calibration

