

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080515\
 Data File : BE090369.D
 Acq On : 5 Aug 2015 21:07
 Operator : TP/UM
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE080515

Quant Time: Aug 06 13:01:22 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	102	0.00
2	1,4-Dioxane	1.000	1.225	-22.5	109	0.00
3	n-Nitrosodimethylamine	1.000	0.924	7.6	95	0.00
4 S	2-Fluorophenol	1.000	0.939	6.1	103	0.00
5 S	Phenol-d6	1.000	0.961	3.9	106	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	103	0.00
7 S	Nitrobenzene-d5	1.000	0.978	2.2	108	0.00
8	Nitrobenzene	1.000	0.949	5.1	110	0.00
9	Naphthalene	1.000	1.004	-0.4	104	0.00
10	Hexachlorobutadiene	1.000	1.079	-7.9	101	0.00
11	2-Methylnaphthalene	1.000	1.038	-3.8	111	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	105	0.00
13 S	2,4,6-Tribromophenol	1.000	1.007	-0.7	118	0.00
14 S	2-Fluorobiphenyl	1.000	1.052	-5.2	109	0.00
15	Acenaphthylene	1.000	1.040	-4.0	111	0.00
16	Acenaphthene	1.000	1.002	-0.2	105	0.00
17	Fluorene	1.000	1.044	-4.4	109	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	103	0.00
19	4-Bromophenyl-phenylether	1.000	1.047	-4.7	110	0.00
20	Hexachlorobenzene	1.000	1.038	-3.8	106	0.00
21	Pentachlorophenol	1.000	0.989	1.1	120	0.00
22	Phenanthrene	1.000	0.986	1.4	102	0.00
23	Anthracene	1.000	1.042	-4.2	110	0.00
24	Fluoranthene	1.000	1.039	-3.9	111	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	105	0.00
26	Pyrene	1.000	0.999	0.1	110	0.00
27 S	Terphenyl-d14	1.000	1.009	-0.9	107	0.00
28	Benzo(a)anthracene	1.000	0.948	5.2	111	0.00
29	Chrysene	1.000	1.035	-3.5	107	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.900	10.0	105	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.005	-0.5	130	0.00
32 I	Perylene-d12	5.000	5.000	0.0	107	0.00
33	Benzo(b)fluoranthene	1.000	1.015	-1.5	120	0.00
34	Benzo(k)fluoranthene	1.000	1.016	-1.6	106	0.00
35 C	Benzo(a)pyrene	1.000	1.022	-2.2	119	0.00
36	Dibenzo(a,h)anthracene	1.000	1.022	-2.2	133	0.00
37	Benzo(g,h,i)perylene	1.000	1.070	-7.0	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0