

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080515\
 Data File : BE090369.D
 Acq On : 5 Aug 2015 21:07
 Operator : TP/UM
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE080515

Quant Time: Aug 06 13:01:22 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.687	0.785	-14.3	109	0.00
3	n-Nitrosodimethylamine	0.773	0.715	7.5	95	0.00
4 S	2-Fluorophenol	0.750	0.704	6.1	103	0.00
5 S	Phenol-d6	1.241	1.197	3.5	106	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
7 S	Nitrobenzene-d5	0.331	0.329	0.6	108	0.00
8	Nitrobenzene	0.349	0.336	3.7	110	0.00
9	Naphthalene	1.104	1.108	-0.4	104	0.00
10	Hexachlorobutadiene	0.191	0.190	0.5	101	0.00
11	2-Methylnaphthalene	0.640	0.665	-3.9	111	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
13 S	2,4,6-Tribromophenol	0.095	0.083	12.6	118	0.00
14 S	2-Fluorobiphenyl	1.375	1.447	-5.2	109	0.00
15	Acenaphthylene	8.098	8.418	-4.0	111	0.00
16	Acenaphthene	1.259	1.261	-0.2	105	0.00
17	Fluorene	1.635	1.707	-4.4	109	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
19	4-Bromophenyl-phenylether	0.194	0.204	-5.2	110	0.00
20	Hexachlorobenzene	1.005	1.042	-3.7	106	0.00
21	Pentachlorophenol	0.044	0.032	27.3#	120	0.00
22	Phenanthrene	1.251	1.233	1.4	102	0.00
23	Anthracene	1.028	1.071	-4.2	110	0.00
24	Fluoranthene	1.231	1.280	-4.0	111	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
26	Pyrene	1.084	1.083	0.1	110	0.00
27 S	Terphenyl-d14	0.737	0.743	-0.8	107	0.00
28	Benzo(a)anthracene	1.108	1.050	5.2	111	0.00
29	Chrysene	1.339	1.386	-3.5	107	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.224	28.0#	105	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	0.959	-1.8	130	0.00
32 I	Perylene-d12	1.000	1.000	0.0	107	0.00
33	Benzo(b)fluoranthene	0.918	0.966	-5.2	120	0.00
34	Benzo(k)fluoranthene	1.392	1.415	-1.7	106	0.00
35 C	Benzo(a)pyrene	0.850	0.890	-4.7	119	0.00
36	Dibenzo(a,h)anthracene	0.766	0.820	-7.0	133	0.00
37	Benzo(g,h,i)perylene	0.923	0.988	-7.0	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0