

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080515\
 Data File : BE090376.D
 Acq On : 6 Aug 2015 1:18
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Aug 06 13:07:16 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	116	0.00
2	1,4-Dioxane	1.000	1.242	-24.2	125	0.00
3	n-Nitrosodimethylamine	1.000	1.018	-1.8	119	0.00
4 S	2-Fluorophenol	1.000	1.233	-23.3	154	0.00
5 S	Phenol-d6	1.000	1.055	-5.5	137	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	117	0.00
7 S	Nitrobenzene-d5	1.000	0.994	0.6	125	0.00
8	Nitrobenzene	1.000	0.963	3.7	128	0.00
9	Naphthalene	1.000	1.005	-0.5	118	0.00
10	Hexachlorobutadiene	1.000	1.074	-7.4	115	0.00
11	2-Methylnaphthalene	1.000	1.067	-6.7	130	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	118	0.00
13 S	2,4,6-Tribromophenol	1.000	1.220	-22.0	166	0.00
14 S	2-Fluorobiphenyl	1.000	1.064	-6.4	124	0.00
15	Acenaphthylene	1.000	1.049	-4.9	125	0.00
16	Acenaphthene	1.000	0.998	0.2	118	0.00
17	Fluorene	1.000	1.005	-0.5	117	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	112	0.00
19	4-Bromophenyl-phenylether	1.000	1.078	-7.8	123	0.00
20	Hexachlorobenzene	1.000	1.015	-1.5	113	0.00
21	Pentachlorophenol	1.000	1.335	-33.5#	192	0.00
22	Phenanthrene	1.000	0.995	0.5	112	0.00
23	Anthracene	1.000	1.063	-6.3	122	-0.02
24	Fluoranthene	1.000	1.098	-9.8	127	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	117	0.00
26	Pyrene	1.000	1.078	-7.8	131	0.00
27 S	Terphenyl-d14	1.000	1.075	-7.5	127	0.00
28	Benzo(a)anthracene	1.000	0.998	0.2	130	0.00
29	Chrysene	1.000	1.010	-1.0	115	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.131	-13.1	155	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.994	0.6	142	0.00
32 I	Perylene-d12	5.000	5.000	0.0	123	0.00
33	Benzo(b)fluoranthene	1.000	1.080	-8.0	148	0.00
34	Benzo(k)fluoranthene	1.000	0.956	4.4	114	0.00
35 C	Benzo(a)pyrene	1.000	1.047	-4.7	141	0.00
36	Dibenzo(a,h)anthracene	1.000	0.998	0.2	147	0.00
37	Benzo(g,h,i)perylene	1.000	1.063	-6.3	135	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0