

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080515\
 Data File : BE090364.D
 Acq On : 5 Aug 2015 18:05
 Operator : TP/UM
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8

Manual Integrations
 APPROVED

MMDadoda
 8/10/2015 10:10:23 AM

Quant Time: Aug 06 02:06:11 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 01:43:02 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	39958	5.00	ng	0.00
6) Naphthalene-d8	10.34	136	177686	5.00	ng	0.00
12) Acenaphthene-d10	14.20	164	102329	5.00	ng	0.00
18) Phenanthrene-d10	16.95	188	239769	5.00	ng	0.00
25) Chrysene-d12	21.10	240	295454	5.00	ng	0.00
32) Perylene-d12	23.40	264	275966	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.22	112	4846	0.66	ng	0.00
5) Phenol-d6	6.77	99	8413	0.89	ng	0.00
7) Nitrobenzene-d5	8.73	82	9979	0.94	ng	0.00
13) 2,4,6-Tribromophenol	15.69	330	1390	1.09	ng	-0.02
14) 2-Fluorobiphenyl	12.82	172	23774	0.81	ng	0.00
27) Terphenyl-d14	19.56	244	36719	0.69	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.20	88	4582	0.82	ng	99
3) n-Nitrosodimethylamine	3.49	42	5186	0.78	ng	# 91
8) Nitrobenzene	8.77	77	10526	0.96	ng	100
9) Naphthalene	10.39	128	32917	0.86	ng	98
10) Hexachlorobutadiene	10.69	225	5720	1.08	ng	99
11) 2-Methylnaphthalene	12.01	142	19052	0.81	ng	99
15) Acenaphthylene	13.91	152	138141	0.78	ng	100
16) Acenaphthene	14.26	154	21340	0.83	ng	97
17) Fluorene	15.26	166	28333	0.89	ng	98
19) 4-Bromophenyl-phenylether	16.14	248	7909	0.88	ng	99
20) Hexachlorobenzene	16.27	284	40240	1.13	ng	99
21) Pentachlorophenol	16.61	266	1312	0.91	ng	96
22) Phenanthrene	16.98	178	51074	0.92	ng	100
23) Anthracene	17.07	178	42075	0.86	ng	99
24) Fluoranthene	18.99	202	50228	0.91	ng	100
26) Pyrene	19.35	202	53161	0.69	ng	100
28) Benzo(a)anthracene	21.08	228	55566	0.97	ng	99
29) Chrysene	21.13	228	64849	0.92	ng	99
30) Bis(2-ethylhexyl)phthalate	21.03	149	11643	0.76	ng	99
31) Indeno(1,2,3-cd)pyrene	25.79	276	45558	1.35	ng	97
33) Benzo(b)fluoranthene	22.70	252	40555	1.02	ng	98
34) Benzo(k)fluoranthene	22.75	252	64615m	0.87	ng	
35) Benzo(a)pyrene	23.30	252	37424	0.99	ng	100
36) Dibenzo(a,h)anthracene	25.80	278	34001	1.33	ng	98
37) Benzo(g,h,i)perylene	26.52	276	41250	0.98	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE080515\
Data File : BE090364.D
Acq On : 5 Aug 2015 18:05
Operator : TP/UM
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC0.8

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM

Quant Time: Aug 06 02:06:11 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 06 01:43:02 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.59 min Scan# 846

Delta R.T. -0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Instrument :

BNA_E

ClientSampleId :

SSTDICC0.8

Tot Ion:152 Resp: 39958

Ion Ratio Lower Upper





152 100

150 153.3 123.0 184.4

115 61.2 48.9 73.3

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM



#2

1,4-Dioxane

Concen: 0.82 ng

RT: 3.20 min Scan# 88

Delta R.T. 0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05




Tgt Ion: 88 Resp: 4582



Ion Ratio Lower Upper

88 100

58 87.4 70.0 105.0

43 33.8 28.4 42.6









#3
n-Nitrosodimethylamine
Concen: 0.78 ng
RT: 3.49 min Scan# 132
Delta R.T. 0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Tot Ion	Ratio	Lower	Upper
42	100		
74	103.2	75.2	112.8
44	8.0	8.7	13.1#

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8


Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM



#4
2-Fluorophenol
Concen: 0.66 ng
RT: 5.22 min Scan# 386
Delta R.T. 0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Tgt Ion	Ratio	Lower	Upper
112	100		
64	71.0	37.1	55.7#
63	36.2	19.5	29.3#





#5

Phenol-d6

Concen: 0.89 ng

RT: 6.77 min Scan# 667

Delta R.T. -0.01 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.8

Tot Ion:	99	Resp:	8413
Ion	Ratio	Lower	Upper
99	100		
42	23.6	18.2	27.4
71	35.1	27.4	41.0

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM



#6

Naphthalene-d8

Concen: 5.00 ng




RT: 10.34 min Scan# 1371


Delta R.T. 0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Tgt Ion:	136	Resp:	177686
Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.7	13.1
54	10.9	9.4	14.0
68	7.6	6.5	9.7





#7

Nitrobenzene-d5

Concen: 0.94 ng

RT: 8.73 min Scan# 1093

Delta R.T. -0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Instrument :

BNA_E


ClientSampleId :

SSTDIC0.8




Tot Ion:	82	Resp:	9979
Ion	Ratio	Lower	Upper
82	100		
128	30.3	25.0	37.6
54	57.3	48.8	73.2

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM



Abundance Ion 82.00 (81.70 to 82.70): BE0
Ion 128.00 (127.70 to 128.70): BE0
Ion 54.00 (53.70 to 54.70): BE0



#8

Nitrobenzene

Concen: 0.96 ng


RT: 8.77 min Scan# 1102

Delta R.T. -0.00 min



Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Tgt Ion:	77	Resp:	10526
Ion	Ratio	Lower	Upper
77	100		
123	31.2	25.1	37.7
65	11.6	9.3	13.9



Abundance Ion 77.00 (76.70 to 77.70): BE0
Ion 123.00 (122.70 to 123.70): BE0
Ion 65.00 (64.70 to 65.70): BE0





#9

Naphthalene

Concen: 0.86 ng

RT: 10.39 min Scan# 1378

Delta R.T. 0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.8

Tot Ion:	128	Resp:	32917
Ion Ratio	Lower	Upper	
128	100		
129	11.5	9.8	14.6
127	12.8	10.6	16.0

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM



#10

Hexachlorobutadiene

Concen: 1.08 ng




RT: 10.69 min Scan# 1417

Delta R.T. 0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Tgt Ion:	225	Resp:	5720
Ion Ratio	Lower	Upper	
225	100		
223	62.0	50.6	75.8
227	63.4	50.7	76.1





#11

2-Methylnaphthalene

Concen: 0.81 ng

RT: 12.01 min Scan# 1650

Delta R.T. -0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.8

Tot Ion:	142	Resp:	19052
Ion Ratio	Lower	Upper	
142	100		
141	89.8	72.6	109.0
115	38.6	31.6	47.4

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM



#12

Acenaphthene-d10

Concen: 5.00 ng




RT: 14.20 min Scan# 2075

Delta R.T. -0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Tgt Ion:	164	Resp:	102329
Ion Ratio	Lower	Upper	
164	100		
162	103.9	81.8	122.8
160	52.2	44.2	66.4









#13
 2,4,6-Tribromophenol
 Concen: 1.09 ng
 RT: 15.69 min Scan# 2206
 Delta R.T. -0.02 min
 Lab File: BE090364.D
 Acq: 5 Aug 2015 18:05

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8

Tot Ion	Ratio	Lower	Upper
330	100		
332	91.4	75.8	113.6
141	154.5	114.4	171.6




Manual Integrations
 APPROVED



MMDadoda
 8/10/2015 10:10:23 AM



#14
 2-Fluorobiphenyl
 Concen: 0.81 ng
 RT: 12.82 min Scan# 1829
 Delta R.T. 0.00 min
 Lab File: BE090364.D
 Acq: 5 Aug 2015 18:05

Tgt Ion	Ratio	Lower	Upper
172	100		
171	33.8	28.5	42.7
170	23.1	19.8	29.8





#15

Acenaphthylene

Concen: 0.78 ng

RT: 13.91 min Scan# 2031

Delta R.T. -0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.8

Tot Ion:	152	Resp:	138141
Ion Ratio	Lower	Upper	
152	100		
151	4.8	3.8	5.8
153	13.1	10.3	15.5

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM



#16

Acenaphthene

Concen: 0.83 ng


RT: 14.26 min Scan# 2084

Delta R.T. -0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Tgt Ion:	154	Resp:	21340
Ion Ratio	Lower	Upper	
154	100		
153	469.4	376.2	564.2
152	278.6	235.0	352.4









#17
Fluorene
Concen: 0.89 ng
RT: 15.26 min Scan# 2181
Delta R.T. 0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
166	100		
165	99.8	81.5	122.3
167	13.8	11.4	17.0




Manual Integrations
APPROVED



MMDadoda
8/10/2015 10:10:23 AM



#18
Phenanthrene-d10
Concen: 5.00 ng
RT: 16.95 min Scan# 2278
Delta R.T. 0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Tgt Ion	Ratio	Lower	Upper
188	100		
94	5.9	5.7	8.5
80	6.1	5.8	8.8









#19
4-Bromophenyl-phenylether
Concen: 0.88 ng
RT: 16.14 min Scan# 2232
Delta R.T. 0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
248	100		
250	97.6	79.0	118.6
141	302.5	239.5	359.3


Manual Integrations
APPROVED


MMDadoda
8/10/2015 10:10:23 AM



#20
Hexachlorobenzene
Concen: 1.13 ng
RT: 16.27 min Scan# 2239
Delta R.T. 0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Tgt Ion	Ratio	Lower	Upper
284	100		
142	36.1	30.1	45.1
249	32.0	25.7	38.5









#21
Pentachlorophenol
Concen: 0.91 ng
RT: 16.61 min Scan# 2259
Delta R.T. 0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
266	100		
268	68.7	58.5	87.7
264	67.4	56.0	84.0




Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM



#22
Phenanthrene
Concen: 0.92 ng
RT: 16.98 min Scan# 2280
Delta R.T. 0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.4	23.2
179	15.4	12.5	18.7









#23
 Anthracene
 Concen: 0.86 ng
 RT: 17.07 min Scan# 2285
 Delta R.T. -0.02 min
 Lab File: BE090364.D
 Acq: 5 Aug 2015 18:05

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8

Tot Ion:178	Resp:	42075
Ion Ratio	Lower	Upper
178	100	
176	18.7	14.9 22.3
179	14.9	12.2 18.4




Manual Integrations
 APPROVED

MMDadoda
 8/10/2015 10:10:23 AM



#24
 Fluoranthene
 Concen: 0.91 ng
 RT: 18.99 min Scan# 2507
 Delta R.T. 0.00 min
 Lab File: BE090364.D
 Acq: 5 Aug 2015 18:05

Tgt Ion:202	Resp:	50228
Ion Ratio	Lower	Upper
202	100	
101	13.0	10.3 15.5
203	17.0	13.7 20.5





#25

Chrysene-d12

Concen: 5.00 ng

RT: 21.10 min Scan# 2865

Delta R.T. -0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Instrument :

BNA_E


ClientSampleId :

SSTDIC0.8




Tot Ion:	240	Resp:	295454
Ion Ratio	Lower	Upper	
240	100		
120	10.1	8.3	12.5
236	26.6	21.0	31.4

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E



#26

Pyrene

Concen: 0.69 ng


RT: 19.35 min Scan# 2575

Delta R.T. 0.00 min



Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Tgt Ion:	202	Resp:	53161
Ion Ratio	Lower	Upper	
202	100		
200	21.1	16.7	25.1
203	17.4	13.8	20.8



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E





#27

Terphenyl-d14

Concen: 0.69 ng

RT: 19.56 min Scan# 2615

Delta R.T. -0.00 min

Lab File: BE090364.D


Acq: 5 Aug 2015 18:05

Instrument :

BNA_E

ClientSampleId :




SSTDIC0.8



Tot Ion:	244	Resp:	36719
Ion Ratio	Lower	Upper	
244	100		
212	7.9	6.6	9.8
122	13.1	10.2	15.4

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM



#28

Benzo(a)anthracene

Concen: 0.97 ng




RT: 21.08 min Scan# 2862


Delta R.T. -0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Tgt Ion:	228	Resp:	55566
Ion Ratio	Lower	Upper	
228	100		
226	28.1	21.9	32.9
229	19.0	15.4	23.0









#29
Chrysene
Concen: 0.92 ng
RT: 21.13 min Scan# 2870
Delta R.T. -0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Instrument :
BNA_E
ClientSampled :
SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
228	100		
226	30.1	23.1	34.7
229	19.8	15.8	23.8




Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM



#30
Bis(2-ethylhexyl)phthalate
Concen: 0.76 ng
RT: 21.03 min Scan# 2854
Delta R.T. -0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Tgt Ion	Ratio	Lower	Upper
149	100		
167	25.3	19.8	29.8
279	3.4	2.4	3.6






#31
Indeno(1,2,3-cd)pyrene
Concen: 1.35 ng
RT: 25.79 min Scan# 3829
Delta R.T. -0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8




Tot Ion	Ratio	Lower	Upper
276	100		
138	24.7	18.8	28.2
277	24.6	18.2	27.4

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM






Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E




#32
Perylene-d12
Concen: 5.00 ng
RT: 23.40 min Scan# 3309
Delta R.T. 0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.0	20.1	30.1
265	21.6	17.2	25.8

Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E





#33

Benzo(b)fluoranthene

Concen: 1.02 ng

RT: 22.70 min Scan# 3155

Delta R.T. -0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.6	18.1	27.1
125	11.1	9.4	14.2

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM



#34

Benzo(k)fluoranthene

Concen: 0.87 ng m




RT: 22.75 min Scan# 3165

Delta R.T. -0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.5	18.1	27.1
125	11.0	9.1	13.7









#35
Benzo(a)pyrene
Concen: 0.99 ng
RT: 23.30 min Scan# 3286
Delta R.T. 0.00 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.8

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.6	26.4
125	12.2	9.8	14.6




Manual Integrations
APPROVED




MMDadoda
8/10/2015 10:10:23 AM



#36
Dibenzo(a,h)anthracene
Concen: 1.33 ng
RT: 25.80 min Scan# 3833
Delta R.T. -0.01 min
Lab File: BE090364.D
Acq: 5 Aug 2015 18:05

Tgt Ion	Ratio	Lower	Upper
278	100		
139	16.6	14.2	21.4
279	24.5	20.1	30.1





#37

Benzo(a,h,i)perylene

Concen: 0.98 ng

RT: 26.52 min Scan# 3989

Delta R.T. -0.00 min

Lab File: BE090364.D

Acq: 5 Aug 2015 18:05

Instrument :

BNA_E

ClientSampleId :

SSTDIC0.8

Tot Ion:	276	Resp:	41250
Ion Ratio	Lower	Upper	
276	100		
277	24.1	20.1	30.1
138	22.1	18.1	27.1

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:23 AM

