

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080515\
 Data File : BE090365.D
 Acq On : 5 Aug 2015 18:41
 Operator : TP/UM
 Sample : SSTDICCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICCC001

Manual Integrations
 APPROVED

MMDadoda
 8/10/2015 10:10:27 AM

Quant Time: Aug 06 02:10:25 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 01:43:02 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	18715	5.00	ng	0.00
6) Naphthalene-d8	10.34	136	83096	5.00	ng	0.00
12) Acenaphthene-d10	14.20	164	46182	5.00	ng	0.00
18) Phenanthrene-d10	16.94	188	111438	5.00	ng	0.00
25) Chrysene-d12	21.10	240	129406	5.00	ng	0.00
32) Perylene-d12	23.40	264	111332	5.00	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.22	112	2618m	0.76	ng	0.00
5) Phenol-d6	6.78	99	4306	0.97	ng	0.00
7) Nitrobenzene-d5	8.73	82	5219	1.05	ng	0.00
13) 2,4,6-Tribromophenol	15.71	330	685	1.18	ng	0.00
14) 2-Fluorobiphenyl	12.82	172	12900	0.97	ng	0.00
27) Terphenyl-d14	19.56	244	18899	0.81	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.20	88	2757	1.11	ng	100
3) n-Nitrosodimethylamine	3.49	42	2880	0.92	ng	100
8) Nitrobenzene	8.77	77	5202	1.02	ng	100
9) Naphthalene	10.39	128	18227	1.02	ng	100
10) Hexachlorobutadiene	10.69	225	3205	1.29	ng	100
11) 2-Methylnaphthalene	12.01	142	10192	0.93	ng	100
15) Acenaphthylene	13.91	152	74138	0.93	ng	100
16) Acenaphthene	14.26	154	11643	1.00	ng	100
17) Fluorene	15.26	166	15266	1.06	ng	100
19) 4-Bromophenyl-phenylether	16.14	248	4235	1.01	ng	100
20) Hexachlorobenzene	16.26	284	22465	1.35	ng	99
21) Pentachlorophenol	16.61	266	605	0.91	ng	100
22) Phenanthrene	16.98	178	27662	1.07	ng	100
23) Anthracene	17.08	178	22310	0.98	ng	100
24) Fluoranthene	18.99	202	26433	1.03	ng	100
26) Pyrene	19.35	202	26993	0.80	ng	100
28) Benzo(a)anthracene	21.08	228	25776	1.03	ng	100
29) Chrysene	21.14	228	35514	1.17	ng	100
30) Bis(2-ethylhexyl)phthalate	21.03	149	5789	0.83	ng	100
31) Indeno(1,2,3-cd)pyrene	25.79	276	20144	1.36	ng	99
33) Benzo(b)fluoranthene	22.70	252	19264	1.15	ng	100
34) Benzo(k)fluoranthene	22.75	252	31924m	1.07	ng	
35) Benzo(a)pyrene	23.30	252	17789	1.11	ng	100
36) Dibenzo(a,h)anthracene	25.81	278	14744	1.42	ng	100
37) Benzo(g,h,i)perylene	26.52	276	19848	1.13	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE080515\
Data File : BE090365.D
Acq On : 5 Aug 2015 18:41
Operator : TP/UM
Sample : SSTDICCC001
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICCC001

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:27 AM

Quant Time: Aug 06 02:10:25 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 06 01:43:02 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.59 min Scan# 846

Delta R.T. 0.00 min

Lab File: BE090365.D

Acq: 5 Aug 2015 18:41

Instrument :

BNA_E

ClientSampleId :

SSTDICCC001

Tot Ion:152 Resp: 18715

Ion Ratio Lower Upper





152 100

150 153.7 123.0 184.4

115 61.1 48.9 73.3

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:27 AM



#2

1,4-Dioxane

Concen: 1.11 ng

RT: 3.20 min Scan# 88

Delta R.T. 0.00 min

Lab File: BE090365.D

Acq: 5 Aug 2015 18:41




Tgt Ion: 88 Resp: 2757


Ion Ratio Lower Upper

88 100

58 87.5 70.0 105.0

43 35.5 28.4 42.6





#3





n-Nitrosodimethylamine
Concen: 0.92 ng
RT: 3.49 min Scan# 132
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Instrument :
BNA_E
ClientSampleId :
SSTDICCC001

Tot Ion	Ratio	Lower	Upper
42	100		
74	94.0	75.2	112.8
44	10.9	8.7	13.1

Manual Integrations
APPROVED




MMDadoda
8/10/2015 10:10:27 AM






#4

2-Fluorophenol
Concen: 0.76 ng m
RT: 5.22 min Scan# 386
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Tgt Ion	Ratio	Lower	Upper
112	100		
64	43.3	37.1	55.7
63	22.8	19.5	29.3









#5
Phenol-d6
Concen: 0.97 ng
RT: 6.78 min Scan# 668
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Tot Ion:	99	Resp:	4306
Ion	Ratio	Lower	Upper
99	100		
42	22.8	18.2	27.4
71	34.2	27.4	41.0

Instrument :
BNA_E
ClientSampleId :
SSTDICCC001


Manual Integrations
APPROVED


MMDadoda
8/10/2015 10:10:27 AM



#6
Naphthalene-d8
Concen: 5.00 ng
RT: 10.34 min Scan# 1371
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Tgt Ion:	136	Resp:	83096
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.7	13.1
54	11.7	9.4	14.0
68	8.1	6.5	9.7





#7

Nitrobenzene-d5

Concen: 1.05 ng

RT: 8.73 min Scan# 1094

Delta R.T. 0.00 min

Lab File: BE090365.D

Acq: 5 Aug 2015 18:41

Instrument :

BNA_E





ClientSampleId :

SSTDICCC001

Tot Ion:	82	Resp:	5219
Ion	Ratio	Lower	Upper
82	100		
128	31.3	25.0	37.6
54	61.0	48.8	73.2

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:27 AM



#8

Nitrobenzene

Concen: 1.02 ng




RT: 8.77 min Scan# 1103


Delta R.T. 0.00 min

Lab File: BE090365.D

Acq: 5 Aug 2015 18:41

Tgt Ion:	77	Resp:	5202
Ion	Ratio	Lower	Upper
77	100		
123	31.4	25.1	37.7
65	11.6	9.3	13.9





#9

Naphthalene

Concen: 1.02 ng

RT: 10.39 min Scan# 1378

Delta R.T. 0.00 min

Lab File: BE090365.D

Acq: 5 Aug 2015 18:41

Instrument :

BNA_E





ClientSampleId :

SSTDICCC001

Tot Ion:	128	Resp:	18227
Ion Ratio	Lower	Upper	
128	100		
129	12.2	9.8	14.6
127	13.3	10.6	16.0

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:27 AM



#10

Hexachlorobutadiene

Concen: 1.29 ng




RT: 10.69 min Scan# 1417


Delta R.T. 0.00 min

Lab File: BE090365.D

Acq: 5 Aug 2015 18:41

Tgt Ion:	225	Resp:	3205
Ion Ratio	Lower	Upper	
225	100		
223	63.2	50.6	75.8
227	63.4	50.7	76.1









#11
2-Methylnaphthalene
Concen: 0.93 ng
RT: 12.01 min Scan# 1651
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Instrument :
BNA_E
ClientSampleId :
SSTDICCC001

Tot Ion	Ratio	Lower	Upper
142	100		
141	90.8	72.6	109.0
115	39.5	31.6	47.4




Manual Integrations
APPROVED


MMDadoda
8/10/2015 10:10:27 AM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.20 min Scan# 2075
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.3	81.8	122.8
160	55.3	44.2	66.4









#13
 2,4,6-Tribromophenol
 Concen: 1.18 ng
 RT: 15.71 min Scan# 2206
 Delta R.T. 0.00 min
 Lab File: BE090365.D
 Acq: 5 Aug 2015 18:41

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC001

Tot Ion	Ratio	Lower	Upper
330	100		
332	95.6	75.8	113.6
141	145.1	114.4	171.6




Manual Integrations
 APPROVED




MMDadoda
 8/10/2015 10:10:27 AM



#14
 2-Fluorobiphenyl
 Concen: 0.97 ng
 RT: 12.82 min Scan# 1829
 Delta R.T. 0.00 min
 Lab File: BE090365.D
 Acq: 5 Aug 2015 18:41

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.6	28.5	42.7
170	24.8	19.8	29.8





#15

Acenaphthylene

Concen: 0.93 ng

RT: 13.91 min Scan# 2031

Delta R.T. 0.00 min

Lab File: BE090365.D

Acq: 5 Aug 2015 18:41

Instrument :

BNA_E





ClientSampleId :

SSTDICCC001

Tot Ion:	152	Resp:	74138
Ion Ratio	Lower	Upper	
152	100		
151	4.8	3.8	5.8
153	12.9	10.3	15.5

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:27 AM



#16

Acenaphthene

Concen: 1.00 ng


RT: 14.26 min Scan# 2084


Delta R.T. 0.00 min

Lab File: BE090365.D

Acq: 5 Aug 2015 18:41

Tgt Ion:	154	Resp:	11643
Ion Ratio	Lower	Upper	
154	100		
153	470.2	376.2	564.2
152	293.7	235.0	352.4









#17
Fluorene
Concen: 1.06 ng
RT: 15.26 min Scan# 2180
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Instrument :
BNA_E
ClientSampleId :
SSTDICCC001

Tot Ion	Ratio	Lower	Upper
166	100		
165	101.8	81.5	122.3
167	14.2	11.4	17.0




Manual Integrations
APPROVED




MMDadoda
8/10/2015 10:10:27 AM



#18
Phenanthrene-d10
Concen: 5.00 ng
RT: 16.94 min Scan# 2277
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Tgt Ion	Ratio	Lower	Upper
188	100		
94	7.1	5.7	8.5
80	7.3	5.8	8.8









#19
4-Bromophenyl-phenylether
Concen: 1.01 ng
RT: 16.14 min Scan# 2231
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Instrument :
BNA_E
ClientSampleId :
SSTDICCC001

Tot Ion	Ratio	Lower	Upper
248	100		
250	98.8	79.0	118.6
141	299.4	239.5	359.3


Manual Integrations
APPROVED


MMDadoda
8/10/2015 10:10:27 AM



#20
Hexachlorobenzene
Concen: 1.35 ng
RT: 16.26 min Scan# 2238
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Tgt Ion	Ratio	Lower	Upper
284	100		
142	37.1	30.1	45.1
249	31.9	25.7	38.5









#21
 Pentachlorophenol
 Concen: 0.91 ng
 RT: 16.61 min Scan# 2258
 Delta R.T. 0.00 min
 Lab File: BE090365.D
 Acq: 5 Aug 2015 18:41

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC001

Tot Ion:	266	Resp:	605
Ion Ratio	Lower	Upper	
266	100		
268	73.1	58.5	87.7
264	70.0	56.0	84.0




Manual Integrations
 APPROVED


MMDadoda
 8/10/2015 10:10:27 AM



#22
 Phenanthrene
 Concen: 1.07 ng
 RT: 16.98 min Scan# 2279
 Delta R.T. 0.00 min
 Lab File: BE090365.D
 Acq: 5 Aug 2015 18:41

Tgt Ion:	178	Resp:	27662
Ion Ratio	Lower	Upper	
178	100		
176	19.3	15.4	23.2
179	15.6	12.5	18.7









#23
 Anthracene
 Concen: 0.98 ng
 RT: 17.08 min Scan# 2285
 Delta R.T. 0.00 min
 Lab File: BE090365.D
 Acq: 5 Aug 2015 18:41

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC001

Tot Ion:178	Resp:	22310
Ion Ratio	Lower	Upper
178	100	
176	18.6	14.9 22.3
179	15.3	12.2 18.4




Manual Integrations
 APPROVED


MMDadoda
 8/10/2015 10:10:27 AM



#24
 Fluoranthene
 Concen: 1.03 ng
 RT: 18.99 min Scan# 2506
 Delta R.T. 0.00 min
 Lab File: BE090365.D
 Acq: 5 Aug 2015 18:41

Tgt Ion:202	Resp:	26433
Ion Ratio	Lower	Upper
202	100	
101	12.9	10.3 15.5
203	17.2	13.7 20.5





#25

Chrysene-d12

Concen: 5.00 ng

RT: 21.10 min Scan# 2865

Delta R.T. 0.00 min

Lab File: BE090365.D

Acq: 5 Aug 2015 18:41

Instrument :

BNA_E


ClientSampleId :

SSTDICCC001




Tot Ion:	240	Resp:	129406
Ion Ratio	Lower	Upper	
240	100		
120	10.4	8.3	12.5
236	26.2	21.0	31.4

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:27 AM



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E



#26

Pyrene

Concen: 0.80 ng


RT: 19.35 min Scan# 2574

Delta R.T. 0.00 min



Lab File: BE090365.D


Acq: 5 Aug 2015 18:41

Tgt Ion:	202	Resp:	26993
Ion Ratio	Lower	Upper	
202	100		
200	20.9	16.7	25.1
203	17.3	13.8	20.8



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E









#27
 Terphenyl-d14
 Concen: 0.81 ng
 RT: 19.56 min Scan# 2615
 Delta R.T. 0.00 min
 Lab File: BE090365.D
 Acq: 5 Aug 2015 18:41

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC001

Tot Ion	Ratio	Lower	Upper
244	100		
212	8.2	6.6	9.8
122	12.8	10.2	15.4




Manual Integrations
 APPROVED




MMDadoda
 8/10/2015 10:10:27 AM



#28
 Benzo(a)anthracene
 Concen: 1.03 ng
 RT: 21.08 min Scan# 2862
 Delta R.T. 0.00 min
 Lab File: BE090365.D
 Acq: 5 Aug 2015 18:41

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.9	32.9
229	19.2	15.4	23.0









#29
Chrysene
Concen: 1.17 ng
RT: 21.14 min Scan# 2870
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Tot Ion	Ratio	Lower	Upper
228	100		
226	28.9	23.1	34.7
229	19.8	15.8	23.8

Instrument :
BNA_E
ClientSampleId :
SSTDICCC001


Manual Integrations
APPROVED


MMDadoda
8/10/2015 10:10:27 AM



#30
Bis(2-ethylhexyl)phthalate
Concen: 0.83 ng
RT: 21.03 min Scan# 2854
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Tgt Ion	Ratio	Lower	Upper
149	100		
167	24.8	19.8	29.8
279	3.0	2.4	3.6









#31
Indeno(1,2,3-cd)pyrene
Concen: 1.36 ng
RT: 25.79 min Scan# 3828
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Instrument :
BNA_E
ClientSampleId :
SSTDICCC001

Tot Ion	Ratio	Lower	Upper
276	100		
138	24.1	18.8	28.2
277	23.4	18.2	27.4




Manual Integrations
APPROVED


MMDadoda
8/10/2015 10:10:27 AM



#32
Perylene-d12
Concen: 5.00 ng
RT: 23.40 min Scan# 3308
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.1	20.1	30.1
265	21.5	17.2	25.8









#33
Benzo(b)fluoranthene
Concen: 1.15 ng
RT: 22.70 min Scan# 3155
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Instrument :
BNA_E
ClientSampleId :
SSTDICCC001

Manual Integrations
APPROVED




MMDadoda
8/10/2015 10:10:27 AM


Tot Ion	Ratio	Lower	Upper
252	100		
253	22.6	18.1	27.1
125	11.8	9.4	14.2



#34
Benzo(k)fluoranthene
Concen: 1.07 ng m
RT: 22.75 min Scan# 3165
Delta R.T. 0.00 min
Lab File: BE090365.D
Acq: 5 Aug 2015 18:41

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.6	18.1	27.1
125	11.4	9.1	13.7





#35

Benzo(a)pyrene

Concen: 1.11 ng

RT: 23.30 min Scan# 3285

Delta R.T. 0.00 min

Lab File: BE090365.D

Acq: 5 Aug 2015 18:41

Instrument :

BNA_E


ClientSampleId :

SSTDICCC001




Tot Ion:	252	Resp:	17789
Ion Ratio	Lower	Upper	
252	100		
253	22.0	17.6	26.4
125	12.2	9.8	14.6

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:27 AM



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#36

Dibenzo(a,h)anthracene

Concen: 1.42 ng

RT: 25.81 min Scan# 3833




Delta R.T. 0.00 min


Lab File: BE090365.D

Acq: 5 Aug 2015 18:41

Tgt Ion:	278	Resp:	14744
Ion Ratio	Lower	Upper	
278	100		
139	17.8	14.2	21.4
279	25.1	20.1	30.1

Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E





#37

Benzo(a,h,i)perylene

Concen: 1.13 ng

RT: 26.52 min Scan# 3988

Delta R.T. 0.00 min

Lab File: BE090365.D


Acq: 5 Aug 2015 18:41

Instrument :

BNA_E

ClientSampleId :

SSTDICCC001



Tot Ion: 276 Resp: 19848

Ion Ratio Lower Upper

276 100

277 25.1 20.1 30.1

138 22.6 18.1 27.1

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:27 AM

