

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080515\
 Data File : BE090368.D
 Acq On : 5 Aug 2015 20:29
 Operator : TP/UM
 Sample : SSTDICC010
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC010

Manual Integrations
 APPROVED

MMDadoda
 8/10/2015 10:10:40 AM

Quant Time: Aug 06 02:14:12 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 01:43:02 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	38952	5.00	ng	0.00
6) Naphthalene-d8	10.34	136	181131	5.00	ng	0.00
12) Acenaphthene-d10	14.20	164	104442	5.00	ng	0.00
18) Phenanthrene-d10	16.94	188	239169	5.00	ng	0.00
25) Chrysene-d12	21.10	240	296330	5.00	ng	0.00
32) Perylene-d12	23.41	264	275322	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.22	112	76230	10.62	ng	0.00
5) Phenol-d6	6.76	99	131041	14.23	ng	-0.01
7) Nitrobenzene-d5	8.72	82	153597	14.18	ng	-0.01
13) 2,4,6-Tribromophenol	15.69	330	35724	12.96	ng	-0.02
14) 2-Fluorobiphenyl	12.82	172	297635	9.89	ng	0.00
27) Terphenyl-d14	19.56	244	470066	8.77	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.20	88	45405	10.26	ng	98
3) n-Nitrosodimethylamine	3.48	42	67096	10.32	ng	# 87
8) Nitrobenzene	8.76	77	163840	14.69	ng	94
9) Naphthalene	10.38	128	376605	9.64	ng	98
10) Hexachlorobutadiene	10.69	225	60315	11.13	ng	100
11) 2-Methylnaphthalene	12.00	142	257797	10.80	ng	98
15) Acenaphthylene	13.91	152	1840706	10.19	ng	99
16) Acenaphthene	14.26	154	256596	9.72	ng	86
17) Fluorene	15.26	166	336218	10.31	ng	98
19) 4-Bromophenyl-phenylether	16.14	248	102625	11.45	ng	97
20) Hexachlorobenzene	16.27	284	443222	12.42	ng	98
21) Pentachlorophenol	16.60	266	44420	11.78	ng	# 89
22) Phenanthrene	16.98	178	570609	10.31	ng	100
23) Anthracene	17.07	178	555253	11.35	ng	100
24) Fluoranthene	18.98	202	656986	11.87	ng	99
26) Pyrene	19.35	202	692170	8.96	ng	99
28) Benzo(a)anthracene	21.09	228	706163	12.29	ng	98
29) Chrysene	21.13	228	723202	10.85	ng	98
30) Bis(2-ethylhexyl)phthalate	21.03	149	330370	9.45	ng	98
31) Indeno(1,2,3-cd)pyrene	25.79	276	752186	13.10	ng	96
33) Benzo(b)fluoranthene	22.70	252	640984	11.55	ng	98
34) Benzo(k)fluoranthene	22.75	252	803077m	10.84	ng	
35) Benzo(a)pyrene	23.30	252	613088	12.03	ng	100
36) Dibenzo(a,h)anthracene	25.81	278	607524	12.44	ng	97
37) Benzo(g,h,i)perylene	26.53	276	636624	12.10	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE080515\
Data File : BE090368.D
Acq On : 5 Aug 2015 20:29
Operator : TP/UM
Sample : SSTDICC010
Misc :
ALS Vial : 9 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:40 AM

Quant Time: Aug 06 02:14:12 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 06 01:43:02 2015
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 7.58 min Scan# 845
Delta R.T. -0.01 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Instrument :
BNA_E
ClientSampleId :
SSTDIC010

Tot Ion	Ratio	Lower	Upper
152	100		
150	158.7	123.0	184.4
115	63.7	48.9	73.3

Manual Integrations
APPROVED




MMDadoda
8/10/2015 10:10:40 AM




#2

1,4-Dioxane
Concen: 10.26 ng
RT: 3.20 min Scan# 88
Delta R.T. 0.00 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
88	100		
58	89.1	70.0	105.0
43	34.1	28.4	42.6





#3





n-Nitrosodimethylamine
 Concen: 10.32 ng
 RT: 3.48 min Scan# 130
 Delta R.T. -0.01 min
 Lab File: BE090368.D
 Acq: 5 Aug 2015 20:29

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC010

Manual Integrations
 APPROVED

MMDadoda
 8/10/2015 10:10:40 AM




Tot Ion	Ratio	Lower	Upper
42	100		
74	107.3	75.2	112.8
44	7.8	8.7	13.1#




#4

2-Fluorophenol
 Concen: 10.62 ng
 RT: 5.22 min Scan# 386
 Delta R.T. 0.00 min
 Lab File: BE090368.D
 Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
112	100		
64	71.2	37.1	55.7#
63	37.2	19.5	29.3#





#5

Phenol-d6

Concen: 14.23 ng

RT: 6.76 min Scan# 665

Delta R.T. -0.01 min

Lab File: BE090368.D

Acq: 5 Aug 2015 20:29

Instrument :

BNA_E


ClientSampleId :

SSTDIC010




Tot Ion:	99	Resp:	131041
Ion Ratio	Lower	Upper	
99	100		
42	23.6	18.2	27.4
71	35.4	27.4	41.0

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:40 AM



Abundance Ion 99.00 (98.70 to 99.70): BE0
Ion 42.00 (41.70 to 42.70): BE0
Ion 71.00 (70.70 to 71.70): BE0



#6

Naphthalene-d8

Concen: 5.00 ng

RT: 10.34 min Scan# 1371




Delta R.T. 0.00 min


Lab File: BE090368.D

Acq: 5 Aug 2015 20:29

Tgt Ion:	136	Resp:	181131
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.7	13.1
54	11.0	9.4	14.0
68	7.8	6.5	9.7

Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BE0
Ion 68.00 (67.70 to 68.70): BE0





#7

Nitrobenzene-d5

Concen: 14.18 ng

RT: 8.72 min Scan# 1091

Delta R.T. -0.01 min

Lab File: BE090368.D

Acq: 5 Aug 2015 20:29

Instrument :

BNA_E

ClientSampleId :

SSTDIC010

Tot Ion: 82 Resp: 153597

Ion Ratio Lower Upper


82 100

128 32.7 25.0 37.6

54 58.7 48.8 73.2

Manual Integrations
APPROVED




MMDadoda
8/10/2015 10:10:40 AM



Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): BE0

Ion 54.00 (53.70 to 54.70): BE0



#8

Nitrobenzene

Concen: 14.69 ng

RT: 8.76 min Scan# 1100

Delta R.T. -0.01 min

Lab File: BE090368.D

Acq: 5 Aug 2015 20:29


Tgt Ion: 77 Resp: 163840

Ion Ratio Lower Upper

77 100

123 35.3 25.1 37.7



65 13.3 9.3 13.9




Abundance Ion 77.00 (76.70 to 77.70): BE0

Ion 123.00 (122.70 to 123.70): BE0

Ion 65.00 (64.70 to 65.70): BE0





#9

Naphthalene

Concen: 9.64 ng

RT: 10.38 min Scan# 1377

Delta R.T. -0.01 min

Lab File: BE090368.D

Acq: 5 Aug 2015 20:29

Instrument :

BNA_E





ClientSampleId :

SSTDIC010

Tot Ion:	128	Resp:	376605
Ion Ratio	Lower	Upper	
128	100		
129	11.0	9.8	14.6
127	13.3	10.6	16.0

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:40 AM



#10

Hexachlorobutadiene

Concen: 11.13 ng




RT: 10.69 min Scan# 1417


Delta R.T. 0.00 min

Lab File: BE090368.D

Acq: 5 Aug 2015 20:29

Tgt Ion:	225	Resp:	60315
Ion Ratio	Lower	Upper	
225	100		
223	63.1	50.6	75.8
227	63.5	50.7	76.1





#11

2-Methylnaphthalene

Concen: 10.80 ng

RT: 12.00 min Scan# 1649

Delta R.T. -0.01 min

Lab File: BE090368.D

Acq: 5 Aug 2015 20:29

Instrument :

BNA_E





ClientSampleId :

SSTDICC010

Tot Ion	Ratio	Lower	Upper
142	100		
141	89.9	72.6	109.0
115	37.7	31.6	47.4

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:40 AM



#12

Acenaphthene-d10

Concen: 5.00 ng




RT: 14.20 min Scan# 2076


Delta R.T. 0.00 min

Lab File: BE090368.D

Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.4	81.8	122.8
160	47.3	44.2	66.4









#13
 2,4,6-Tribromophenol
 Concen: 12.96 ng
 RT: 15.69 min Scan# 2206
 Delta R.T. -0.02 min
 Lab File: BE090368.D
 Acq: 5 Aug 2015 20:29

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC010

Tot Ion	Ratio	Lower	Upper
330	100		
332	97.2	75.8	113.6
141	152.3	114.4	171.6




Manual Integrations
 APPROVED




MMDadoda
 8/10/2015 10:10:40 AM



#14
 2-Fluorobiphenyl
 Concen: 9.89 ng
 RT: 12.82 min Scan# 1828
 Delta R.T. -0.00 min
 Lab File: BE090368.D
 Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.6	28.5	42.7
170	23.3	19.8	29.8





#15

Acenaphthylene

Concen: 10.19 ng

RT: 13.91 min Scan# 2032

Delta R.T. 0.00 min

Lab File: BE090368.D

Acq: 5 Aug 2015 20:29

Tot Ion:152 Resp: 1840706

Ion	Ratio	Lower	Upper
152	100		
151	4.9	3.8	5.8
153	13.1	10.3	15.5

152 100

151 4.9 3.8 5.8

153 13.1 10.3 15.5

Instrument :





BNA_E

ClientSampleId :

SSTDIC010

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:40 AM



#16

Acenaphthene

Concen: 9.72 ng

RT: 14.26 min Scan# 2085

Delta R.T. 0.00 min

Lab File: BE090368.D

Acq: 5 Aug 2015 20:29


Tgt Ion:154 Resp: 256596


Ion	Ratio	Lower	Upper
154	100		
153	458.2	376.2	564.2
152	237.4	235.0	352.4

154 100

153 458.2 376.2 564.2

152 237.4 235.0 352.4









#17
Fluorene
Concen: 10.31 ng
RT: 15.26 min Scan# 2181
Delta R.T. 0.00 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Instrument :
BNA_E
ClientSampleId :
SSTDICC010

Tot Ion	Ratio	Lower	Upper
166	100		
165	99.7	81.5	122.3
167	13.6	11.4	17.0




Manual Integrations
APPROVED




MMDadoda
8/10/2015 10:10:40 AM



#18
Phenanthrene-d10
Concen: 5.00 ng
RT: 16.94 min Scan# 2278
Delta R.T. 0.00 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
188	100		
94	5.7	5.7	8.5
80	5.9	5.8	8.8









#19
4-Bromophenyl-phenylether
Concen: 11.45 ng
RT: 16.14 min Scan# 2232
Delta R.T. 0.00 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Instrument :
BNA_E
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED


MMDadoda
8/10/2015 10:10:40 AM


Tot Ion	Ratio	Lower	Upper
248	100		
250	97.1	79.0	118.6
141	292.0	239.5	359.3



#20
Hexachlorobenzene
Concen: 12.42 ng
RT: 16.27 min Scan# 2239
Delta R.T. 0.00 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
284	100		
142	36.2	30.1	45.1
249	31.0	25.7	38.5









#21
 Pentachlorophenol
 Concen: 11.78 ng
 RT: 16.60 min Scan# 2258
 Delta R.T. -0.02 min
 Lab File: BE090368.D
 Acq: 5 Aug 2015 20:29

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC010

Tot Ion	Ratio	Lower	Upper
266	100		
268	58.3	58.5	87.7#
264	67.2	56.0	84.0




Manual Integrations
 APPROVED


MMDadoda
 8/10/2015 10:10:40 AM



#22
 Phenanthrene
 Concen: 10.31 ng
 RT: 16.98 min Scan# 2280
 Delta R.T. 0.00 min
 Lab File: BE090368.D
 Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.4	23.2
179	15.5	12.5	18.7









#23
 Anthracene
 Concen: 11.35 ng
 RT: 17.07 min Scan# 2285
 Delta R.T. -0.02 min
 Lab File: BE090368.D
 Acq: 5 Aug 2015 20:29

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC010

Tot Ion	Ratio	Lower	Upper
178	100		
176	18.8	14.9	22.3
179	15.5	12.2	18.4




Manual Integrations
 APPROVED


MMDadoda
 8/10/2015 10:10:40 AM



#24
 Fluoranthene
 Concen: 11.87 ng
 RT: 18.98 min Scan# 2506
 Delta R.T. -0.00 min
 Lab File: BE090368.D
 Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
202	100		
101	13.7	10.3	15.5
203	17.5	13.7	20.5





#25

Chrysene-d12

Concen: 5.00 ng

RT: 21.10 min Scan# 2865

Delta R.T. -0.00 min

Lab File: BE090368.D


Acq: 5 Aug 2015 20:29

Instrument :

BNA_E

ClientSampleId :




SSTDIC010



Tot Ion:	240	Resp:	296330
Ion Ratio	Lower	Upper	
240	100		
120	11.6	8.3	12.5
236	27.2	21.0	31.4

Manual Integrations
APPROVED

MMDadoda
8/10/2015 10:10:40 AM



#26

Pyrene


Concen: 8.96 ng

RT: 19.35 min Scan# 2575



Delta R.T. 0.00 min


Lab File: BE090368.D

Acq: 5 Aug 2015 20:29



Tgt Ion:	202	Resp:	692170
Ion Ratio	Lower	Upper	
202	100		
200	21.6	16.7	25.1
203	17.7	13.8	20.8









#27
 Terphenyl-d14
 Concen: 8.77 ng
 RT: 19.56 min Scan# 2615
 Delta R.T. -0.00 min
 Lab File: BE090368.D
 Acq: 5 Aug 2015 20:29

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC010

Tot Ion	Ratio	Lower	Upper
244	100		
212	8.0	6.6	9.8
122	12.1	10.2	15.4




Manual Integrations
 APPROVED


MMDadoda
 8/10/2015 10:10:40 AM



#28
 Benzo(a)anthracene
 Concen: 12.29 ng
 RT: 21.09 min Scan# 2863
 Delta R.T. 0.01 min
 Lab File: BE090368.D
 Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.8	21.9	32.9
229	20.1	15.4	23.0









#29
Chrysene
Concen: 10.85 ng
RT: 21.13 min Scan# 2870
Delta R.T. -0.00 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Instrument :
BNA_E
ClientSampleId :
SSTDIC010

Tot Ion	Ratio	Lower	Upper
228	100		
226	31.0	23.1	34.7
229	19.7	15.8	23.8




Manual Integrations
APPROVED


MMDadoda
8/10/2015 10:10:40 AM



#30
Bis(2-ethylhexyl)phthalate
Concen: 9.45 ng
RT: 21.03 min Scan# 2855
Delta R.T. 0.01 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
149	100		
167	25.7	19.8	29.8
279	3.3	2.4	3.6









#31
Indeno(1,2,3-cd)pyrene
Concen: 13.10 ng
RT: 25.79 min Scan# 3831
Delta R.T. 0.01 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Instrument :
BNA_E
ClientSampleId :
SSTDIC010

Tot Ion	Ratio	Lower	Upper
276	100		
138	25.4	18.8	28.2
277	25.0	18.2	27.4




Manual Integrations
APPROVED


MMDadoda
8/10/2015 10:10:40 AM



#32
Perylene-d12
Concen: 5.00 ng
RT: 23.41 min Scan# 3310
Delta R.T. 0.01 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.8	20.1	30.1
265	21.7	17.2	25.8









#33
Benzo(b)fluoranthene
Concen: 11.55 ng
RT: 22.70 min Scan# 3155
Delta R.T. -0.00 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Instrument :
BNA_E
ClientSampleId :
SSTDICC010

Tot Ion	Ratio	Lower	Upper
252	100		
253	21.8	18.1	27.1
125	10.9	9.4	14.2




Manual Integrations
APPROVED


MMDadoda
8/10/2015 10:10:40 AM



#34
Benzo(k)fluoranthene
Concen: 10.84 ng m
RT: 22.75 min Scan# 3166
Delta R.T. 0.00 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.3	18.1	27.1
125	10.6	9.1	13.7









#35
Benzo(a)pyrene
Concen: 12.03 ng
RT: 23.30 min Scan# 3287
Delta R.T. 0.01 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Instrument :
BNA_E
ClientSampleId :
SSTDIC010

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.6	26.4
125	11.9	9.8	14.6




Manual Integrations
APPROVED




MMDadoda
8/10/2015 10:10:40 AM



#36
Dibenzo(a,h)anthracene
Concen: 12.44 ng
RT: 25.81 min Scan# 3835
Delta R.T. 0.00 min
Lab File: BE090368.D
Acq: 5 Aug 2015 20:29

Tgt Ion	Ratio	Lower	Upper
278	100		
139	16.1	14.2	21.4
279	23.9	20.1	30.1





#37
 Benzo(a,h,i)perylene
 Concen: 12.10 ng
 RT: 26.53 min Scan# 3992
 Delta R.T. 0.01 min
 Lab File: BE090368.D
 Acq: 5 Aug 2015 20:29

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC010

Tot Ion	Ratio	Lower	Upper
276	100		
277	23.8	20.1	30.1
138	21.7	18.1	27.1

Manual Integrations
 APPROVED

MMDadoda
 8/10/2015 10:10:40 AM

