

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080517\
 Data File : BE093683.D
 Acq On : 5 Aug 2017 20:06
 Operator : SJ/JU
 Sample : I4427-06
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 LASB-13A-2

Quant Time: Aug 06 03:32:27 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 02 10:58:31 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	2983	0.40	ng	0.00
7) Naphthalene-d8	10.71	136	14739	0.40	ng	0.00
13) Acenaphthene-d10	14.53	164	8719	0.40	ng	0.00
19) Phenanthrene-d10	17.24	188	22328	0.40	ng	0.00
27) Chrysene-d12	21.40	240	22537	0.40	ng	-0.01
34) Perylene-d12	23.91	264	20140	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.52	112	2669	0.30	ng	0.00
5) Phenol-d6	7.09	99	4103	0.31	ng	0.00
8) Nitrobenzene-d5	9.06	82	3229	0.29	ng	0.00
11) 2-Methylnaphthalene-d10	12.29	152	5570	0.24	ng	0.00
14) 2,4,6-Tribromophenol	16.00	330	593	0.33	ng	0.00
15) 2-Fluorobiphenyl	13.15	172	8052	0.23	ng	-0.01
25) Fluoranthene-d10	19.27	212	56302	0.22	ng	0.00
29) Terphenyl-d14	19.86	244	9898	0.24	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
26) Fluoranthene	19.30	202	2317	0.030	ng	# 92
28) Pyrene	19.67	202	2357	0.032	ng	# 87
30) Benzo(a)anthracene	21.39	228	1860	0.028	ng	# 81
31) Chrysene	21.45	228	1846	0.025	ng	# 83
32) Bis(2-ethylhexyl)phthalate	21.30	149	22791	0.252	ng	# 98
35) Benzo(b)fluoranthene	23.15	252	2478	0.034	ng	# 54
37) Benzo(a)pyrene	23.80	252	1566	0.024	ng	# 44

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080517\
Data File : BE093683.D
Acq On : 5 Aug 2017 20:06
Operator : SJ/JU
Sample : I4427-06
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
LASB-13A-2

Quant Time: Aug 06 03:32:27 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 02 10:58:31 2017
Response via : Initial Calibration

