

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080517\
 Data File : BE093672.D
 Acq On : 5 Aug 2017 10:34
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 07 18:08:47 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 02 10:58:31 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	-0.01
2	1,4-Dioxane	0.648	0.653	-0.8	111	0.00
3	n-Nitrosodimethylamine	0.607	0.592	2.5	114	0.01
4 S	2-Fluorophenol	1.192	1.312	-10.1	126	0.00
5 S	Phenol-d6	1.794	1.951	-8.8	124	-0.01
6	bis(2-Chloroethyl)ether	1.508	1.466	2.8	110	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
8 S	Nitrobenzene-d5	0.305	0.328	-7.5	129	0.00
9	Naphthalene	4.630	4.545	1.8	111	0.00
10	Hexachlorobutadiene	0.172	0.169	1.7	113	-0.02
11 SURR	2-Methylnaphthalene-d10	0.632	0.634	-0.3	113	0.00
12	2-Methylnaphthalene	0.766	0.754	1.6	111	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
14 S	2,4,6-Tribromophenol	0.093	0.102	-9.7	149	0.00
15 S	2-Fluorobiphenyl	1.601	1.559	2.6	107	-0.02
16	Acenaphthylene	1.912	2.028	-6.1	122	0.00
17	Acenaphthene	1.349	1.347	0.1	112	0.00
18	Fluorene	1.803	1.751	2.9	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.126	0.116	7.9	96	0.00
21	Hexachlorobenzene	0.129	0.131	-1.6	109	0.00
22	Pentachlorophenol	0.055	0.075	-36.4#	166#	0.00
23	Phenanthrene	0.880	0.894	-1.6	115	0.00
24	Anthracene	1.203	1.300	-8.1	119	0.00
25 SURR	Fluoranthene-d10	4.541	4.824	-6.2	113	0.00
26	Fluoranthene	1.401	1.467	-4.7	112	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	114	-0.01
28	Pyrene	1.292	1.295	-0.2	115	0.00
29 S	Terphenyl-d14	0.719	0.703	2.2	110	0.00
30	Benzo(a)anthracene	1.176	1.305	-11.0	131	0.00
31	Chrysene	1.311	1.303	0.6	113	0.00
32	Bis(2-ethylhexyl)phthalate	1.603	3.052	-90.4#	253#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.281	-7.9	122	0.00
34 I	Perylene-d12	1.000	1.000	0.0	124	0.00
35	Benzo(b)fluoranthene	1.427	1.367	4.2	122	0.00
36	Benzo(k)fluoranthene	1.459	1.331	8.8	117	0.00
37 C	Benzo(a)pyrene	1.293	1.279	1.1	127	0.00
38	Dibenzo(a,h)anthracene	1.248	1.206	3.4	122	0.00
39	Benzo(g,h,i)perylene	1.266	1.197	5.5	120	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080517\
Data File : BE093672.D
Acq On : 5 Aug 2017 10:34
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 07 18:08:47 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 02 10:58:31 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		