

Data Path : Z:\HPCHEM1\BNA E\Data\BE080517\
 Data File : BE093675.D
 Acq On : 5 Aug 2017 14:59
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 06 03:11:17 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 02 10:58:31 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.648	0.600	7.4	105	0.00
3	n-Nitrosodimethylamine	0.607	0.585	3.6	116	0.00
4 S	2-Fluorophenol	1.192	1.273	-6.8	126	0.00
5 S	Phenol-d6	1.794	1.881	-4.8	123	0.00
6	bis(2-Chloroethyl)ether	1.508	1.447	4.0	111	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
8 S	Nitrobenzene-d5	0.305	0.327	-7.2	128	0.00
9	Naphthalene	4.630	4.510	2.6	109	0.00
10	Hexachlorobutadiene	0.172	0.170	1.2	113	-0.02
11 SURR	2-Methylnaphthalene-d10	0.632	0.631	0.2	112	0.00
12	2-Methylnaphthalene	0.766	0.754	1.6	111	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.093	0.101	-8.6	147	0.00
15 S	2-Fluorobiphenyl	1.601	1.576	1.6	107	-0.02
16	Acenaphthylene	1.912	2.013	-5.3	120	0.00
17	Acenaphthene	1.349	1.333	1.2	110	0.00
18	Fluorene	1.803	1.748	3.1	105	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	124	0.00
20	4-Bromophenyl-phenylether	0.126	0.111	11.9	98	0.00
21	Hexachlorobenzene	0.129	0.123	4.7	109	0.00
22	Pentachlorophenol	0.055	0.072	-30.9#	171#	0.00
23	Phenanthrene	0.880	0.861	2.2	118	0.00
24	Anthracene	1.203	1.368	-13.7	133	0.00
25 SURR	Fluoranthene-d10	4.541	4.703	-3.6	117	0.00
26	Fluoranthene	1.401	1.420	-1.4	116	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	119	-0.01
28	Pvrene	1.292	1.268	1.9	117	0.00
29 S	Terphenyl-d14	0.719	0.693	3.6	114	0.00
30	Benzo(a)anthracene	1.176	1.241	-5.5	130	0.00
31	Chrysene	1.311	1.283	2.1	116	0.00
32	Bis(2-ethylhexyl)phthalate	1.603	2.828	-76.4#	244#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.274	-7.3	127	0.00
34 I	Perylene-d12	1.000	1.000	0.0	130	0.00
35	Benzo(b)fluoranthene	1.427	1.342	6.0	125	0.00
36	Benzo(k)fluoranthene	1.459	1.363	6.6	125	0.00
37 C	Benzo(a)pyrene	1.293	1.267	2.0	131	0.00
38	Dibenzo(a,h)anthracene	1.248	1.198	4.0	126	0.00
39	Benzo(a,h,i)perylene	1.266	1.173	7.3	123	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE080517\
Data File : BE093675.D
Acq On : 5 Aug 2017 14:59
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 06 03:11:17 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 02 10:58:31 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			