

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080517\
 Data File : BE093675.D
 Acq On : 5 Aug 2017 14:59
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 06 03:11:17 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 02 10:58:31 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	114	0.00
2	1,4-Dioxane	0.400	0.370	7.5	105	0.00
3	n-Nitrosodimethylamine	0.400	0.385	3.8	116	0.00
4 S	2-Fluorophenol	0.400	0.427	-6.7	126	0.00
5 S	Phenol-d6	0.400	0.419	-4.7	123	0.00
6	bis(2-Chloroethyl)ether	0.400	0.384	4.0	111	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	112	0.00
8 S	Nitrobenzene-d5	0.400	0.429	-7.2	128	0.00
9	Naphthalene	0.400	0.390	2.5	109	0.00
10	Hexachlorobutadiene	0.400	0.397	0.8	113	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.400	0.0	112	0.00
12	2-Methylnaphthalene	0.400	0.394	1.5	111	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.400	0.445	-11.2	147	0.00
15 S	2-Fluorobiphenyl	0.400	0.394	1.5	107	-0.02
16	Acenaphthylene	0.400	0.421	-5.2	120	0.00
17	Acenaphthene	0.400	0.395	1.3	110	0.00
18	Fluorene	0.400	0.388	3.0	105	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	124	0.00
20	4-Bromophenyl-phenylether	0.400	0.356	11.0	98	0.00
21	Hexachlorobenzene	0.400	0.397	0.8	109	0.00
22	Pentachlorophenol	0.400	0.521	-30.3#	171	0.00
23	Phenanthrene	0.400	0.391	2.3	118	0.00
24	Anthracene	0.400	0.455	-13.7	133	0.00
25 SURR	Fluoranthene-d10	0.400	0.414	-3.5	117	0.00
26	Fluoranthene	0.400	0.406	-1.5	116	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	119	-0.01
28	Pyrene	0.400	0.392	2.0	117	0.00
29 S	Terphenyl-d14	0.400	0.386	3.5	114	0.00
30	Benzo(a)anthracene	0.400	0.422	-5.5	130	0.00
31	Chrysene	0.400	0.391	2.3	116	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.706	-76.5#	244	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.429	-7.2	127	0.00
34 I	Perylene-d12	0.400	0.400	0.0	130	0.00
35	Benzo(b)fluoranthene	0.400	0.376	6.0	125	0.00
36	Benzo(k)fluoranthene	0.400	0.374	6.5	125	0.00
37 C	Benzo(a)pyrene	0.400	0.392	2.0	131	0.00
38	Dibenzo(a,h)anthracene	0.400	0.384	4.0	126	0.00
39	Benzo(g,h,i)perylene	0.400	0.371	7.3	123	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080517\
Data File : BE093675.D
Acq On : 5 Aug 2017 14:59
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 06 03:11:17 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 02 10:58:31 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				