

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080715\
 Data File : BE090394.D
 Acq On : 7 Aug 2015 13:00
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 07 23:08:32 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.687	0.656	4.5	93	0.00
3	n-Nitrosodimethylamine	0.773	0.762	1.4	103	0.00
4 S	2-Fluorophenol	0.750	0.766	-2.1	114	0.00
5 S	Phenol-d6	1.241	1.228	1.0	111	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
7 S	Nitrobenzene-d5	0.331	0.315	4.8	106	0.00
8	Nitrobenzene	0.349	0.318	8.9	107	0.00
9	Naphthalene	1.104	1.105	-0.1	106	0.00
10	Hexachlorobutadiene	0.191	0.197	-3.1	108	0.00
11	2-Methylnaphthalene	0.640	0.629	1.7	108	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
13 S	2,4,6-Tribromophenol	0.095	0.076	20.0	113	0.00
14 S	2-Fluorobiphenyl	1.375	1.384	-0.7	109	0.00
15	Acenaphthylene	8.098	7.849	3.1	108	0.00
16	Acenaphthene	1.259	1.244	1.2	109	0.00
17	Fluorene	1.635	1.689	-3.3	112	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
19	4-Bromophenyl-phenylether	0.194	0.184	5.2	107	0.00
20	Hexachlorobenzene	1.005	0.956	4.9	105	0.00
21	Pentachlorophenol	0.044	0.033	25.0	133	0.00
22	Phenanthrene	1.251	1.220	2.5	109	0.00
23	Anthracene	1.028	1.001	2.6	111	-0.02
24	Fluoranthene	1.231	1.173	4.7	109	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
26	Pyrene	1.084	1.041	4.0	113	0.00
27 S	Terphenyl-d14	0.737	0.712	3.4	111	0.00
28	Benzo(a)anthracene	1.108	1.044	5.8	119	0.00
29	Chrysene	1.339	1.306	2.5	108	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.198	36.3#	100	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	0.795	15.6	116	0.00
32 I	Perylene-d12	1.000	1.000	0.0	111	0.00
33	Benzo(b)fluoranthene	0.918	0.950	-3.5	121	0.00
34	Benzo(k)fluoranthene	1.392	1.304	6.3	101	0.00
35 C	Benzo(a)pyrene	0.850	0.792	6.8	110	0.00
36	Dibenzo(a,h)anthracene	0.766	0.700	8.6	117	0.00
37	Benzo(g,h,i)perylene	0.923	0.853	7.6	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0