

Data Path : Z:\HPCHEM1\BNA E\Data\BE080717\
 Data File : BE093699.D
 Acq On : 7 Aug 2017 13:38
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE080717

Quant Time: Aug 07 14:33:31 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 07 14:30:21 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.603	0.609	-1.0	112	0.00
3	n-Nitrosodimethylamine	0.625	0.629	-0.6	110	0.01
4 S	2-Fluorophenol	1.294	1.299	-0.4	114	0.00
5 S	Phenol-d6	1.976	1.914	3.1	113	0.00
6	bis(2-Chloroethyl)ether	1.530	1.532	-0.1	116	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.342	0.318	7.0	112	0.00
9	Naphthalene	4.587	4.575	0.3	115	0.00
10	Hexachlorobutadiene	0.170	0.169	0.6	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.643	0.645	-0.3	115	0.00
12	2-Methylnaphthalene	0.774	0.764	1.3	114	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.110	0.091	17.3	105	0.00
15 S	2-Fluorobiphenyl	1.577	1.600	-1.5	117	0.00
16	Acenaphthylene	2.011	1.937	3.7	115	0.00
17	Acenaphthene	1.355	1.330	1.8	113	0.00
18	Fluorene	1.752	1.709	2.5	114	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.113	0.115	-1.8	120	0.00
21	Hexachlorobenzene	0.121	0.118	2.5	110	0.00
22	Pentachlorophenol	0.084	0.070	16.7	102	0.00
23	Phenanthrene	0.852	0.827	2.9	115	0.00
24	Anthracene	1.294	1.231	4.9	113	0.00
25 SURR	Fluoranthene-d10	4.564	4.317	5.4	104	0.00
26	Fluoranthene	1.397	1.309	6.3	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pvrene	1.310	1.320	-0.8	102	0.00
29 S	Terphenyl-d14	0.713	0.729	-2.2	104	0.00
30	Benzo(a)anthracene	1.269	1.228	3.2	99	0.00
31	Chrysene	1.295	1.275	1.5	99	0.00
32	Bis(2-ethylhexyl)phthalate	2.721	2.287	16.0	82	0.00
33	Indeno(1,2,3-cd)pyrene	1.248	1.259	-0.9	100	0.00
34 I	Perylene-d12	1.000	1.000	0.0	99	0.00
35	Benzo(b)fluoranthene	1.387	1.386	0.1	98	0.00
36	Benzo(k)fluoranthene	1.386	1.402	-1.2	103	0.00
37 C	Benzo(a)pyrene	1.309	1.307	0.2	99	0.00
38	Dibenzo(a,h)anthracene	1.225	1.220	0.4	99	0.00
39	Benzo(a,h,i)perylene	1.229	1.215	1.1	99	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE080717\
Data File : BE093699.D
Acq On : 7 Aug 2017 13:38
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE080717

Quant Time: Aug 07 14:33:31 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 07 14:30:21 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			