

Data Path : Z:\HPCHEM1\BNA E\Data\BE080717\
 Data File : BE093707.D
 Acq On : 7 Aug 2017 18:48
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 08 01:06:52 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 07 14:30:21 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.603	0.593	1.7	112	0.00
3	n-Nitrosodimethylamine	0.625	0.633	-1.3	113	0.01
4 S	2-Fluorophenol	1.294	1.233	4.7	111	0.00
5 S	Phenol-d6	1.976	1.850	6.4	111	0.00
6	bis(2-Chloroethyl)ether	1.530	1.488	2.7	115	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
8 S	Nitrobenzene-d5	0.342	0.302	11.7	109	0.00
9	Naphthalene	4.587	4.515	1.6	117	0.00
10	Hexachlorobutadiene	0.170	0.169	0.6	118	0.00
11 SURR	2-Methylnaphthalene-d10	0.643	0.640	0.5	118	0.00
12	2-Methylnaphthalene	0.774	0.759	1.9	117	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
14 S	2,4,6-Tribromophenol	0.110	0.086	21.8	101	0.00
15 S	2-Fluorobiphenyl	1.577	1.611	-2.2	120	0.00
16	Acenaphthylene	2.011	1.946	3.2	118	0.00
17	Acenaphthene	1.355	1.355	0.0	118	0.00
18	Fluorene	1.752	1.731	1.2	117	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
20	4-Bromophenyl-phenylether	0.113	0.122	-8.0	123	0.00
21	Hexachlorobenzene	0.121	0.127	-5.0	112	0.00
22	Pentachlorophenol	0.084	0.064	23.8	89	0.00
23	Phenanthrene	0.852	0.846	0.7	112	0.00
24	Anthracene	1.294	1.218	5.9	106	0.00
25 SURR	Fluoranthene-d10	4.564	4.333	5.1	100	0.00
26	Fluoranthene	1.397	1.318	5.7	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
28	Pvrene	1.310	1.314	-0.3	98	0.00
29 S	Terphenyl-d14	0.713	0.723	-1.4	99	0.00
30	Benzo(a)anthracene	1.269	1.208	4.8	93	0.00
31	Chrysene	1.295	1.316	-1.6	99	0.00
32	Bis(2-ethylhexyl)phthalate	2.721	1.996	26.6#	69	0.00
33	Indeno(1,2,3-cd)pyrene	1.248	1.274	-2.1	98	0.00
34 I	Perylene-d12	1.000	1.000	0.0	95	0.00
35	Benzo(b)fluoranthene	1.387	1.424	-2.7	97	0.00
36	Benzo(k)fluoranthene	1.386	1.389	-0.2	99	0.00
37 C	Benzo(a)pyrene	1.309	1.314	-0.4	96	0.00
38	Dibenzo(a,h)anthracene	1.225	1.240	-1.2	98	0.00
39	Benzo(a,h,i)perylene	1.229	1.222	0.6	96	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE080717\
Data File : BE093707.D
Acq On : 7 Aug 2017 18:48
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 08 01:06:52 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 07 14:30:21 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			